

Department of Chemical Engineering
Indian Institute of Technology, Delhi

November 07, 2012

Subject: Quotation for 1260 INFINITY REFRACTIVE INDEX DETECTOR

Following Specification are required for 1260 Infinity Refractive Index Detector

- | | | |
|-------------------------------------|---|---|
| 1. Detection Mode | : | Refractive index Detector with Deflection Method detection. |
| 2. Short Term Noise | : | 2.5×10^{-9} RIU or better. |
| 3. Drift | : | Not more than $< 200 \times 10^{-9}$ RIU / Hr. |
| 4. Refractive Range | : | 1.00 to 1.75 or better. |
| 5. Flow Cell | : | Should be 8 μ l or better. |
| 6. Date accusation rate | : | 37 Hz or better. |
| 7. This system should be compatible | : | With Agilent EZChrom Elite Version 3.3.2 SP2 software. |
| 8. Warranty | : | 12 (Months) |

Please send the above quotation latest by 29 / 11 / 2012.


Terms and Conditions:

- 1 Quotations must be made in sealed envelopes. Technical and Commercial bids must be sent separately in two sealed envelopes and then put together in one envelope. The quotes must reach the following address by 29th November, 2012 by 17: 00 hours latest.

Dr. A. S. Rathore
Department of Chemical Engineering,
Block II, Room No. 94,
Indian Institute of Technology, Delhi
Hauz Khas, New Delhi – 110016

- 2 Price must be quoted CIF New Delhi.
- 3 Please specify warranty periods.
- 4 Indian agency certificate must be enclosed. Proprietary certificate might be enclosed if applicable.
- 5 Payment through L/C.
- 6 Validity of quotation should be at least 3 months.
- 7 Period of delivery should be mentioned.
- 8 Educational discount should also be mentioned.

Remarks: The Institute reserves the right to accept or reject any all the quotations without assigning any reason thereof.



Dr. Anurag S. Rathore
Department of Chemical Engineering
IIT Delhi
Tel: 9650770650
Email: asrathore@biotechmz.com