

**Corrigendum to the Tender reference NO: IITD/EBSC/2513**

Please read the “Sensitivity” specification in the Tender No. 2513 as per the following:

Sensitivity	Full scan sensitivity in MS Mode with Reserpine 10 pg/microliter at S/N>10:1; and MS/MS Mode : Reserpine 150 fg/microliter at S/N>500:1 or preferably better.