

**NANOSCALE RESEARCH FACILITY (NRF)
INDIAN INSTITUTE OF TECHNOLOGY DELHI
HAUZ KHAS, NEW DELHI- 110016**

Date: Nov. 16, 2012

NOTICE INVITING QUOTATIONS

Ref: No. NRF/UV-LT-PL/2012

Extended Due date: Jan. 18, 2012

Please send your quotations to the undersigned in a sealed cover super scribed with our Ref. no. & due date for the following items (UV low temp Photoluminescence system with multiple options).

S.No.	Name of item with full technical specifications	quantity
1	<p>UV low temp Photoluminescence system for semiconductor thin films and nanostructures:</p> <p>Excitations source: Xe-arc lamp (450 W) (ozone free, spectral range 300 nm - 1700 nm) along with external cw lasers such as 325 nm He-Cd laser (with variable output power: $\leq 5\text{mW}$ to $\geq 20\text{ mW}$) and 514 nm Argon laser (with variable output power: $\leq 5\text{mW}$ to $\geq 20\text{ mW}$), cw laser 785 nm (maximum output power: 100 mW); for lamp the excitation range must be continuously tunable within 200 nm – 950 nm.</p> <p>Laser spot size: $\leq 1\ \mu\text{m}$ Spectral range: 300-1000 nm Wavelength accuracy: $\pm 0.25\text{ nm}$ or better Monochromators: Two (one at source side other at detector side) double grating aberration corrected holographic grating (1800 gr/mm or better), wavelength range: from $\leq 200\text{ nm}$ to $\geq 1000\text{ nm}$, resolution 0.05 nm, focal length $\geq 500\text{ mm}$, motorized.</p> <p>Emission detector: TEC cooled ($\leq -70\text{ }^\circ\text{C}$) CCD detector, Pixel: 1024x256, pixel size: $26\ \mu\text{m} \times 26\ \mu\text{m}$, Spectral range: $\leq 200\text{ nm}$ to $\geq 1000\text{ nm}$, Quantum optical efficiency $\geq 50\%$ between 500 nm to 800 nm, dark noise $\leq 0.0025\text{ e-}/\text{pixel}/\text{sec}$.</p> <p>Electronics: Data acquisition system must have at least 16 bit DA/AD board with PCI interface, AD sampling $\geq 500\text{ kHz}$.</p> <p>Sample heating/cooling option: Sample temperature should be variable between 5 K to 300 K using closed cycle He cryostat with an set point accuracy of 1K and temperature stability of 0.1 K with a suitable temperature controller.</p> <p>Sample mapping: Sub-micron mapping of nanostructures (spatial resolution at least 250 nm). The system should able to do spectral mapping at least for 2 inch wafer. All the XYZ stages must be motorized with step resolution of 100 nm and computer controlled.</p> <p>Sample holders: Our samples will be semiconductors thin film or</p>	01

	<p>nanostructures, solid state samples (sample size at least 2 inch diameter and 1 mm thickness); also include cuvet cell holder.</p> <p>Software: i. Fully functional software for data acquisition and data analysis using a Windows based operating system (Windows XP or later versions).</p> <p>i. Scientific publication-quality graphics and layout capabilities should be available within the control and analysis software environment</p> <p>ii. Computer: Windows based operating system with flat panel Monitor (24 inch or larger), 500 GB HDD, CD/DVD writer, USB ports (8), should be able to export files to the clipboard or save as JPEG, PNG, BMP, TIFF etc.</p> <p>Power: 220-250 Vac 50 Hz</p>	
2	All the required accessories for confocal (10X, 20X and 100X) Raman measurements (both stokes and antistokes); spectral resolution $\leq 1 \text{ cm}^{-1}/\text{pixel}$ in UV and visible.	1
3	Desirable option: Attachment for coupling with Atomic Force Microscopy system should be quoted separately.	1

Quoted price should include all the calibration samples, all the accessories needed for the system, and all the necessary installation and training provided by the vendor. **PL mapping on a semiconductor material should be demonstrable.** A compliance chart is required for all the technical specifications mentioned above. **Users list:** Vendors should provide user list for the said item in India.

Warranty: (Required) On-site comprehensive including part replacement for 2 years. Additional warranty per annum may be quoted. This should be clearly shown in the technical as well as financial bid.

Terms and conditions covering submission of quotations

1. **DELIVERY:** All prices quoted for FOB
2. **TERMS OF PAYMENT:** Letter of credit
3. **VALIDITY OF QUOTATIONS:** **three months or more**
4. **CORRESPONDENCE:** No correspondence regarding acceptance /rejection of quotation will be entertained.
5. **SUBMISSION OF QUOTATIONS:** Separate quotations should be submitted for technical bid and commercial bid in two separate and clearly marked envelopes.
6. **DISCOUNTS/REBATES:** Special discounts/rebate wherever admissible keeping in the view that supplies are being made for an Educational institute may be indicated in the offer.

7. DIRECTOR'S RIGHT:

Director, IIT Delhi reserves the right of acceptance or rejection of any or all quotations without assigning any reason.

Please specify terms and conditions. The quotations must have a validity of 3 months. Sealed quotations (separate technical and financial) may be send to the following address.

Dr.J.P. Singh
Associate Professor
Block VI, Room no. 121 (NRF)
Indian Institute of Technology Delhi
Hauz Khas, New Delhi 110016
INDIA
E-mail: jpsingh@physics.iitd.ac.in