

Department of Chemical Engineering
Indian Institute of Technology, Delhi

January, 31st 2013

Subject: Quotation for O2 & CO2 off GAS ANALYZER

Following Specification are required for O2 & CO2 off GAS ANALYZER

O2 gas range : 1-50%

Co2 gas range: 0-25%

O2 detector : Zirconium dioxide

O2 sensor accuracy : $\pm 3\%$

Co2 detector: Infrared

Co2 sensor accuracy : $\pm 3\%$

Operating temp range : 15-40C

Operating pressure range : 11-18 psi

Operating Humidity : 0-100% RH

Calibration : One point

Third party software Integration : Yes

Please send the above quotation latest by 14 / 02 / 2013.

Terms and Conditions:

- 1 Quotations must be made in sealed envelopes. Technical and Commercial bids must be sent separately in two sealed envelopes and then put together in one envelope. The quotes must reach the following address by 04^h February, 2013 by 17: 00 hours latest.**

**Dr. A. S. Rathore
Department of Chemical Engineering,
Block II, Room No. 94,
Indian Institute of Technology, Delhi
Hauz Khas, New Delhi – 110016**

- 2 Price must be quoted CIF New Delhi.**
- 3 Please specify warranty periods.**
- 4 Indian agency certificate must be enclosed. Proprietary certificate might be enclosed if applicable.**

- 5 Payment through L/C.**
- 6 Validity of quotation should be at least 3 months.**
- 7 Period of delivery should be mentioned.**
- 8 Educational discount should also be mentioned.**

Remarks: The Institute reserves the right to accept or reject any all the quotations without assigning any reason thereof.

Dr. Anurag S. Rathore
Department of Chemical Engineering
IIT Delhi
Tel: 9650770650
Email: asrathore@biotechmz.com