

**INDIAN INSTITUTE OF TECHNOLOGY DELHI
HAUZ KHAS, NEW DELHI-110016**

Dated: 13/01/2023

Subject: - Subject: - Corrigendum for “Flue Gas Analyzer” Tender ID- “2022_IIT_732236_1” & tender reference No“ IITD/DESE(SP-4064)/2022”
”

S. No	Technical Specification Given as	Technical Specification Read as(Corrigendum)
1.	Flue Gas Analyzer	CO, CO ₂ (NDIR), O ₂ , NO₂ , NO
2.	CO Resolution	1 ppm
7.	HC NO₂ Resolution	< 0.3 ppm
8.	HC NO₂ Corrected	Yes
15.	Warranty of the flue gas analyzer	2 years
16.	Gas Collecting Metallic Probe	Quantity: 1

Addendum

Technical Specification should be add in technical specification		
S. No	Component	Details
21.	CO Range	0 – 4500 ppm
22.	NO Range	0 – 4000 ppm
23.	NO ₂ Range	0 – 400 ppm
24.	CO ₂ Range	0 – 40% (Vol.)
25.	O ₂ Range	0 – 21% (Vol.)
26.	Gas Collecting Metallic Probe Length	> 600mm

COMPLIANCE SHEET

Technical Specification			
S. No	Component	Details	COMPLIANCE Y/N
21.	CO Range	0 – 4500 ppm	
22.	NO Range	0 – 4000 ppm	
23.	NO ₂ Range	0 – 400 ppm	
24.	CO ₂ Range	0 – 40% (Vol.)	
25.	O ₂ Range	0 – 21% (Vol.)	
26.	Gas Collecting Metallic Probe Length	> 600mm	

All Prospective Bidders are requested to take note on the above change and submit their bids as per the new clause.