

## **CORRIGENDUM**

### **NIQ “Procurement of Ultrapure Water Purification System” (NID 4093)**

Specification of above item NIQ may be read as under:

1. Instant flow rate(with application Pak final filter)..... 1.0-1.5 LPM
2. Production rate .....5 L/h
3. Water storage tank should be of HDPE(High Density Polyethylene) with level sensors to show the water level in the main systems display
4. RO system at the upstream of Water purification system to treat the tap water which is used for the preparation of type II water.
5. Other things will be remain the same.

**Last date for the submission of quotation is extended up to May 16, 2014 (5:00PM)**

**Dr. Gazala Habib**  
Department of Civil Engineering  
Indian Institute of Technology, Delhi  
Hauz Khas, New Delhi- 110016  
Phone No: 011-2659-1192