

Department of Chemistry

Corrigendum (18/8/2012)

Confocal Raman microscope (NIT no: 2113)

one correction in the main specification section
regarding CCD detector;

The Pixel size should be $\leq 16 \mu\text{m} \times 16 \mu\text{m}$

instead of 16 mm x 16 mm



Prof A K Ganguli