

**INDIAN INSTITUTE OF TECHNOLOGY DELHI
HAUZ KHAS, NEW DELHI-110016**

Dated: 10/12/2015

Subject: - Corrigendum for Purchase of "(DC power supply unit)

	Section and Point No.	Existing Clause		Revised Clause	
1.	Page No.1 (Technical specification)	Input	230V AC $\pm 10\%$, 3 ϕ , 4 wire , up to 4kW	Input	440V AC $\pm 10\%$, 3 ϕ , 4 wire , up to 4kW

Prospective Bidders are requested to take note on the above change and submit their bids as per the revised clause.