Notice Inviting Quotation (E-Publishing mode)

INDIAN INSTITUTE OF TECHNOLOGY DELHI HAUZ KHAS, NEW DELHI-110016

Dated: 13/02/2017

Open Tender Notice No.

Indian Institute of Technology Delhi is in the process of purchasing following item(s) as per details as given as under.

Details of the item	Low Temperature (Chiller) unit for rotavapour
Earnest Money Deposit to be submitted	Rs. 0
Warranty	As applicable
Performance security	0

Tender Documents may be downloaded from Central Public Procurement Portal <u>http://eprocure.gov.in/epublish/app</u>. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <u>http://eprocure.gov.in/epublish/app</u>. The portal enrolment is free of cost. This is an offline tender process where bidders needs to submit physical bids at office of the indenter (Ref. Para 7 of Schedule, Next page).

Tenderers can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, Click on "GO" button to view all IIT Delhi tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document offline on the website <u>http://eprocure.gov.in/epublish/app</u> as per the schedule given in the next page.

SCHEDULE

EMD AmountRs. 0 (Rupees 0 only)(If applicable)(To be paid through RTGS/NEFT. IIT Del are as under: Name of the Bank A/C : IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 100703572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 100713572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank A/C :: IITD Revenue A SBI A/C No. : SBIN0001077 MICR Code :: 110002156 Swift No. : SBININBB547 (This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the UT the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100169.Contact Phone Numbers(+91)-11-26591502	
are as under: Name of the Bank A/C : IITD Revenue A SBI A/C No. : 10773572622 Name of the Bank : State Bank of Ir Hauz Khas, 110016 IFSC Code IFSC Code : SBIN0001077 MICR Code : 110002156 Swift No. : SBINNBB547 (This is mandatory that UTR Number is prooffline quotation/bid. (Kindly refer to the U' the Declaration Sheet at Annexure-II) 2. Issue of Tender Document 13/02/17 to 27/02/17 (except Saturdays / Sund Holidays) 3. Last date for receipt of queries 24/02/17 4. Date of pre bid meeting 5. Last Date and Time for receipts of Bids 01/03/17 Hrs. on - 4:00 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Name of the Bank A/C: IITD Revenue A SBI A/C No.SBI A/C No.: 10773572622 Name of the Bank: State Bank of Ir HauzHauzKhas,110016IFSC Code: SBIN0001077 MICR CodeIISC Code: SBINNNBB547 (This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the U the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	hi Bank details
SBI A/C No. : 10773572622 Name of the Bank : State Bank of Ir Hauz Khas, 110016 IFSC Code : SBIN0001077 MICR Code : 110002156 Swift No. : SBININBB547 (This is mandatory that UTR Number is pro-offline quotation/bid. (Kindly refer to the UT the Declaration Sheet at Annexure-II) 13/02/17 to 27/02/17 (except Saturdays / Sund Holidays) 3. Last date for receipt of queries 24/02/17 4. Date of pre bid meeting NA 5. Last Date and Time for receipts of Bids Upto 27/02/17 Hrs. on 5:30 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Name of the Bank: State Bank of Ir HauzHauzKhas, 110016IFSC Code: SBIN0001077MICR Code: 110002156Swift No.: SBININBB547(This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the UT the Declaration Sheet at Annexure-II)2.Issue of Tender Document3.Last date for receipt of queries4.Date of pre bid meeting5.Last Date and Time for receipts of Bids6.Opening of Technical Bid7.Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of Communication Ravi P Singh , Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	Account
HauzHauzKhas, 110016 IFSC CodeHauzKhas, 110016 IFSC CodeIFSC Code: SBIN0001077 MICR Code: I10002156 Swift No.: SBININBB547 (This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the UT the Declaration Shet at Annexure-II)Image: Subscript of the Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)Image: Subscript of Queries24/02/17Image: Subscript of Queries24/02/17	
110016IFSC Code: SBIN0001077MICR Code: 110002156Swift No.: SBININBB547(This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the UT the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Technology, Hauz Khas, New Delhi - 110016	
Image: Section of the section of th	New Delhi-
MICR Code: 110002156Swift No.: SBININBB547(This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the UT the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Swift No.: SBININBB547 (This is mandatory that UTR Number is pro offline quotation/bid. (Kindly refer to the U the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of Bids and Institute of Technology, Hauz Khas, New Delhi - 110016Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Image: Second	
offline quotation/bid. (Kindly refer to the U' the Declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Technology, Hauz Khas, New Delhi - 110016	
Image: constraint of the declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	vided in the
Image: constraint of the declaration Sheet at Annexure-II)2.Issue of Tender Document13/02/17 to 27/02/17 (except Saturdays / Sund Holidays)3.Last date for receipt of queries24/02/174.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	FR Column of
 Holidays) 3. Last date for receipt of queries 24/02/17 4. Date of pre bid meeting NA 5. Last Date and Time for receipts of Bids Upto 27/02/17 Hrs. on 5:30 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 	
 Holidays) 3. Last date for receipt of queries 24/02/17 4. Date of pre bid meeting NA 5. Last Date and Time for receipts of Bids Upto 27/02/17 Hrs. on 5:30 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 	lavs and
 3. Last date for receipt of queries 24/02/17 4. Date of pre bid meeting 5. Last Date and Time for receipts of Bids Upto 27/02/17 Hrs. on 5:30 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 	
4.Date of pre bid meetingNA5.Last Date and Time for receipts of BidsUpto 27/02/17 Hrs. on 5:30 pm6.Opening of Technical Bid01/03/17 Hrs. on - 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
5. Last Date and Time for receipts of Bids Upto 27/02/17 Hrs. on 5:30 pm 6. Opening of Technical Bid 01/03/17 Hrs. on - 4:00 pm 7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
6.Opening of Technical Bid01/03/17 Hrs. on – 4:00 pm7.Place of Bid Submission & Opening of BidsRavi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 1100168.Address of CommunicationRavi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
7. Place of Bid Submission & Opening of Bids Ravi P Singh Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016 8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
8. Address of Communication Ravi P Singh, Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Department of Chemistry, Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Indian Institute of Technology, Hauz Khas, New Delhi - 110016	
Hauz Khas, New Delhi - 110016	
9. Contact Phone Numbers (+91)-11- 26591502	
10. Fax Number (+91)-11- 26581102	
11. E-mail Address ravips@chemistry.iitd.ac.in	

*Tender document can also be downloaded from IIT Delhi Website (www.iitd.ac.in/tenders).

Department of Chemistry

Indian Institute of Technology Hauz Khas, New Delhi-110 016

NOTICE INVITING QUOTATIONS

Dated: XX/XX/XXXX

Subject : Low temperature (Chiller) unit for rotavapour

Invitation for Tender Offers

Indian Institute of Technology Delhi invites offline Bids (Technical bid and Commercial bid) from eligible and experienced OEM (Original Equipment Manufacturer) OR OEM Authorized Dealer for **supply**, **installation & integration of Low temperature (Chiller) unit for rotavapour** with (warranty period as stated at page #1 of this tender) on site comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document, which is available on CPP Portal <u>http://eprocure.gov.in/epublish/app</u>.

TECHNICAL SPECIFICATION:

Sl. No.	Technical Specifications		
		Specification of Recirculating Cooler	
	1	Working temperature range (°C)	-20 40
	2	Temperature stability (°C)	±0.5
	3	Cooling Capacity (KW)	At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.1KW,
	6	Pump capacity flow rate (I/min)	15
	7	Pump capacity flow pressure (bar)	0.35
	8	Filling volume liters	34.5
	9	Digital interface	RS232
	10	Dimensions W x L x H (cm)	23-25 x48 50 x 58-60
	11	Weight (kg)	Less than 39Kg
		Desirable Features	
		Splash-proof keypad with integrated mains switch	
		Removable venting grid for cleaning of the condenser	

Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) 6 Pump capacity flow rate (l/min) 7 Pump capacity flow rete (l/min) 7 Pump capacity flow rete (l/min) 7 Pump capacity flow rate (l/min) 7 Pump capacity flow rete (l/min) 7 Pump capacity flow pressure (bar) 0.35 3 10 Digital interface 8 Filling volume litters 3 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Destrable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class accor			Filling level indicator	
+80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length U Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) 6 Pump capacity flow rate (l/min) 15 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 3 Digital interface RS232 10 Dimensions W x L H (cm) 2.25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Inblings , Insulat			-	
No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length 2 Temperature stability (*C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 10°C 0.1KW, 6 Pump capacity flow rate (l/min) 7 Pump capacity flow rate (l/min) 8 Filling volume liters 3 Cooling Capacity flow rate (l/min) 1 Weight (kg) 0 Less than 39Kg 0 Desirable Features 5 Splash-proof keypad with integrated mains switch 8 Filling level indicator 9 Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Image: class is required 1 <			· · ·	
IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (*C) 2 Temperature stability (*C) 3 Cooling Capacity (KW) At 20°C 6 Pump capacity flow rate (I/min) 7 Pump capacity flow pressure (bar) 0.35 34.5 9 Digital interface R5232 10 10 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES 1 Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Vorking temperature range (*C) -20 40 2 Temperature stability (*C) ±0.5				
ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20				
Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (*C) -20 40 2 Temperature stability (*C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-10°C 0.15KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25W,At 10°C 0.25W,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.25KW,At 0°C 0.25KW,At-20°C 0.35KW,At 0°C 0.25KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25KW				
Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (*C) -20 40 2 Temperature stability (*C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-10°C 0.15KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25W,At 10°C 0.25W,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.25KW,At 0°C 0.25KW,At-20°C 0.35KW,At 0°C 0.25KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25KW,At-20°C 0.25KW,At 0°C 0.25KW,At 0°C 0.25KW				
Specification of Recirculating Cooler 1 Working temperature range (°C) 20 40 2 Temperature stability (°C) 40.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At 10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.1KW, 6 Pump capacity flow rate (I/min) 15 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface RS232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Filling level indicator Permissible temperature in return line +80°C No side vents is required Interface I No side vents is required I Fubigs , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) 40.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At				
1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.25KW,At-10°C 0.25KW,At-10°C 0.25KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.25KW,At-10°C 0.25KW,At-10°C 0.25KW,At-10°C 0.15KW,At-20°C 0.25KW,At-20°C 0.25KW,At-			rubings, insulation should be quoted separately for a	a wieter length
1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.25KW,At-10°C 0.25KW,At-10°C 0.25KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.25KW,At-10°C 0.25KW,At-10°C 0.25KW,At-10°C 0.15KW,At-20°C 0.25KW,At-20°C 0.25KW,At-			Specification of Recirculating Cooler	
2 Temperature stability (*C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.1KW, 6 Pump capacity flow pressure (bar) 0.35 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface R5232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C +80°C No side vents is required IP class according to IEC 60529: 21 M ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At				
2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.1KW, 6 Pump capacity flow pressure (bar) 0.35 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface R5232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At		1	Working temperature range (°C)	-20 40
3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.15KW,At-20°C 0.11KW, 6 Pump capacity flow rete (I/min) 15 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface R5232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 Image: Color of the close color close color close color color close color color close color clos		2		±0.5
0°C 0°C 0.2 KW, At-10°C 0.15KW, At-20°C 0.15KW, At-20°C 0.1KW, 0.35 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface RS232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES ACCESSORIES 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.3KW,At 0°C 0.3KW,At 0°C 0.3KW,At 0°C 0.3KW,At 0°C 0.3KW,At 10°C 0.15KW,At-20°C				
0.15KW,At-20°C 0.15KW,At-20°C 0.15KW, 6 Pump capacity flow rate (l/min) 7 Pump capacity flow pressure (bar) 8 Filling volume liters 9 Digital interface 9 Digital interface 9 Digital interface 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length ACCESSORIES 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C				
Image: space state of the				
6 Pump capacity flow rate (l/min) 15 7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface R5232 10 Dimensions Wx L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Image: Condenser Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C				
7 Pump capacity flow pressure (bar) 0.35 8 Filling volume liters 34.5 9 Digital interface RS232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length 1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.3KW,At 0°C 0.3KW,At 10°C 0.15KW,At-20°C		6	Pump capacity flow rate (I/min)	
8 Filling volume liters 34.5 9 Digital interface RS232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required Image: Condense is required IP class according to IEC 60529: 21 Image: Condense is required ACCESSORIES Image: Condense is required Image: Condense is in the image: Condense: Condense is in the image: Condense is in t		7		0.35
9 Digital interface RS232 10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features		8		34.5
10 Dimensions W x L x H (cm) 23-25 x48 50 x 58-60 11 Weight (kg) Less than 39Kg Desirable Features		9	Digital interface	RS232
11 Weight (kg) Less than 39Kg Desirable Features		10		23-25 x48 50 x 58-60
Desirable Features Splash-proof keypad with integrated mains switch Removable venting grid for cleaning of the condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW)		11		Less than 39Kg
Image: switch Image: switch Removable venting grid for cleaning of the condenser Image: switch Image: switch Filling level indicator Image: switch Permissible temperature in return line +80°C Image: switch Permissible temperature in return line +80°C Image: switch Image: switch Image: switch Permissible temperature in return line +80°C Image: switch Image: switch Image: switch Image: switch swi				
Image: switch Image: switch Removable venting grid for cleaning of the condenser Image: switch Image: switch Filling level indicator Image: switch Permissible temperature in return line +80°C Image: switch Permissible temperature in return line +80°C Image: switch Image: switch Image: switch Permissible temperature in return line +80°C Image: switch Image: switch Image: switch Image: switch swi				
condenser Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C				
Filling level indicator Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 IP class according to IEC 60529: 21 ACCESSORIES Image: Constraint of the second seco			Removable venting grid for cleaning of the	
Permissible temperature in return line +80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-20°C			condenser	
+80°C No side vents is required IP class according to IEC 60529: 21 ACCESSORIES ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C			Filling level indicator	
IP class according to IEC 60529: 21 ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler Working temperature range (°C) 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-20°C				
ACCESSORIES ACCESSORIES Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C			No side vents is required	
Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-20°C			IP class according to IEC 60529: 21	
Tubings , Insulation should be quoted separately for 3 Meter length Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.15KW,At-20°C				
Specification of Recirculating Cooler 1 Working temperature range (°C) 2 Temperature stability (°C) 3 Cooling Capacity (KW) At 20°C 0°C 0.2KW,At-10°C 0.15KW,At-20°C			ACCESSORIES	
1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C			Tubings , Insulation should be quoted separately for 3	3 Meter length
1 Working temperature range (°C) -20 40 2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C			Specification of Recirculating Cooler	
2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C				
2 Temperature stability (°C) ±0.5 3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C		1	Working temperature range (°C)	-20 40
3 Cooling Capacity (KW) At 20°C 0.3KW,At 0°C 0.2KW,At-10°C 0.15KW,At-20°C				
0°C 0.2KW,At-10°C 0.15KW,At-20°C				
0.15KW,At-20°C				
6 Pump capacity flow rate (I/min) 15		6	Pump capacity flow rate (I/min)	15

7	Pump capacity flow pressure (bar)	0.35
8	Filling volume liters	34.5
9	Digital interface	RS232
10	Dimensions W x L x H (cm)	23-25 x48 50 x 58-60
11	Weight (kg)	Less than 39Kg
	Desirable Features	
	Splash-proof keypad with integrated mains	
	switch	
	Removable venting grid for cleaning of the	
	condenser	
	Filling level indicator	
	Permissible temperature in return line	
	+80°C	
	No side vents is required	
	IP class according to IEC 60529: 21	
	ACCESSORIES	
	Tubings , Insulation should be quoted separately f	or 3 Meter length

Sl. No.	List of optional items
	NA

A complete set of tender documents* may be Download by prospective bidder free of cost from the website <u>http://eprocure.gov.in/epublish/app</u>. Bidder has to make payment of requisite fees (i.e. Tender fees (if any) and EMD) online through RTGS/NEFT only.

Sl. No.	Specification
1.	Due date : The tender has to be submitted off-line before the due date. The offers received after
	the due date and time will not be considered.
2.	Preparation of Bids: The offer/bid should be submitted in two bid systems (i.e.) Technical bid
	and financial bid. The technical bid should consist of all technical details along with commercial
	terms and conditions. Financial bid should indicate item wise price for the items mentioned in the
	technical bid.
3.	EMD (if applicable): The tenderer should submit an EMD amount through RTGS/NEFT. The
	Technical Bid without EMD would be considered as UNRESPONSIVE and will not be accepted.
	The EMD will be refunded without any interest to the unsuccessful bidders after the award of
	contract. Refer to Schedule (at page 1 of this document) for its actual place of submission.
4.	Refund of EMD: The EMD will be returned to unsuccessful Tenderer only after the Tenders are
	finalized. In case of successful Tenderer, it will be retained till the successful and complete
	installation of the equipment.
5.	Opening of the tender : The offline bid will be opened by a committee duly constituted for this
	purpose. Offline bids (complete in all respect) received along with UTR info of EMD fund
	transfer (if any) will be opened as mentioned at "Annexure: Schedule" in presence of bidders
	representative if available. Only one representative will be allowed to participate in the tender
	opening. Bid received without EMD (if present) will be rejected straight way. The technical bid
	will be opened offline first and it will be examined by a technical committee (as per specification
	and requirement). The financial offer/bid will be opened only for the offer/bid which technically
	meets all requirements as per the specification, and will be opened in the presence of the vendor's
	representatives subsequently for further evaluation. The bidders if interested may participate on
	the tender opening Date and Time. The bidder should produce authorization letter from their
	company to participate in the tender opening.
6.	Acceptance/ Rejection of bids: The Committee reserves the right to reject any or all offers
_	without assigning any reason.
7.	Pre-qualification criteria:
	(i) Bidders should be the manufacturer / authorized dealer. Letter of Authorization from original
	equipment manufacturer (OEM) on the same and specific to the tender should be enclosed.
	(ii) An undertaking from the OEM is required stating that they would facilitate the bidder on a
	regular basis with technology/product updates and extend support for the warranty as well. (Ref.
	Annexure-II)
	(iii) OEM should be Nationally/Internationally reputed Company.
	(iv) Non-compliance of tender terms, non-submission of required documents, lack of clarity of the
	specifications, contradiction between bidder specification and supporting documents etc. may lead
	to rejection of the bid.
	(v) In the tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEM itself
	can bid but both cannot bid simultaneously for the same item/product in the same tender.
	(vi) If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid
0	on behalf of another Principal/OEM in the same tender for the same item/product.
8.	Performance Security : The supplier shall require to submit the performance security in the form
	of irrevocable bank guarantee issued by any Indian Nationalized Bank for an amount which is
	stated at page #1 of the tender document within 21 days from the date of receipt of the purchase
	order/LC and should be kept valid for a period of 60 days beyond the date of completion of
0	warranty period.
9.	Force Majeure: The Supplier shall not be liable for forfeiture of its performance security,
	liquidated damages or termination for default, if and to the extent that, it's delay in performance or

	other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
	• For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
	• If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
10.	Risk Purchase Clause : In event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from the other source on the total risk of the supplier under risk purchase clause.
11.	Packing Instructions : Each package will be marked on three sides with proper paint/indelible ink,
11.	the following:
	i. Item Nomenclature
	ii. Order/Contract No.
	iii. Country of Origin of Goods
	iv. Supplier's Name and Address
	v. Consignee details
	vi. Packing list reference number
12.	Delivery and Documents:
	Delivery of the goods should be made within a maximum of 08 to 16 weeks from the date of placement of purchase order. Within 24 hours of shipment, the supplier shall notify the purchaser and the insurance company by cable/telex/fax/e mail the full details of the shipment including contract number, railway receipt number/ AAP etc. and date, description of goods, quantity, name of the consignee, invoice etc. The supplier shall mail the following documents to the purchaser with
	a copy to the insurance company:1. 4 Copies of the Supplier invoice showing contract number, goods' description, quantity
	2. unit price, total amount;
	3. Insurance Certificate if applicable;
	4. Manufacturer's/Supplier's warranty certificate;
	5. Inspection Certificate issued by the nominated inspection agency, if any
	6. Supplier's factory inspection report; and
	7. Certificate of Origin (if possible by the beneficiary);
	 8. Two copies of the packing list identifying the contents of each package. 9. The above documents should be received by the Purchaser before arrival of the Goods (except
	where the Goods have been delivered directly to the Consignee with all documents) and, if not received, the Supplier will be responsible for any consequent expenses.
13.	Delayed delivery: If the delivery is not made within the due date for any reason, the Committee
15.	will have the right to impose penalty 1% per week and the maximum deduction is 10% of the contract value / price.
14.	Prices : The price should be quoted in net per unit (after breakup) and must include all packing and delivery charges. The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. However the percentage of taxes & duties shall be clearly indicated. The price should be quoted without custom duty and excise duty, since IIT Delhi is exempted from payment of Excise Duty and is eligible for concessional rate of custom duty. Necessary certificate will be issued on demand.

	Le serve of immente the main should be maded on FOD Davis calls. Under an early
	In case of imports, the price should be quoted on FOB Basis only. Under special
	circumstances (eg. perishable chemicals), when the item is imported on CIF/CIP, please indicate
	CIF/CIP charges separately upto IIT Delhi indicating the mode of shipment. IIT Delhi will make
	necessary arrangements for the clearance of imported goods at the Airport/Seaport. Hence the
	price should not include the above charges.
15.	Progress of Supply: Wherever applicable, supplier shall regularly intimate progress of supply, in
	writing, to the Purchaser as under:
	1. Quantity offered for inspection and date;
	2. Quantity accepted/rejected by inspecting agency and date;
	3. Quantity dispatched/delivered to consignees and date;
	4. Quantity where incidental services have been satisfactorily completed with date;
	5. Quantity where rectification/repair/replacement effected/completed on receipt of any
	communication from consignee/Purchaser with date;
	6. Date of completion of entire Contract including incidental services, if any; and
	7. Date of receipt of entire payments under the Contract (In case of stage-wise inspection, details
	required may also be specified).
16.	Inspection and Tests: Inspection and tests prior to shipment of Goods and at final acceptance are as
	follows:
	• After the goods are manufactured and assembled, inspection and testing of the goods shall be
	carried out at the supplier's plant by the supplier, prior to shipment to check whether the goods
	are in conformity with the technical specifications attached to the purchase order.
	Manufacturer's test certificate with data sheet shall be issued to this effect and submitted along
	with the delivery documents. The purchaser shall be present at the supplier's premises during
	such inspection and testing if need is felt. The location where the inspection is required to be
	conducted should be clearly indicated. The supplier shall inform the purchaser about the site
	preparation, if any, needed for installation of the goods at the purchaser's site at the time of
	submission of order acceptance.
	• The acceptance test will be conducted by the Purchaser, their consultant or other such person
	nominated by the Purchaser at its option after the equipment is installed at purchaser's site in the
	presence of supplier's representatives. The acceptance will involve trouble free operation and
	ascertaining conformity with the ordered specifications and quality. There shall not be any
	additional charges for carrying out acceptance test. No malfunction, partial or complete failure
	of any part of the equipment is expected to occur. The Supplier shall maintain necessary log in
	respect of the result of the test to establish to the entire satisfaction of the Purchaser, the
	successful completion of the test specified.
	• In the event of the ordered item failing to pass the acceptance test, a period not exceeding one
	weeks will be given to rectify the defects and clear the acceptance test, failing which the
	Purchaser reserve the right to get the equipment replaced by the Supplier at no extra cost to the
	Purchaser.
	• Successful conduct and conclusion of the acceptance test for the installed goods and equipment
	shall also be the responsibility and at the cost of the Supplier.
17.	Resolution of Disputes : The dispute resolution mechanism to be applied pursuant shall be as
	follows:
	• In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to
	any matter arising out of or connected with this agreement, such disputes or difference shall be
	settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules there under
	and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings.
	The dispute shall be referred to the Director, Indian Institute of Technology (IIT) Delhi and if he is
	unable or unwilling to act, to the sole arbitration of some other person appointed by him willing to
	act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and

	-
	binding on all parties to this order.
	• In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled
	by arbitration in accordance with provision of sub-clause (a) above. But if this is not acceptable to
	the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United
	Nations Commission on International Trade Law) Arbitration Rules.
	• The venue of the arbitration shall be the place from where the order is issued.
18.	• The venue of the abbitration shall be the place from where the order is issued. • Applicable Law: The place of jurisdiction would be New Delhi (Delhi) INDIA.
19.	Right to Use Defective Goods
19.	If after delivery, acceptance and installation and within the guarantee and warranty period, the
	operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to
	continue to operate or use such goods until rectifications of defects, errors or omissions by repair or
20	by partial or complete replacement is made without interfering with the Purchaser's operation.
20.	Supplier Integrity
	The Supplier is responsible for and obliged to conduct all contracted activities in accordance with
	the Contract using state of the art methods and economic principles and exercising all means
	available to achieve the performance specified in the contract.
21.	Training
	The Supplier is required to provide training to the designated Purchaser's technical and end user
	personnel to enable them to effectively operate the total equipment.
22.	Installation & Demonstration
	The supplier is required to done the installation and demonstration of the equipment within one
	month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be
	the same as per the supply of materials.
	In case of any mishappening/damage to equipment and supplies during the carriage of supplies
	from the origin of equipment to the installation site, the supplier has to replace it with new
	equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance
	company as per his convenience. IITD will not be liable to any type of losses in any form.
23.	Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the
201	supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse"
	(final destinations) on "All Risks" basis including War Risks and Strikes. The insurance shall be
	valid for a period of not less than 3 months after installation and commissioning. In case of orders
	placed on FOB/FCA basis, the purchaser shall arrange Insurance. If orders placed on
24.	CIF/CIP basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include:
24.	
	• Furnishing of 01 set of detailed operations & maintenance manual.
	• Arranging the shifting/moving of the item to their location of final installation within IITD
	premises at the cost of Supplier through their Indian representatives.
25.	Warranty:
	(i) Warranty period shall be (as stated at page #2 of this tender) from date of installation of Goods
	at the IITD site of installation. The Supplier shall, in addition, comply with the performance
	and/or consumption guarantees specified under the contract. If for reasons attributable to the
	Supplier, these guarantees are not attained in whole or in part, the Supplier shall at its
	discretion make such changes, modifications, and/or additions to the Goods or any part thereof
	as may be necessary in order to attain the contractual guarantees specified in the Contract at its
	own cost and expense and to carry out further performance tests. The warranty should be
	comprehensive on site.
	(ii) The Purchaser shall promptly notify the Supplier in writing of any claims arising under this
	warranty. Upon receipt of such notice, the Supplier shall immediately within in 02 days

	 arrange to repair or replace the defective goods or parts thereof free of cost at the ultimate destination. The Supplier shall take over the replaced parts/goods at the time of their replacement. No claim whatsoever shall lie on the Purchaser for the replaced parts/goods thereafter. The period for correction of defects in the warranty period is 02 days. If the supplier having been notified fails to remedy the defects within 02 days, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expenses and without prejudice to any other rights, which the purchaser may have against the supplier under the contract. (iii) The warranty period should be clearly mentioned. The maintenance charges (AMC) under different schemes after the expiry of the warranty should also be mentioned. The comprehensive warranty will commence from the date of the satisfactory installation/commissioning of the equipment against the defect of any manufacturing, workmanship and poor quality of the components. (iv) After the warranty period is over, Annual Maintenance Contract (AMC)/Comprehensive
	Maintenance Contract (CMC) up to next two years should be started. The AMC/CMC
26	charges will not be included in computing the total cost of the equipment.
26.	Governing Language The contract shall be written in English language. English language version of the Contract shall govern its interpretation. All correspondence and other documents pertaining to the Contract, which are exchanged by the parties, shall be written in the same language.
27.	Applicable Law
	The Contract shall be interpreted in accordance with the laws of the Union of India and all disputes
- 20	shall be subject to place of jurisdiction.
28.	Notices
	• Any notice given by one party to the other pursuant to this contract/order shall be sent to the other party in writing or by cable, telex, FAX or e mail and confirmed in writing to the other party's address.
	• A notice shall be effective when delivered or on the notice's effective date, whichever is later.
29.	Taxes
	Suppliers shall be entirely responsible for all taxes, duties, license fees, octroi, road permits, etc., incurred until delivery of the contracted Goods to the Purchaser. However, VAT in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order.
30.	Duties
	IIT Delhi is exempted from paying custom duty under notification No.51/96 (partially or full) and necessary "Custom Duty Exemption Certificate" can be issued after providing following information and Custom Duty Exemption Certificate will be issued to the shipment in the name of the Institute, no certificate will be issued to third party: a) Shipping details i.e. Master Airway Bill No. and House Airway No. (if exists)
	b) Forwarder details i.e. Name, Contact No., etc.
	IIT Delhi is exempted from paying Excise Duty and necessary Excise Duty Exemption Certificate
	will be provided for which following information are required.
	b) Quotation with details of Basic Price, Rate, Tax & Amount on which ED is applicable
	c) Supply Order Copy
21	d) Proforma-Invoice Copy.
31.	Payment: i. For Indigenous supplies, 100% payment shall be made by the Purchaser against delivery,
	inspection, successful installation, commissioning and acceptance of the equipment at IITD in

good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in Clause 9 of tender terms and conditions.

ii.	For imported items Payment will be made through irrevocable Letter of Credit (LC). Letter of
	Credit (LC) will be established in favour of foreign Supplier after the submission of
	performance security. The letter of credit (LC) will be established on the exchange rates as
	applicable on the date of establishment. For Imports, LC will be opened for 100% FOB/CIF
	value. 80% of the LC amount shall be released on presentation of complete and clear shipping
	documents and 20% of the LC amount shall be released after the installation and demonstration
	of the equipment at the INST site of installation in faultless working condition for period of 60
	days from the date of the satisfactory installation and subject to the production of unconditional
	performance bank guarantee as specified in Clause 8 of tender terms and conditions.

- iii. Indian Agency commission (IAC), if any shall be paid after satisfactory installation & commissioning of the goods at the destination at the exchange rate prevailing on the date of negotiation of LC documents, subject to DGS&D registration for restricted items.
- iv. All the bank charges within India will be borne by the Institute and outside India will be borne by the Supplier.

32. **User list:** Brochure detailing technical specifications and performance, list of industrial and educational establishments where the items enquired have been supplied must be provided. (Ref. Annexure-III)

- Manuals and Drawings
 (i) Before the goods and equipment are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.
 - (ii) The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract.
 - (iii) Unless and otherwise agreed, the goods equipment shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser.
- 36. **Application Specialist**: The Tenderer should mention in the Techno-Commercial bid the availability and names of Application Specialist and Service Engineers in the nearest regional office. (Ref. to Annexure-III)
- 37. **Site Preparation**: The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter.

The supplier shall visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre-installation requirements.

38. Spare Parts

33.

The Supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

- ii. Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this election shall not relieve the Supplier of any warranty obligations under the Contract; and
- iii. In the event of termination of production of the spare parts:
- iv. Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements; and
- v. Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if requested.
- Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the

	Goods, such as gaskets, plugs, washers, belts etc. Other spare parts and components shall be supplied as promptly as possible but in any case within six months of placement of order.
39.	Defective Equipment : If any of the equipment supplied by the Tenderer is found to be substandard, refurbished, un-merchantable or not in accordance with the description/specification or otherwise faulty, the committee will have the right to reject the equipment or its part. The prices of such equipment shall be refunded by the Tenderer with 18% interest if such payments for such equipment have already been made. All damaged or unapproved goods shall be returned at suppliers cost and risk and the incidental expenses incurred thereon shall be recovered from the supplier. Defective part in equipment, if found before installation and/or during warranty period, shall be replaced within 45 days on receipt of the intimation from this office at the cost and risk of supplier including all other charges. In case supplier fails to replace above item as per above terms & conditions, IIT Delhi may consider "Banning" the supplier.
40.	Termination for Default
	 The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default sent to the Supplier, terminate the Contract in whole or part: i. If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the order, or within any extension thereof granted by the Purchaser; or ii If the Supplier fails to perform any other obligation(s) under the Contract. iii If the Supplier, in the judgment of the Purchaser has engaged in corrupt or fraudulent practices in competing for or in executing the Contract.
	 For the purpose of this Clause: "Corrupt practice" means the offering, giving, receiving or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution. "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Borrower, and includes collusive practice among Bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the Borrower of the benefits of free and open competition;"
	• In the event the Purchaser terminates the Contract in whole or in part, the Purchaser may procure, upon such terms and in such manner, as it deems appropriate, Goods or Services similar to those undelivered, and the Supplier shall be liable to the Purchaser for any excess costs for such similar Goods or Services. However, the Supplier shall continue the performance of the Contract to the extent not terminated.
41.	Shifting : After 1-2 years once our new Academic Block will be ready, the supplier has to shift and reinstall the instrument free of cost (if required).
42.	Downtime: During the warranty period not more than 5% downtime will be permissible. For every day exceeding permissible downtime, penalty of 1/365 of the 5% FOB value will be imposed. Downtime will be counted from the date and time of the filing of complaint with in the business hours.
43.	Training of Personnel: The supplier shall be required to undertake to provide the technical training to the personnel involved in the use of the equipment at the Institute premises, immediately after completing the installation of the equipment for a minimum period of one week at the supplier's cost.
44.	Disputes and Jurisdiction: Any legal disputes arising out of any breach of contract pertaining to
15	this tender shall be settled in the court of competent jurisdiction located within New Delhi.
45.	Compliancy certificate : This certificate must be provided indicating conformity to the technical specifications. (Annexure-I)
46.	"In case of CIF/CIP shipments, kindly provide the shipment information atleast 2 days in advance

before landing the shipment alongwith the documents i.e. invoice, packing list, forwarder Name,								
address, contact No. in India to save demurrage charges (imposed by Indian Customs) .								
Otherwise these charges will be recovered from the supplier/Indian Agent."								

COMPLIANCE SHEET

TECHNICAL SPECIFICATION

Sl.	Technical Specifications	Compliance
No.		Y/N

I have also enclosed all relevant documents in support of my claims, (as above) in the following pages.

Signature of Bidder

Name: _____ Designation: _____

Organization Name: _____

Contact No. : _____

<< Organization Letter Head >> DECLARATION SHEET

We, _______ hereby certify that all the information and data furnished by our organization with regard to this tender specification are true and complete to the best of our knowledge. I have gone through the specification, conditions and stipulations in details and agree to comply with the requirements and intent of specification.

This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certified that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty.

The prices quoted in the financial bids are subsidized due to academic discount given to IIT Delhi.

We, further specifically certify that our	NAME & ADDRESS OF
organization has not been Black Listed/De	THE Vendor/ Manufacturer / Agent
Listed or put to any Holiday by any	
Institutional Agency/ Govt. Department/	
Public Sector Undertaking in the last three	
years.	
1 Phone	
2 Fax	
3 E-mail	
4 Contact Person Name	
5 Mobile Number	
6 TIN Number	
7 PAN Number	
(In case of on-line payment of Tender	
Fees)	
8 UTR No. (For Tender Fee)	
(In case of on-line payment of EMD)	
9 UTR No. (For EMD)	

(Signature of the Tenderer)

Name:

Seal of the Company

List of Govt. Organization/Deptt.

List of Government Organizations for whom the Bidder has undertaken such work during last three years (must be supported with work orders)							
Name of the organization	Name of Contact Person	Contact No.					

Name of application specialist / Service Engineer who have the technical competency to handle and support the quoted product during the warranty period.								
Name of the organizationName of Contact PersonContact No.								

Signature of Bidder

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

Bid Submission

Offline Bid Submission:

The Offline bids (complete in all respect) must be sealed in two Envelops as explained below:-

Envelope – 1 (Following documents to be provided as single PDF file)										
Sl. No.	Sl. No. Documents Content									
1.	Technical Bid	Compliance Sheet as per Annexure - I	.PDF							
2.		Organization Declaration Sheet as per Annexure - II	.PDF							
3.		List of organizations/ clients where the same products have been supplied (in last two years) along with their contact number(s). (Annexure-III)	.PDF							
4.		Technical supporting documents in support of all claims made at Annexure-I (Annexure-IV)	.PDF							
	Envelope – 2									
Sl. No.	Sl. No. TYPES Content									
1.	Financial Bid	Price bid should be submitted in PDF format.	.PDF							

<Department/Centre Name> Indian Institute of Technology Delhi Hauz Khas, New Delhi-110016

Date: XX/XX/XXXX

Subject: Purchase of <Item>

S. No.	Currency	Description of Item &	Qty.	Unit	Agency	Discount	Ex-works	Packing +	FOB	Insurance	CIF Price
		Specification	in	Price	Commission		price	Handling	Price	+ Frieght	(f+g)
			Units				(d=a+b-c)	+ DOC $+$	(f=d+e)	(g)	
				(a)	(b)	(c)		Inland			
								Frieght			
								(e)			
1											

For indigenous items please quote as per following format.

S. No.	Description of Item &	Qty. in Units	Unit Price in	Excise Duty %	CST/VAT%	Octroi%	Total Price in
	Specification		Rs.				Rs.
1.							
2.							