

Environmental Engineering Laboratory
Department of Civil Engineering
Notice Inviting Quotation for “Stainless Steel Cyclone Separator”

Date: 21/07/2017

Sealed Quotations are invited for “**Stainless Steel cyclone Separator.**” as per the specifications described below on or before **04/08/2017, 17:00 Hours**. Two bids (separate Quotation for technical and Financial bids) put together in common outer envelope Subscribing “**Quotation for Stainless Steel Cyclone Separator**” should be addressed and submitted to

Dr. GAZALA HABIB

Block-4, Room No:#303, Environmental Engineering
Department of Civil Engineering, Indian Institute Of Technology Delhi
Contact: 011- 2659 - 1192
Hauz Khas, New Delhi – 110016

Technical Specification of “Stainless Steel Cyclone Separator”

Type	Quantity	Technical Specifications:
Stainless Steel Cyclone Separator	1	Stainless Steel cyclone Separator.
		Weight: Should be less than – 0.5kg
		Cutsie: 2.5µm
		Designed for Flow: 4 to 5Lpm
		Removable top for easy Cleaning.
		Available with straight arm or 90 degree Inlet
		Removable outlet can be configured in diffrenet directions
		Inlet tube: 8 mm OD Outlet tube: 8mm OD Thread
Remark: Product must be Certified by USEPA. Product shuld be design to meet USPEA Specification.		

NOTES:

- Comprehensive onsite 3 year Warranty.
- Charges of AMC (Post Warranty) for five year must be provided separately.
- The vendor will provide details of Organization/Institute where they are distributing last 5 – 10 Years.
- The vender should provide authorization Certificate from manufacturer.
- Payment in Foreign currency will be made by Letter of credit as per Institute’s rules.
- Delivery period should not be exceeds 4-6 weeks and should be clearly indicated in the quotation.
- The decision of Departmental purchase finalization committee for selection or rejection Of any bid will be final.

DR.GAZALA HABIB

Assistant Professor
Department of Civil Engineering

