

SCHOOL OF BIOLOGICAL SCIENCES
Indian Institute of Technology Delhi

February 22, 2012

NIQ for CO₂ INCUBATOR

- 1) Microprocessor controlled 170-200 Liters direct heated CO₂ Incubator with High Temperature Decontamination facility @ 120degC having temperature control from 4 deg. C above ambient to 50°C, with control accuracy $\pm 0.1^{\circ}\text{C}$.
- 2) It should have six-sided direct heating with fanless, gentle convection circulation to provide stable temperature control, excellent uniformity and rapid recovery with no over shoot.
- 3) It should have CO₂ control range from 0.2 to 20% with control accuracy and uniformity of $\pm 0.1\%$ and should have rapid recovery of at least 0.7% per minute.
- 4) It should have Infra-red (IR) CO₂ sensor with programmable auto-zero function provide superior accuracy & stability.
- 5) It should come with minimum 4 adjustable height shelves & humidity reservoir (removable) to achieve at least 95% RH.
- 6) It should have independent door heater eliminate condensation on inner door surface.
- 7) It should have digital display for set parameters with audio visual alarm.
- 8) It should have HEPA filter on CO₂ inlet.
- 9) It should have RS232 communication port.
- 10) It should have 25mm access port.
- 11) It should have non-volatile memory which must guarantee data integrity regardless of length of time or frequency of power interruption.
- 12) It should have following additional safety such as back-up microprocessor, separate over temperature cut-out, alarms setpoint reset automatically, password protection etc.
- 13) It should come with CO₂ gas cylinder with regulator.
- 14) It should be ISO 9001 & CE Certification.

General instructions:

1. Letter from the manufacturer specifically to quote for this tender is to be attached for authenticity of dealership/ agency and the dealer should be authorized service provider.
2. Vendor should get a fresh certificate directly from their product principal's clearly mentioning about warranty for two years of the equipment to be delivered. The lowest quotation however, does not depend upon the warranty period.
3. Validity of the quotation should be at least three months.
4. Vendors will do the installation and demonstration of the equipment at IIT Delhi premises without additional charges.
5. Taxes, terms and conditions should be clearly mentioned.
6. In the case the items are proprietary products of the company, a proprietary item certificate stating the same may be provided.

7. Specifications form should be similar to the given specifications sheet.
8. A compliance statement for required specifications should be attached.
9. Payment terms and conditions should be clearly mentioned. No advance payment is encouraged by IIT Delhi.
12. Firm **MUST** provide a compliance statement vis-à-vis specifications in a “tabular form” clearly stating the compliance and giving justification, if any supported by technical literature with clear reference of page number, paragraph or lines. This statement must be signed, with the company seal, by the tendered for its authenticity and acceptance that any incorrect or ambiguous information found submitted will result in disqualification of the tender. The quotation should be complete in all respects. (as per IIT-Delhi rules).

The Institute/ purchase committee has the right to accept or reject any bid or all quotations without assigning any reason whatsoever.

Sealed quotations in separate envelopes of Technical and Commercial bids kept in one sealed outer envelope (super-scribed “Quotation CO₂ incubator”) should be addressed to **Dr. Archana Chugh, Kusuma School of Biological Sciences, IIT-Delhi, Hauz Khas, New Delhi 110016** and should **reach the Kusuma School of Biological Sciences office, IIT-Delhi by 10:00 hrs on March 9, 2012.**