

NOTICE INVITING QUOTATION (NIQ)

Date: 15/11/2011

Sealed quotations on company letterhead are invited for the purchase of following items by IIT Delhi confirming to technical specifications and terms and conditions as given below:

Technical Specifications for the Purchase of Load Cells

Item	Quantity	Specifications
Load Cells (2 kN, 5 kN and 10 kN capacity)	2 Nos. each	1. Capable of measuring compressive and tensile forces including dynamic/cyclic loading. 2. Max. operating force: 120% 3. Nominal operating temperature range: -10°C to +70°C 4. Made of Stainless steel (body) 5. Nominal sensitivity: 1 mV/V 6. Connection cable of about 5 m length is required.

Price should be quoted to CIF New Delhi. (IIT Delhi)

The technical bid and price bid should be packed and sealed separately. Technical catalogue supporting technical specifications must be attached with the technical bid. If the technical bid and price bid are in a same envelope or the technical catalogue is not attached, then the bid would be treated as invalid.





Vendors are asked to quote for all capacities. However, depending on the budget, the order may be placed for some of the above capacity only and the number may also vary.

The last date for receiving the quotation is December 08, 2011 (5.30 PM).

The validity of quotation should be three months.

Sealed quotation should be submitted to:

Prof. K. S. Rao
Room No-223, Block-V (Soil Lab)
Civil Engineering Department
IIT Delhi, Hauz Khas, New Delhi-110016.


(Prof. K. S. Rao) ~~~~ (Dr. R. Ayothiraman)  (Dr. J. T. Shahu)  (Dr. T. Chackraborty)