

**NIQ for synthesized microwave source (MI00900/CEQPT)**

Feb 28, 2012

Quotes are invited for two microwave synthesized sources with the specifications:

1. Frequency range should cover : 0.1 to 18 GHz.
2. Switching time (to settle at any frequency after receiving command) should be < 1 millisecond.
3. Output power at least 5 dBm
4. Interface : USB or SPI ( or similar serial interface) .


General Terms and conditions:

1. Vendors are requested to send their quotes (technical specs and prices separately) to the office of Dr A Basu, Rm. III-319, C.A.R.E., I.I.T. Delhi-110016, before 5 pm, Mar 14, 2012.
2. Quote should contain terms of payment, and should be valid for at least 90 days.
3. Payment after delivery is preferred but L/C can be arranged

  
( Prof D.Chadha, EE, Chairperson )

  
( Dr M P Abegaonkar - CARE )

  
( Prof S K Koul -CARE)

  
( Dr A.Basu -CARE)