

Environmental Engineering Laboratory
Department of Civil Engineering
Notice Inviting Quotation for “DC operated Vacuum diaphragm pump”

Date: 24/04/2018

Sealed Quotations are invited for “**DC operated Vacuum diaphragm pump**” as per the specifications described below on or before **10/05/2018, 17:00 Hours**. Two bids (separate Quotation for technical and Financial bids) put together in common outer envelope Subscribing “**Quotation for DC operated Vacuum diaphragm pump**” should be addressed and submitted to

Dr. GAZALA HABIB

Block-4, Room No:#303, Environmental Engineering
Department of Civil Engineering, Indian Institute Of Technology Delhi
Contact: 011- 2659 - 1192
Hauz Khas, New Delhi – 110016

Technical Specification of “DC operated Vacuum diaphragm pump”

Type	Quantity	Technical Specifications:
DC operated Vacuum diaphragm pump	4	Motor: Brushless, Ball bearing, dual shaft, oil free operation
		Weight: 0.1-3.0 Kg, Portable
		Operating voltage: 12.0 V / 24.0 V DC
		Operating Temperature: 0-50 °C or higher
		Flow rate: 12 LPM or higher at a relative vacuum of 250 mm Hg

NOTES:

- Comprehensive onsite 1 year Warranty.
- Charges of AMC (Post Warranty) for five year must be provided separately.
- The vendor will provide details of Organization/Institute where they are distributing last 5 – 10 Years.
- The vender should provide authorization Certificate from manufacturer.
- Payment in Foreign currency will be made by Letter of credit as per Institute’s rules.
- Delivery period should not be exceeds 4-6 weeks and should be clearly indicated in the quotation.
- The decision of Departmental purchase finalization committee for selection or rejection Of any bid will be final.

DR.GAZALA HABIB

Assistant Professor
Department of Civil Engineering