

NOTICE INVITING QUOTATION (NIQ)

Date: 22.11.2019

Sealed quotations on company letterhead are invited for the purchase of following item by IIT Delhi confirming to technical specifications and terms and conditions as given below:

Technical Specification for the Purchase of Equipments

Items and its Description
LOADING FRAME FOR MOUNTING OF HYDRAULIC JACK/ACTUATOR
Specification: <ul style="list-style-type: none">• A base plate made up of MS sheet of minimum 20mm thickness suitably reinforced to accommodate tank of dimension 2.35m x 1m x 1m (L x W x H) and load capacity of 50kN.• Two threaded columns to fix jack/actuator crosshead and height adjustable from 0.5m to 1m manually.• Manual locking arrangement for crosshead.• Base has provision for shifting of loading pillars from centre to 500mm offset from centre.• Capacity: 50kN Vertical Clearance: 1.2m Cross head adjustment: 0.5m to 1.0m Horizontal Clearance: 1.1m

TERMS AND CONDITIONS:

- ❖ The technical bid and price bid should be packed and sealed separately. If the technical bid and price bid are in a same envelope, then the bid would be treated as invalid.
- ❖ Minimum three supply order of similar work of any government organisation in last five years should be attached along with the technical bid. If the supply order is not attached with technical bid, then bid would be treated as invalid.
- ❖ Cost should be on CIF at IIT Delhi, New Delhi basis. The rate quoted should be inclusive of installation, commissioning, training required and one year of maintenance. Service support should be provided at no additional cost for a period of at least 12 months.
- ❖ IIT Delhi reserves the right to accept or reject any or all quotations without assigning any reason.
- ❖ Warranty terms must be clearly stated in the bid.
- ❖ The last date for receiving the quotation is November 29, 2019 (5:30 PM).
- ❖ The validity of quotation should be three months.
- ❖ All the calibration charts of the instruments should be supplied at the time of instrument delivery.
- ❖ The maximum time for delivery of the instruments in running condition is 45 days from the date of placing the order.

Bappaditya Manna
(Dr. Bappaditya Manna) 22/11/2019
Associate Professor
Civil Engineering Department
Indian Institute of Technology Delhi
Hauz Khas, New Delhi-110016