Notice Inviting Quotation (E-Procurement mode) INDIAN INSTITUTE OF TECHNOLOGY DELHI HAUZ KHAS, NEW DELHI-110016

Dated: 07/06/2017

Open Tender Notice No.IITD/BCVL(SP-1039)/2017

Indian Institute of Technology Delhi is in the process of purchasing following item(s) as per details as given as under.

Details of the item	Soil moisture and temperature measurement system
Earnest Money Deposit to be submitted	Not Applicable
Warranty	3 Years

Tender Documents may be downloaded from Central Public Procurement Portal <u>http://eprocure.gov.in/eprocure/app</u>. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <u>http://eprocure.gov.in/eprocure/app</u>. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at 'Instructions for online Bid Submission'.

Tenderers can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, Click on "GO" button to view all IIT Delhi tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website <u>http://eprocure.gov.in/eprocure/app</u> as per the schedule given in the next page.

No manual bids will be accepted. All quotation (both Technical and Financial should be submitted in the E-procurement portal).

Schedule

Name of Organization	Indian Institute of Technology
Tender Type (Open/Limited/EOI/Auction/Single)	OPEN
Tender Category (Services/Goods/works)	Goods
Type/Form of Contract (Work/Supply/ Auction/Service/Buy/Empanelment/ Sell)	Supply
Product Category (Civil Works/Electrical Works/Fleet Management/ Computer Systems)	Civil Engineering Instrument
Source of Fund (Institute/Project)	Budget Code <u>Fabrication</u> / Project Code <u>RP03329G</u>
Is Multi Currency Allowed	YES
Payment Mode (Online/Offline)	Offline
Date of Issue/Publishing	07/06/2017 (12:00Hrs)
Document Download/Sale Start Date	07/06/2017 (12:00Hrs)
Document Download/Sale End Date	28/06/2017 (15:00Hrs)
Date for Pre-Bid Conference	None
Venue of Pre-Bid Conference	None
Last Date and Time for Uploading of Bids	28/06/2017 (15:00Hrs)
Date and Time of Opening of Technical Bids	30/06/2017 (15:00Hrs)
Tender Fee	None
EMD	Not applicable
No. of Covers (1/2/3/4)	02
Bid Validity days (180/120/90/60/30)	180 days
Address for Communication	Dr. Tanusree Chakraborty, Room V-202 Department of Civil Engineering, Indian Institute of Technology (IIT) Delhi, HauzKhas, New Delhi - 110016, India.
Contact No.	(+91-11) 2659 1268
Email Address	tanusree@civil.iitd.ac.in

Chairman Purchase Committee (Buyer Member)

Instructions for Online Bid Submission:

As per the directives of Department of Expenditure, this tender document has been published on the Central Public Procurement Portal (<u>URL:https://eprocure.gov.in/eprocure/app</u>). The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

More information useful for submitting online bids on the CPP Portal may be obtained at:

https://eprocure.gov.in/eprocure/app

REGISTRATION

- 1) Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal (URL:<u>https://eprocure.gov.in/eprocure/app</u>) by clicking on the link "Click here to Enroll". Enrolment on the CPP Portal is free of charge.
- 2) As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
- 3) Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- 4) Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate(Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.), with their profile.
- 5) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.
- 6) Bidder then logs in to the site through the secured log-in by entering their userID / password and the password of the DSC / eToken.

SEARCHING FOR TENDER DOCUMENTS

- There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.
- 2) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.
- 3) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

PREPARATION OF BIDS

- 1) Bidder should take into account any corrigendum published on the tender document before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the

bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.

- 3) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF formats. Bid documents may be scanned with 100 dpi with black and white option.
- 4) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Space" area available to them to upload such documents. These documents may be directly submitted from the "My Space" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

SUBMISSION OF BIDS

- 1) Bidder should log into the site well in advance for bid submission so that he/she upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- 2) The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.
- 3) Bidder has to select the payment option as "on-line" to pay the tender fee / EMD as applicable and enter details of the instrument. Whenever, EMD / Tender fees is sought, bidders need to pay the tender fee and EMD separately on-line through RTGS (Refer to Schedule, Page No.2).
- 4) A standard BoQ format has been provided with the tender document to be filled by all the bidders. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. Bidders are required to download the BoQ file, open it and complete the white colored (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.

OR

In some cases Financial Bids can be submitted in PDF format as well (in lieu of BOQ).

- 5) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- 6) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 7) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 8) Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- 9) Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.

ASSISTANCE TO BIDDERS

- 1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- 2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is 1800 233 7315.

General Instructions to the Bidders

- 1) The tenders will be received online through portal <u>https://eprocure.gov.in/eprocure/app</u>. In the Technical Bids, the bidders are required to upload all the documents in .pdf format.
- 2) Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/etoken in the company's name is a prerequisite for registration and participating in the bid submission activities through https://eprocure.gov.in/eprocure/app. Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site https://eprocure.gov.in/eprocure/app under the link "Information about DSC".
- 3) Tenderer are advised to follow the instructions provided in the 'Instructions to the Tenderer for the esubmission of the bids online through the Central Public Procurement Portal foe-Procurement at https://eprocure.gov.in/eprocure/app.

Department of Civil Engineering Indian Institute of Technology HauzKhas, New Delhi-110 016

NOTICE INVITING QUOTATIONS

Dated: 07/06/2017

Subject: Purchase of soil moisture and temperature measurement system

Invitation for Tender Offers

Indian Institute of Technology Delhi invites online Bids (Technical bid and Commercial bid) from eligible and experienced OEM (Original Equipment Manufacturer) OR OEM Authorized Dealer for **supply, fabrication, installation & integration of soil moisture and temperature measurement system** with three year possible warranty or minimum one year comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document, which is available on CPP Portal <u>https://eprocure.gov.in/eprocure/app</u>

TECHNICAL SPECIFICATION:

Item	Description
Soil Moi	sture and temperature measurement system
Ι	Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground leve and up to 15 meter depth.
II	Corrosion free extension cables for placing the sensors at different level below the ground.
III	 A high accuracy data monitoring system to receive all the data from sensors continuously and simultaneously, save and stored the data for long period and transfer all the data into a PC through suitable communication network. The data monitoring system includes: a) A data logger with a) high resolution display b) multiple number of sensor inpu channels for simultaneous data acquisition c) in built memory for long term data monitoring / storage d) multiple output channels to link communication network protocol to deliver continuous and multiple data output. b) Multiple communication interfaces protocol option to transfer the data from data logge to a PC or notebook. c) All the accessories which is required to support the data monitoring system. d) Suitable guidelines and manuals for total system.
IV	A suitable user-friendly, easy access software which will compatible with communication interface networks. The software should capable to display the transferred real-time as well a previous data in the form of tables, charts etc. and analyse the same to deliver outputs with graphical interface.

A complete set of tender documents* may be Download by prospective bidder free of cost from the website <u>http://eprocure.gov.in/eprocure/app</u>. Bidder has to make payment of requisite fees (i.e. Tender fees (if any) and EMD) online through RTGS/NEFT only.

Terms & Conditions Details

Sl.No.	Specification
1.	Due date : The tender has to be submitted before the due date. The offers received after the due date and time will not be considered. No manual bids will be considered.
2.	Preparation of Bids : The offer/bid should be submitted in two bid systems (i.e.) Technical bid and financial bid. The technical bid should consist of all technical details along with commercial terms and conditions. Financial bid should indicate item wise price for the items mentioned in the technical bid in the given format .Financial Bids to be submitted in PDF format.The Technical bid and the financial bid should be submitted Online.
3.	EMD (if applicable): The tenderer should submit an EMD amount through RTGS/NEFT. The Technical Bid without EMD would be considered as UNRESPONSIVE and will not be accepted. The EMD will be refunded without any interest to the unsuccessful bidders after the award of contract. Refer to Schedule (at page 1 of this document) for its actual place of submission.
4.	Refund of EMD : The EMD will be returned to unsuccessful Tenderer only after the Tenders are finalized. In case of successful Tenderer, it will be retained till the successful and complete installation of the equipment.
5.	Opening of the tender : The online bid will be opened by a committee duly constituted for this purpose. Online bids (complete in all respect) received along with EMD (if any) will be opened as mentioned at "Annexure: Schedule" in presence of bidders representative if available. Only one representative will be allowed to participate in the tender opening. Bid received without EMD (if present) will be rejected straight way. The technical bid will be opened online first and it will be examined by a technical committee (as per specification and requirement). The financial offer/bid will be opened only for the offer/bid which technically meets all requirements as per the specification, and will be opened in the presence of the vendor's representatives subsequently for further evaluation. The bidders if interested may participate on the tender opening Date and Time. The bidder should produce authorization letter from their company to participate in the tender opening.
6.	Acceptance/ Rejection of bids: The Committee reserves the right to reject any or all offers without assigning any reason.
7.	 Pre-qualification criteria: (i)Bidders should be the manufacturer / authorized dealer. Letter of Authorization from original equipment manufacturer (OEM) on the same and specific to the tender should be enclosed. (ii)An undertaking from the OEM is required stating that they would facilitate the bidder on a regular basis with technology/product updates and extend support for the warranty as well. (iii) OEM should be internationally reputed Branded Company. (iv) Non-compliance of tender terms, non-submission of required documents, lack of clarity of the specifications, contradiction between bidder specification and supporting documents etc. may lead to rejection of the bid.
	 (v) In the tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender. (vi) If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product. (vii) Certification is required regarding previous supply and fabrication of soil moisture and temperature measurement system to other government and non-government institutions. The vendor must provide contact addresses of their previous customers. The current customer will contact the respective institution for verification.

8.	Risk Purchase Clause: In event of failure of supply of the item/equipment within the stipulated
	delivery schedule, the purchaser has all the right to purchase the item/equipment from the other
	source on the total risk of the supplier under risk purchase clause.
9.	Packing Instructions: Each package will be marked on three sides with proper paint/indelible ink,
	the following:
	i. Item Nomenclature
	ii. Order/Contract No.
	iii. Country of Origin of Goods
	iv. Supplier's Name and Address
	v. Consignee details
	vi. Packing list reference number
10.	Delivery and Documents:
10	Delivery of the goods should be made within a maximum of 08 to 16 weeks from the date of
	placement of purchase order and the opening of LC. Within 24 hours of shipment, the supplier shall
	notify the purchaser and the insurance company if any by cable/telex/fax/e mail the full details of the
	shipment including contract number, railway/ shipment receipt number/ AAP etc. and date,
	description of goods, quantity, name of the consignee, invoice etc. The supplier shall mail the
	following documents to the purchaser with a copy to the insurance company or certified authority
	from IIT Delhi:
	1. 4 Copies of the Supplier invoice showing contract number, goods' description, quantity
	2. unit price, total amount;
	3. Insurance Certificate if applicable;
	4. Manufacturer's/Supplier's warranty certificate;
	5. Inspection Certificate issued by the nominated inspection agency, if any
	6. Supplier's factory inspection report; and
	7. Certificate of Origin (if possible by the beneficiary);
	8. Two copies of the packing list identifying the contents of each package.
	9. The above documents should be received by the Purchaser before arrival of the Goods (except
	where the Goods have been delivered directly to the Consignee with all documents) and, if not
	received, the Supplier will be responsible for any consequent expenses.
11.	Delayed delivery: If the delivery is not made within the due date for any reason, the Committee
	will have the right to impose penalty 1% per week and the maximum deduction is 10% of the
	contract value / price.
12.	
	and delivery charges. The offer/bid should be exclusive of taxes and duties, which will be paid by
	the purchaser as applicable. However the percentage of taxes & duties shall be clearly indicated.
	The price should be quoted without custom duty and excise duty, since IIT Delhi is exempted
	from payment of Excise Duty and is eligible for concessional rate of custom duty. Necessary
	certificate will be issued on demand.
	In case of imports, the price should be quoted on FOB/FCA origin Airport Basis only. Under
	special circumstances (eg. perishable chemicals), when the item is imported on CIF/CIP, please indicate CIF/CIP charges congretally unto UT Dalhi indicating the mode of chingment UT Dalhi
	indicate CIF/CIP charges separately upto IIT Delhi indicating the mode of shipment. IIT Delhi will make necessary arrangements for the clearance of imported goods at the Airport/Separate
	will make necessary arrangements for the clearance of imported goods at the Airport/Seaport. Hence the price should not include the above charges. At any circumstances, it is the
	responsibility of the foreign supplier to handover the material to our forwarder at the origin
	airport after completing all the inland clearing. No Ex- Works consignment will be
	entertained.
	"In case of CIF/CIP shipments, kindly provide the shipment information at least 2 days in

	advance before landing the shipment along with the documents i.e. invoice, packing list,
	forwarder Name, address, contact No. in India to save penalty/demurrage charges (imposed by
	Indian Customs). Otherwise these charges will be recovered from the supplier/Indian Agent."
13.	Notices: For the purpose of all notices, the following shall be the address of the Purchaser and
	Supplier.
	Purchaser: Dr. TanusreeChakraborty,
	Department of Civil Engineering,
	Indian Institute of Technology
	HauzKhas, New Delhi - 110016.
	Supplier: (To be filled in by the supplier)
	(All supplier's should submit its supplies information as per Annexure-II).
14.	Progress of Supply: Wherever applicable, supplier shall regularly intimate progress of supply, in
	writing, to the Purchaser as under:
	1. Quantity offered for inspection and date;
	2. Quantity accepted/rejected by inspecting agency and date;
	3. Quantity dispatched/delivered to consignees and date;
	4. Quantity where incidental services have been satisfactorily completed with date;
	5. Quantity where rectification/repair/replacement effected/completed on receipt of any
	communication from consignee/Purchaser with date;
	6. Date of completion of entire Contract including incidental services, if any; and
	7. Date of receipt of entire payments under the Contract (In case of stage-wise inspection, details
	required may also be specified).
15	Inspection and Tests: Inspection and tests prior to shipment of Goods and at final acceptance are
15.	as follows:
	• After the goods are manufactured and assembled, inspection and testing of the goods shall be
	carried out at the supplier's plant by the supplier, prior to shipment to check whether the goods
	are in conformity with the technical specifications attached to the purchase order.
	Manufacturer's test certificate with data sheet shall be issued to this effect and submitted along
	with the delivery documents. The purchaser shall be present at the supplier's premises during such inspection and testing if need is falt. The leastion where the inspection is required to be
	such inspection and testing if need is felt. The location where the inspection is required to be
	conducted should be clearly indicated. The supplier shall inform the purchaser about the site
	preparation, if any, needed for installation of the goods at the purchaser's site at the time of
	submission of order acceptance.
	• The acceptance test will be conducted by the Purchaser, their consultant or other such person
	nominated by the Purchaser at its option after the equipment is installed at purchaser's site in the
	presence of supplier's representatives. The acceptance will involve trouble free operation and
	ascertaining conformity with the ordered specifications and quality. There shall not be any
	additional charges for carrying out acceptance test. No malfunction, partial or complete failure of
	any part of the equipment is expected to occur. The Supplier shall maintain necessary log in
	respect of the result of the test to establish to the entire satisfaction of the Purchaser, the
	successful completion of the test specified.
	• In the event of the ordered item failing to pass the acceptance test, a period not exceeding one
	weeks will be given to rectify the defects and clear the acceptance test, failing which the
	Purchaser reserve the right to get the equipment replaced by the Supplier at no extra cost to the
	Purchaser.
	9

	• Successful conduct and conclusion of the acceptance test for the installed goods and equipment
16	shall also be the responsibility and at the cost of the Supplier.
16.	Resolution of Disputes : The dispute resolution mechanism to be applied pursuant shall be as follows:
	 In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Indian Arbitration& Conciliation Act, 1996, the rules there under and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Director, Indian Institute of Technology (IIT) Delhi and if he is unable or unwilling to act, to the sole arbitration of some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order. In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by arbitration in accordance with provision of sub-clause (a) above. But if this is not acceptable to the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules.
	• The venue of the arbitration shall be the place from where the order is issued.
17.	
18.	Right to Use Defective Goods If after delivery, acceptance and installation and within the guarantee and warranty period, the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.
19.	Supplier Integrity
	The Supplier is responsible for and obliged to conduct all contracted activities in accordance with the Contract using state of the art methods and economic principles and exercising all means available to achieve the performance specified in the contract.
20.	
	end user personnel to enable them to effectively operate the total equipment.
21.	Installation & Demonstration The supplier is required to done the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation or provide all necessary guidelines for installations. Otherwise the penalty clause will be the same as per the supply of materials.
	In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form.
22.	Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the Supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis including War Risks and Strikes. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. In case of orders placed on FOB/FCA basis, the purchaser shall arrange Insurance. If orders placed on CIF/CIP basis, the insurance should be up to IIT Delhi.
23.	Incidental services: The incidental services also include:
	• Furnishing of 01 set of detailed operations & maintenance manual.
	• Arranging the shifting/moving of the item to their location of final installation within IITD

r	
	premises at the cost of Supplier through their Indian representatives.
24.	Warranty:
	 Warranty period shall be (as stated at page #2 of this tender) from date of installation of Good at the IITD site of installation. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the contract. If for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall at its discretion make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own contant expense and to carry out further performance tests. The warranty should be comprehensive on site. Note: If a different period of warranty has been specified in the 'Technical Specifications' of the state of the sta
	 then the period mentioned above shall stand modified to that extent. 2. The Purchaser shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall immediately within in 02 days arrang to repair or replace the defective goods or parts thereof free of cost at the ultimate destination. The Supplier shall take over the replaced parts/goods at the time of their replacement. No claim whatsoever shall lie on the Purchaser for the replaced parts/goods thereafter. The period for correction of defects in the warranty period is 02 days. If the supplier having been notified fail to remedy the defects within 02 days, the purchaser may proceed to take such remedial action a may be necessary, at the supplier's risk and expenses and without prejudice to any other rights which the purchaser may have against the supplier under the contract. 3. The warranty period should be clearly mentioned. The maintenance charges (AMC) under different schemes after the expiry of the warranty should also be mentioned. The comprehensive warranty will commence from the date of the satisfactory installation/commissioning of the equipment against the defect of any manufacturing, workmanship and poor quality of the components. 4. After the warranty period is over, Annual Maintenance Contract (AMC)/Comprehensive
	Maintenance Contract (CMC) up to next two years should be started. The AMC/CMC charge
	will not be included in computing the total cost of the equipment.
25.	Governing Language The contract shall be written in English language. English language version of the Contract sha govern its interpretation. All correspondence and other documents pertaining to the Contract, whic are exchanged by the parties, shall be written in the same language.
26.	Applicable Law
20.	The Contract shall be interpreted in accordance with the laws of the Union of India and all dispute shall be subject to place of jurisdiction.
27.	Notices
	 Any notice given by one party to the other pursuant to this contract/order shall be sent to the other party in writing or by cable, telex, FAX or e mail and confirmed in writing to the other party's address. A notice shall be effective when delivered or on the notice's effective date, whichever a statement of the other party is address.
	later.
28.	Taxes Suppliers shall be entirely responsible for all taxes, duties, license fees, octroi, road permits, etc incurred until delivery of the contracted Goods to the Purchaser. However, VAT in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order.
29.	Duties IIT Delhi is exempted from paying custom duty under notification No.51/96 (partially or full) an
	necessary "Custom Duty Exemption Certificate" can be issued after providing following

	information and Custom Duty Exemption Certificate will be issued to the shipment in the name of
	the Institute, (no certificate will be issued to third party): The procured product should be used for
	teaching, scientific and research work only. a. Shipping details i.e. Master Airway Bill No. and House Airway No. (if exists)
	b. Forwarder details i.e. Name, Contact No., etc.
	b. Forwarder details i.e. Maine, Contact No., etc.
	IIT Delhi is exempted from paying Excise Duty and necessary Excise Duty Exemption Certificate will be provided for which following information are required.
	c. Quotation with details of Basic Price, Rate, Tax & Amount on which ED is applicable
	d. Supply Order Copy
	e. Proforma-Invoice Copy.
30.	Agency Commission : Agency commission if any will be paid to the Indian agent in Rupees on receipt of the equipment and after satisfactory installation. Agency Commission will not be paid in foreign currency under any circumstances. The details should be explicitly shown in Tender even in case of Nil commission. The tenderer should indicate the percentage of agency commission to be paid to the Indian agent.
31	Payment:
	 (i) For imported items Payment will be made through irrevocable Letter of Credit (LC) Cash Against Documents (CAD)/Against delivery/after satisfactory installation by T.T. Letter of Credit (LC) will be established in favour of foreign Supplier after the submission of
	performance security. The letter of credit (LC) will be established on the exchange rates as applicable on the date of establishment. For Imports, LC will be opened for 100% FOB/CIF value. 80% of the LC amount shall be released on presentation of complete and clear shipping documents and 20% of the LC amount shall be released after the installation and demonstration of the equipment at the INST site of installation in faultless working condition for period of 60 days from the date of the satisfactory installation and subject to the production of unconditional performance bank guarantee as specified in Clause 8 of tender terms and
	 conditions. (ii) For Indigenous supplies, 100% payment shall be made by the Purchaser against delivery, inspection, successful installation, commissioning and acceptance of the equipment at IITD in good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in Clause 9 of tender terms and
	conditions.
	 (iii) Indian Agency commission (IAC), if any shall be paid after satisfactory installation & commissioning of the goods at the destination at the exchange rate prevailing on the date of negotiation of LC documents, subject to DGS&D registration for restricted items. (iv) All the bank charges within India will be borne by the Institute and outside India will be
	borne by the Supplier.
32.	
33.	Manuals and Drawings
	(i) Before the goods and equipment are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.(ii) The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract.
24	(iii) Unless and otherwise agreed, the goods equipment shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser.
34.	Application Specialist: The Tenderer should mention in the Techno-Commercial bid the

	ailability and names of Application Specialist and Service Engineers in the nearest regional fice. (Ref. to Annexure-III)
	te Preparation : The supplier shall inform to the Institute about the site preparation, if any,
nee	eded for the installation of equipment, immediately after the receipt of the purchase order. The
	pplier must provide complete details regarding space and all the other infrastructural
rec	quirements needed for the equipment, which the Institute should arrange before the arrival of
the	e equipment to ensure its timely installation and smooth operation thereafter.
	e supplier may visit the Institute and see the site where the equipment is to be installed and may
off	fer his advice and render assistance to the Institute in the preparation of the site and other pre-
ins	stallation requirements.
36. Sp	are Parts
Th	e Supplier may be required to provide any or all of the following materials, notifications, and
inf	formation pertaining to spare parts manufactured or distributed by the Supplier:
i. i	Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this
	election shall not relieve the Supplier of any warranty obligations under the Contract; and
ii.	In the event of termination of production of the spare parts:
iii.	Advance notification to the Purchaser of the pending termination, in sufficient time to permit the
	Purchaser to procure needed requirements; and
iv.	Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and
:	specifications of the spare parts, if requested.
	pplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the
	oods, such as gaskets, plugs, washers, belts etc. Other spare parts and components shall be
	pplied as promptly as possible but in any case within six months of placement of order.
	efective Equipment: If any of the equipment supplied by the Tenderer is found to be
	bstandard, refurbished, un-merchantable or not in accordance with the description/specification
	otherwise faulty, the committee will have the right to reject the equipment or its part. The
-	ces of such equipment shall be refunded by the Tenderer with 18% interest if such payments
	such equipment have already been made. All damaged or unapproved goods shall be returned
	suppliers cost and risk and the incidental expenses incurred thereon shall be recovered from the
-	pplier. Defective part in equipment, if found before installation and/or during warranty period,
	all be replaced within 45 days on receipt of the intimation from this office at the cost and risk of
	pplier including all other charges. In case supplier fails to replace above item as per above terms
	conditions, IIT Delhi may consider "Banning" the supplier.
	ermination for Default
	e Purchaser may, without prejudice to any other remedy for breach of contract, by written notice
of	default sent to the Supplier, terminate the Contract in whole or part:
	i. If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the
	order, or within any extension thereof granted by the Purchaser; or
	ii. If the Supplier fails to perform any other obligation(s) under the Contract.
	iii. If the Supplier, in the judgment of the Purchaser has engaged in corrupt or fraudulent
	practices in competing for or in executing the Contract.
•	For the purpose of this Clause:
	i. "Corrupt practice" means the offering, giving, receiving or soliciting of anything of
	value to influence the action of a public official in the procurement process or in
	contract execution.
	ii. "Fraudulent practice" means a misrepresentation of facts in order to influence a
	procurement process or the execution of a contract to the detriment of the Borrower,
	and includes collusive practice among Bidders (prior to or after bid submission)
	designed to establish bid prices at artificial non-competitive levels and to deprive the
	Borrower of the benefits of free and open competition;"

	• In the event the Purchaser terminates the Contract in whole or in part, the Purchaser may procure, upon such terms and in such manner, as it deems appropriate, Goods or Services similar to those undelivered, and the Supplier shall be liable to the Purchaser for any excess costs for such similar Goods or Services. However, the Supplier shall continue the performance of the Contract to the extent not terminated.	
39.	39. Downtime: During the warranty period not more than 5% downtime will be permissible. F	
	every day exceeding permissible downtime, penalty of 1/365 of the 5% FOB value will be	
	imposed. Downtime will be counted from the date and time of the filing of complaint with in the	
	business hours.	
40.	0. Training of Personnel: The supplier shall be required to undertake to provide the technical	
	training to the personnel involved in the use of the equipment immediately after completing the	
	installation of the equipment for a minimum period of one week at the supplier's cost.	
41.	Disputes and Jurisdiction: Any legal disputes arising out of any breach of contract pertaining to	
	this tender shall be settled in the court of competent jurisdiction located within New Delhi.	
42.	Compliancy certificate: This certificate must be provided indicating conformity to the technical	
	specifications.	

COMPLIANCE SHEET

TECHNICAL SPECIFICATION

Item	Description	Compliance
Soil Moi	isture and temperature measurement system	Y/N
Ι	15 Nos Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.	
II	Corrosion free extension cables for placing the sensors at different level below the ground.	
Ш	 A high accuracy data monitoring system to receive all the data from sensors continuously and simultaneously, save and stored the data for long period and transfer all the data into a PC through suitable communication network. The data monitoring system includes: a) A data logger with a) high resolution display b) multiple number of sensor input channels for simultaneous data acquisition c) in built memory for long term data monitoring / storage d) multiple output channels to link communication network protocol to deliver continuous and multiple data output. b) Multiple communication interfaces protocol option to transfer the data from data logger to a PC or notebook. c) All the accessories which is required to support the data monitoring system. d) Suitable guidelines and manuals for total system. 	
IV	A suitable user-friendly, easy access software which will compatible with communication interface networks. The software should capable to display the transferred real-time as well as previous data in the form of tables, charts etc. and analyse the same to deliver outputs with graphical interface.	

I have also enclosed all relevant documents in support of my claims, (as above) in the following pages.

Signature of Bidder

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

<< Organization Letter Head >> DECLARATION SHEET

We, _______ hereby certify that all the information and data furnished by our organization with regard to this tender specification are true and complete to the best of our knowledge. I have gone through the specification, conditions and stipulations in details and agree to comply with the requirements and intent of specification.

This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certified that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty.

The prices quoted in the financial bids are subsidized due to academic discount given to IIT Delhi.

We, further specifically certify that our	NAME & ADDRESS OF
organization has not been Black Listed/De	THE Vendor/ Manufacturer / Agent
Listed or put to any Holiday by any Institutional	
Agency/ Govt. Department/ Public Sector	
Undertaking in the last three years.	
1 Phone	
2 Fax	
3 E-mail	
4 Contact Person Name	
5 Mobile Number	
6 TIN Number	
7 PAN Number	
(In case of on-line payment of Tender Fees)	
8 UTR No. (For Tender Fee)	
(In case of on-line payment of EMD)	
9 UTR No. (For EMD)	
10 Kindly provide bank details of the	
bidder in the following format:	
a) Name of the Bank	
b) Account Number	
c) Kindly attach scanned copy of one	
Cheque book page to enable us to return	
the EMD to unsuccessful bidder	

(Signature of the Tenderer)

Name:

Seal of the Company

List of Government Organizations for whom the Bidder has undertaken such work during last three years (must be supported with work orders)							
Name of the organization	Name of Contact Person	Contact No.					

List of Govt. Organization/Deptt.

Name of application specialist / Service Engineer who have the technical competency to handle and support the quoted product during the warranty period.								
Name of the organization	Name of Contact Person	Contact No.						

Signature of Bidder

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

Bid Submission

i. <u>Online Bid Submission :</u>

The Online bids (complete in all respect) must be uploaded online in **Two** Envelops as explained below:-

Sl. No.	Documents	Content	File Types				
1.	Technical Bid	Compliance Sheet as per Annexure - I	.PDF				
2.		Organization Declaration Sheet as per Annexure - II	.PDF				
3.		List of organizations/ clients where the same products have been supplied (in last two years) along with their contact number(s). (Annexure-III)	.PDF				
4.		Technical supporting documents in support of all claims made at Annexure-I (Annexure-IV)	.PDF				
	Envelope – 2						
Sl. No.	TYPES	Content					
1.	Financial Bid	Price bid should be submitted in PDF format.	.PDF				

Department of Civil Engineering Indian Institute of Technology Delhi Hauz Khas, New Delhi-110016

Date: 07/06/2017

Subject: Purchase of Soil moisture and temperature measurement system

U										
Currency	Description and Specification of the	Qty. in	Unit	Agency	Discount	Ex-works	Packing	FOB/F	Insurance	CIF
	Item	Units	Price	Commission		price	+	CAAirp	+ Frieght	Price
						(d=a+b-c)	Handling	ort	(g)	(f+g)
			(a)	(b)	(c)		+ DOC +	Price		
							Inland	(f=d+e)		
							Freight +			
							FCA			
							Charges			
							(e)			
	Soil moisture and temperature sensor									
	for measuring the moisture content and									
	temperature with high level accuracy.									
	The sensors need to be placed every 1									
	meter distance below ground level and									
	up to 15 meter depth.									
	Corrosion free extension cables for									
	placing the sensors at different level									
	below the ground.									
	A high accuracy data monitoring									
	system to receive all the data from									
	sensors continuously and									
	simultaneously, save and stored the									
	Currency	ItemItemSoil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Corrosion free extension cables for placing the sensors at different level below the ground.A high accuracy data monitoring system to receive all the data from sensors continuously and	ItemUnitsItemUnitsSoil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Corrosion free extension cables for placing the sensors at different level below the ground.A high accuracy data monitoring system to receive all the data from sensors continuously and	ItemUnitsPriceItemUnitsPrice(a)(a)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free sensors at different level below the ground.A high accuracy data monitoring system to receive all the data from sensors continuously andImage: Corrosion free sensors continuously and	ItemUnitsPriceCommission(a)(b)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Content of the sensor of the sensors at different level below the ground.Image: Content of the sensor of the sensors at different level below the ground.A high accuracy data monitoring system to receive all the data from sensors continuously andImage: Content of the sensor of the sens	ItemUnitsPriceCommission(a)(b)(c)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the data from sensors continuously andImage: Corrosion free extension cables for placing the data from sensors continuously andImage: Corrosion free extension cables for placing the data from <td>ItemUnitsPriceCommissionprice(a)(b)(c)(c)(d=a+b-c)(a)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.Image: Corrosion free extension cables for placing the sensors at different level below the ground.</br></td> <td>ItemUnitsPriceCommissionprice+(a)(b)(c)(c)price+Handling+ DOC +InlandFreight(b)(c)(c)(c)Soil moisture and temperature sensor(c)(c)(c)(c)for measuring the moisture content and(c)(c)(c)(c)temperature with high level accuracy.(c)(c)(c)(c)The sensors need to be placed every 1(c)(c)(c)(c)meter distance below ground level and(c)(c)(c)(c)Data(c)(c)(c)(c)(c)Corrosion free extension cables for placing the sensors at different level below the ground.(c)(c)(c)A high accuracy data monitoring system to receive all the data from sensors continuously and(c)(c)(c)Note(c)<t< td=""><td>ItemUnitsPriceCommission (a)price (b)+ (c)CAAirp Handling + DOC + Inland Freight + FCA Charges (e)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level sensors continuously andImage: Soil work with the sensors with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensor with the sensor work with the sensor w</td><td>ItemUnitsPriceCommissionprice+CAAirp+ Frieght(a)(b)(c)(c)<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ FCA \\ FCA \\ ChargesPrice$\begin{pmatrix} d=a+b-c \\ FCA \\ FCA$</math></math></math></math></math></math></math></math></math></math></math></math></math></math></math></math></td></t<></td>	ItemUnitsPriceCommissionprice(a)(b)(c)(c)(d=a+b-c)(a)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 	ItemUnitsPriceCommissionprice+(a)(b)(c)(c)price+Handling+ DOC +InlandFreight(b)(c)(c)(c)Soil moisture and temperature sensor(c)(c)(c)(c)for measuring the moisture content and(c)(c)(c)(c)temperature with high level accuracy.(c)(c)(c)(c)The sensors need to be placed every 1(c)(c)(c)(c)meter distance below ground level and(c)(c)(c)(c)Data(c)(c)(c)(c)(c)Corrosion free extension cables for placing the sensors at different level below the ground.(c)(c)(c)A high accuracy data monitoring system to receive all the data from sensors continuously and(c)(c)(c)Note(c) <t< td=""><td>ItemUnitsPriceCommission (a)price (b)+ (c)CAAirp Handling + DOC + Inland Freight + FCA Charges (e)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level sensors continuously andImage: Soil work with the sensors with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensor with the sensor work with the sensor w</td><td>ItemUnitsPriceCommissionprice+CAAirp+ Frieght(a)(b)(c)(c)<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ FCA \\ ChargesPrice<math>\begin{pmatrix} d=a+b-c \\ FCA \\ FCA \\ ChargesPrice$\begin{pmatrix} d=a+b-c \\ FCA \\ FCA$</math></math></math></math></math></math></math></math></math></math></math></math></math></math></math></math></td></t<>	ItemUnitsPriceCommission (a)price (b)+ (c)CAAirp Handling + DOC + Inland Freight + FCA Charges (e)Soil moisture and temperature sensor for measuring the moisture content and temperature with high level accuracy. The sensors need to be placed every 1 meter distance below ground level and up to 15 meter depth.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level below the ground.Image: Soil work with the sensors at different level sensors continuously andImage: Soil work with the sensors with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensors with the sensors at different level below the ground.Image: Soil work with the sensor with the sensor with the sensor work with the sensor w	ItemUnitsPriceCommissionprice+CAAirp+ Frieght(a)(b)(c)(c) $\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ hlandling \\ Frieght + \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ FCA \\ FCA \\ ChargesPrice\begin{pmatrix} d=a+b-c \\ FCA \\ FCA$

	I				T	ı
data for long period and transfer all the data into a PC through suitable						
communication network. The data						
monitoring system includes:						
 a) A data logger with a) high resolution display b) multiple number of sensor input channels for simultaneous data acquisition c) in built memory for long term data monitoring / storage d) multiple output channels to link communication network protocol to deliver continuous and multiple data output. b) Multiple communication interfaces protocol option to transfer the data from data logger to a PC or notebook. c) All the accessories which is required to support the data monitoring system. d) Suitable guidelines and manuals for total system. 						
 A suitchle user friendly ease						
A suitable user-friendly, easy access software which will compatible with						
communication interface networks. The						
software should capable to display the						
transferred real-time as well as						
previous data in the form of tables,						
charts etc. and analyse the same to						
deliver outputs with graphical interface.						

Note: The above financial template should be strictly followed. Any deviation from the above template (in terms of description and specification of the item) may lead to cancellation of the tender.