#### Notice Inviting Quotation (E-Procurement mode) कोटेशन को आमंत्रित करने की सूचना (इ-प्रोक्योर्मेंट मोड)

#### INDIAN INSTITUTE OF TECHNOLOGY DELHI भारतीय प्रौद्योगिकी संस्थान दिल्ली HAUZ KHAS, NEW DELHI-110016 हौज ख़ास, नई दिल्ली -110016

Dated:18/12/2018

#### **Open Tender Notice No.: IITD/BCVL(SP-1994)/2018**

Indian Institute of Technology Delhi is in the process of purchasing following item(s) as per details as given as under.

Details of the item आइटम का विवरण	Handheld XRF Analyzer for Geochemical Analysis of Soils
Earnest Money Deposit to be submitted बयाना जमा करने के लिए जमा राशि	NIL
Warranty वारंटी अवधि	2 Years + One year Extended Warranty
Performance security निष्पादन सुरक्षा	NIL

Tender Documents may be downloaded from Central Public Procurement Portal <u>http://eprocure.gov.in/eprocure/app</u>.Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <u>http://eprocure.gov.in/eprocure/app</u>. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at 'Instructions for online Bid Submission'.

Tenderers can access tender documents on the website(For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, Click on "GO" button to view all IIT Delhi tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website <u>http://eprocure.gov.in/eprocure/app</u> as per the schedule given in the next page.

No manual bids will be accepted. All quotation (both Technical and Financial should be submitted in the E-procurement portal).

SCHEDULE						
Name of Organization	Indian Institute of Technology Delhi					
Tender Type (Open/Limited/EOI/Auction/Single)	OPEN					
Tender Category (Services/Goods/works)	GOODS					
Type/Form of Contract (Work/Supply/ Auction/Service/Buy/Empanelment/ Sell)	BUY					
Product Category (Civil Works/Electrical Works/Fleet Management/ Computer Systems)	OTHERS					
Source of Fund (Institute/Project)	Budget Code Equipment/ Project Code #RP03614					
Is Multi Currency Allowed	YES					
Date of Issue/Publishing	18/12/2018 (17:00Hrs)					
Document Download/Sale Start Date	18/12/2018 (17:00Hrs)					
Document Download/Sale End Date	08/01/2019 (15:00Hrs)					
Date for Pre-Bid Conference						
Venue of Pre-Bid Conference						
Last Date and Time for Uploading of Bids	08/01/2018 (15:00Hrs)					
Date and Time of Opening of Technical Bids	10/01/2019 (15:00Hrs)					
Tender Fee EMD	Rs.NIL/- (For Tender Fee)Rs.NIL/-(For EMD)(To be paid through RTGS/NEFT. IIT Delhi Bank details are as under:Name of the Bank A/C: IITD Revenue AccountSBI A/C No.: 10773572622Name of the Bank: State Bank of India, IIT Delhi, Hauz Khas, New Delhi-110016IFSC Code: SBIN0001077MICR Code: 110002156Swift No.: SBININBB547(This is mandatory that UTR Number is provided in the on- line quotation/bid. (Kindly refer to the UTR Column of the Declaration Sheet at Annexure-II)					
No. of Covers (1/2/3/4)	02					
Bid Validity days (180/120/90/60/30)	180days (From last date of opening of tender)					
Address for Communication	Prof. Manoj Datta, Principal Investigator Department of Civil Engineering, IIT Delhi, New Delhi-16.					
Contact No.	+91-11-2659 1183					
Fax No.	+91-11-2658 1117					
Email Address	mdatta@civil.iitd.ac.in					

## Chairman Purchase Committee (Buyer Member)

### Instructions for Online Bid Submission/ऑनलाइन बोली (बिड) के लिए निर्देश:

As per the directives of Department of Expenditure, this tender document has been published on the Central Public Procurement Portal (<u>URL:http://eprocure.gov.in/eprocure/app</u>). The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

व्यय विभाग के निर्देशों के अनुसार, यह निविदा दस्तावेज केंद्रीय सार्वजनिक प्रापण पोर्टल (यूआरएल: http: //eprocure.gov.in/eprocure/app) पर प्रकाशित किया गया है। बोलीदाताओं को मान्य डिजिटल हस्ताक्षर प्रमाण पत्र का उपयोग करते हुए सीपीपी पोर्टल पर इलेक्ट्रॉनिक रूप से अपनी बोलियों की सॉफ्ट प्रतियां जमा करना आवश्यक है। सीपीपी पोर्टल पर पंजीकरण करने के लिए निविदाकर्ताओं की सहायता करने के लिए नीचे दिए गए निर्देशों का मतलब है, सीपीपी पोर्टल पर आवश्यकताओं के अनुसार अपनी बोलियां तैयार करें और अपनी बोलियां ऑनलाइन जमा करें।

More information useful for submitting online bids on the CPP Portal may be obtained at:

अधिक जानकारी सीपीपी पोर्टल पर ऑनलाइन बोलियां जमा करने के लिए उपयोगी हो सकती है: <u>http://eprocure.gov.in/eprocure/app</u>

#### **REGISTRATION**

 Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal (URL:<u>http://eprocure.gov.in/eprocure/app</u>) by clicking on the link "Click here to Enroll". Enrolment on the CPP Portal is free of charge.

बोलीदाताओं को "नामांकन के लिए यहां क्लिक करें" लिंक पर क्लिक करके सेंट्रल पब्लिक प्रोक्युरमेंट पोर्टल (यूआरएल: http://eprocure.gov.in/eprocure/app) के ई-प्रोक्योरमेंट मॉड्यूल पर भर्ती करना आवश्यक है। सीपीपी पोर्टल पर नामांकन नि: शुल्क है

2) As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.

नामांकन प्रक्रिया के भाग के रूप में, बोलीदाताओं को अपने खाते के लिए एक अद्वितीय उपयोगकर्ता नाम चुनना होगा और एक पासवर्ड प्रदान करना होगा।

3) Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.

बोलीदाताओं को सलाह दी जाती है कि पंजीकरण प्रक्रिया के भाग के रूप में अपना वैध ईमेल पता और मोबाइल नंबर पंजीकृत करें। इन का उपयोग सीपीपी पोर्टल से किसी भी संचार के लिए किया जाएगा।

4) Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.), with their profile.

नामांकन पर, बोलीदाताओं को सीसीए इंडिया द्वारा मान्यता प्राप्त किसी प्रमाणन प्राधिकरण द्वारा जारी किए गए अपने मान्य डिजिटल हस्ताक्षर प्रमाण पत्र (कक्षा द्वितीय या कक्षा III प्रमाण पत्र के साथ महत्वपूर्ण उपयोग पर हस्ताक्षर करने) की आवश्यकता होगी (जैसे सिफी / टीसीएस / एनकोड / ई-मुद्राआदि) , उनके प्रोफाइल के साथ

- 5) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.
  केवल एक मान्य डीएससी एक बोलीदाता द्वारा पंजीकृत होना चाहिए। कृपया ध्यान दें कि निविदाकर्ता यह सुनिश्चित करने
- 6) Bidder then logs in to the site through the secured log-in by entering their userID / password and the password of the DSC / eToken.

के लिए ज़िम्मेदार हैं कि वे अपने डीएससी को दूसरों को उधार नहीं देते हैं जिससे दुरुपयोग हो सकता है।

बोलीदाता फिर अपने यूजर आईडी / पासवर्ड और डीएससी / ईटीकेन के पासवर्ड को दर्ज करके सुरक्षित लॉग-इन के माध्यम से साइट पर लॉग ऑन करता है।

#### SEARCHING FOR TENDER DOCUMENTS/निविदा दस्तावेजों के लिए खोजना

 There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.

सीपीपी पोर्टल में निर्मित विभिन्न खोज विकल्प हैं, ताकि बोलीदाताओं को कई मापदंडों से सक्रिय निविदाएं खोज सकें। इन मापदंडों में निविदा आईडी, संगठन का नाम, स्थान, तिथि, मूल्य आदि शामिल हो सकते हैं। निविदाओं के लिए उन्नत खोज का एक विकल्प भी है, जिसमें बोलीदाता कई नामों को जोड़ सकते हैं जैसे संगठन का नाम, अनुबंध का स्थान, स्थान, सीपीपी पोर्टल पर प्रकाशित निविदा की खोज के लिए तारीख, अन्य कीवर्ड आदि।

2) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.

बोलीदाताओं ने एक बार निविदाएं चुनी हैं जिसमें वे रुचि रखते हैं,उसका वे आवश्यक दस्तावेज / निविदा कार्यक्रम डाउनलोड कर सकते हैं। ये निविदाएं 'मेरी निविदाओं' फ़ोल्डर में ले जाईजा सकती हैं। इससे सीपीपी पोर्टल को बोलीदाताओं को एसएमएस / ई-मेल के माध्यम से सूचित किया जा सकता है, यदि निविदा दस्तावेज मेंकोई शुद्धि जारी कि गई है।

3) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

बोलीदाता को प्रत्येक निविदा को निर्दिष्ट अद्वितीय निविदा आईडी का नोट बनाना चाहिए, अगर वे हेल्पडेस्क से कोई स्पष्टीकरण / सहायता प्राप्त करना चाहते हैं।

#### <u>PREPARATION OF BIDS /</u> बोली (बिड) की तैयारी

1) Bidder should take into account any corrigendum published on the tender document before submitting their bids.

बोलीदाता को अपनी बोलियां जमा करने से पहले निविदा दस्तावेज पर प्रकाशित किसी भी शुद्धि को ध्यान में रखना चाहिए।

2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.

कृपया बोली के भाग के रूप में जमा किए जाने वाले दस्तावेजों को समझने के लिए निविदा विज्ञापन और निविदा दस्तावेज ध्यान से देखें। कृपया उन अंकों की संख्या पर ध्यान दें जिन में बोली दस्तावेज जमा करना है, दस्तावेजों की संख्या – जिसमें प्रत्येक दस्तावेज के नाम और सामग्री शामिल हैं, जिन्हें प्रस्तुत करने की आवश्यकता है। इनमें से कोई भी विचलन बोली को अस्वीकार कर सकता है।

3) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF formats. Bid documents may be scanned with 100 dpi with black and white option.

बोलीदाता, अग्रिम में, निविदा दस्तावेज / अनुसूची में बताए अनुसार प्रस्तुत करने के लिए बोली दस्तावेज तैयार करना चाहिए और आम तौर पर, वे पीडीएफ / एक्सएलएस / आरएआर / डीडब्ल्यूएफ स्वरूपों में हो सकते हैं। बोली दस्तावेजों को 100 डीपीआई के साथ काले और सफेद विकल्प स्कैन किया जा सकता है। 4) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Space" area available to them to upload such documents. These documents may be directly submitted from the "My Space" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

मानक दस्तावेजों के एक ही सेट को अपलोड करने के लिए आवश्यक समय और प्रयास से बचने के लिए जो प्रत्येक बोली के भाग के रूप में जमा करने के लिए आवश्यक हैं, ऐसे मानक दस्तावेज अपलोड करने का प्रावधान (जैसे पैन कार्ड कॉपी, वार्षिक रिपोर्ट, लेखा परीक्षक प्रमाण पत्र आदि) ) बोलीदाताओं को प्रदान किया गया है। ऐसे दस्तावेजों को अपलोड करने के लिए बोलीकर्ता उनके लिए उपलब्ध "मेरा स्पेस" क्षेत्र का उपयोग कर सकते हैं। बोली जमा करते समय ये दस्तावेज़ सीधे "मेरा स्पेस" क्षेत्र से जमा किए जा सकते हैं, और उन्हें बार-बार अपलोड करने की ज़रूरत नहीं है इससे बोली जमा प्रक्रिया के लिए आवश्यक समय में कमी आएगी।

## <u>SUBMISSION OF BIDS/</u>बोली (बिड) का जमा करना

1) Bidder should log into the site well in advance for bid submission so that he/she upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.

बोलीदाता को बोली प्रस्तुति के लिए अच्छी तरह से साइट पर लॉग इन करना चाहिए ताकि वह समय पर बोली अपलोड कर सके या फिर बोली प्रस्तुत करने के समय से पहले। अन्य मुद्दों के कारण किसी भी देरी के लिए बोलीदाता जिम्मेदार होगा।

2) The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.

बोलीदाता को निविदा दस्तावेज में दर्शाए अनुसार एक-एक करके आवश्यक बोली दस्तावेजों को डिजिटल हस्ताक्षर और अपलोड करना होगा।

3) Bidder has to select the payment option as "on-line" to pay the tender fee / EMD as applicable and enter details of the instrument. Whenever, EMD / Tender fees is sought, bidders need to pay the tender fee and EMD separately on-line through RTGS (Refer to Schedule, Page No.2).

बोलीदाता को निविदा शुल्क / ईएमडी को भुगतान के लिए "ऑन लाइन" के रूप में भुगतान विकल्प चुनना होगा और उपकरण का विवरण दर्ज करना होगा। जब भी, ईएमडी / निविदा शुल्क की मांग की जाती है, बोलीदाताओं को टेंडर शुल्क और ईएमडी अलग–अलग आरटीजीएस के माध्यम से ऑन लाइन पर भुगतान करने की आवश्यकता होती है (अनुसूची, पेज नं .2 देखें)।

4) A standard BoQ format has been provided with the tender document to be filled by all the bidders. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. Bidders are required to download the BoQ file, open it and complete the white colored (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.

एक मानक BoQप्रारूप को सभी बोलीदाताओं द्वारा भरने के लिए निविदा दस्तावेज प्रदान किया गया है। बोलीदाताओं को इस बात का ध्यान रखना चाहिएकि उन्हें आवश्यक प्रारूप में अपनी वित्तीय बोली जमा करनी चाहिए और कोई अन्य प्रारूप स्वीकार्य नहीं है। बोलीकर्ताओं को BoQफाइल को डाउनलोड करने, इसे खोलने और अपने संबंधित वित्तीय उद्धरण और अन्य विवरण (जैसे बोलीदाता का नाम) के साथ सफेद रंगीन (असुरक्षित) कोशिकाओं को पूरा करना आवश्यक है। कोई भी अन्य कक्ष नहीं बदला जाना चाहिए। एक बार विवरण पूरा हो जाने पर, बोलीदाता को इसे सहेजना होगा और इसे ऑनलाइन जमा करना होगा, बिना फ़ाइल नाम बदलना। यदि BOQ फ़ाइल को बोलीदाता द्वारा संशोधित किया गया है, तो बोली को खारिज कर दिया जाएगा।

### OR/या

In some cases Financial Bids can be submitted in PDF format as well (in lieu of BOQ).

कुछ मामलों में वित्तीय बोलियां पीडीएफ प्रारूप में भी जमा की जा सकती हैं (BOQ के बदले)

5) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.

सर्वर का समय (जो बोलीदाताओं के डैशबोर्ड पर प्रदर्शित होता है) बोलीदाताओं द्वारा बोलियों को खोलने के लिए समय सीमा को संदर्भित करने के लिए मानक समय के रूप में माना जाएगा। बोलीदाताओं को खोलना आदि। बोलीदाताओं को बोली प्रस्तुत करने के दौरान इस समय का पालन करना चाहिए।

6) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done.

बोलीदाताओं द्वारा प्रस्तुत सभी दस्तावेज पीकेआई एन्क्रिप्शन तकनीकों का उपयोग करके एन्क्रिप्ट किया जाएगा जिससे डेटा की गोपनीयता सुनिश्चित हो सके। दर्ज किए गए डेटा को अनधिकृत व्यक्तियों द्वारा बोली खोलने के समय तक नहीं देखा जा सकता है। बोलियों की गोपनीयता को सुरक्षित सॉकेट लेयर 128 बिट एन्क्रिप्शन तकनीक का उपयोग कर रखा जाता है। संवेदनशील क्षेत्रों का डेटा संग्रहण एन्क्रिप्शन किया जाता है।

7) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.

अपलोड किए गए निविदा दस्तावेज केवल अधिकृत बोलीदाता द्वारा निविदा खोलने के बाद ही पठनीय हो सकते हैं।

8) Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.

बोलियों के सफल और समय पर जमा होने पर, पोर्टल एक सफल बोली प्रस्तुत करने का संदेश देगा और एक बोली सारांश बोली संख्या के साथ प्रदर्शित किया जाएगा। और अन्य सभी प्रासंगिक विवरणों के साथ बोली प्रस्तुत करने की तारीख और समय।

9) Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.

कृपया अनुपालन पत्रक की एक पीडीएफ फाइल में सभी प्रासंगिक दस्तावेजों के स्कैन किए गए पीडीएफ़ को जोड़ दें।

#### ASSISTANCE TO BIDDERS /बोलीदाताओं को सहायता

1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.

निविदा दस्तावेज से संबंधित कोई भी प्रश्न और इसमें निहित नियमों और शर्तों को निविदा आमंत्रण प्राधिकरण को निविदा के लिए या निविदा में वर्णित प्रासंगिक संपर्क व्यक्ति से संबोधित किया जाना चाहिए।

2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is 1800 233 7315.

ऑनलाइन बोली प्रस्तुत करने या सामान्य में सीपीपी पोर्टल से संबंधित प्रश्नों की प्रक्रिया से संबंधित कोई भी प्रश्न 24x7 सीपीपी पोर्टल हैल्पडेस्क पर निर्देशित किया जा सकता है। हेल्पडेस्क के लिए संपर्क संख्या 1800 233 7315 है

## General Instructions to the Bidders / बोलीदाताओं के लिए सामान्य निर्देश

1) The tenders will be received online through portal <u>http://eprocure.gov.in/eprocure/app</u>. In the Technical Bids, the bidders are required to upload all the documents in .pdf format.

निविदाएं पोर्टल http://eprocure.gov.in/eprocure/app के माध्यम से ऑनलाइन प्राप्त होंगी तकनीकी बोलियों में, बोलीदाताओं को सभी दस्तावेजों को। पीडीएफ प्रारूप में अपलोड करना होगा।

Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/e-token 2) in the company's name is a prerequisite for registration and participating in the bid submission activities through https://eprocure.gov.in/eprocure/app. Digital Signature Certificates can be obtained from the authorized certifying agencies. details of which are available in the web site https://eprocure.gov.in/eprocure/app under the link "Information about DSC". कंपनी के नाम में स्मार्ट कार्ड / ई-टोकन के रूप में मान्य क्लास ॥ / ॥। डिजिटल हस्ताक्षर प्रमाण पत्र (डीएससी) के

पंजीकरण के लिए एक शर्त है और https://eprocure.gov.in/eprocure/ के माध्यम से बोली प्रस्तुत करने की गतिविधियों में भाग लेसकते है। डिजिटल हस्ताक्षर प्रमाण पत्र अधिकृत प्रमाणित एजेंसियों से प्राप्त की जा सकती है, जिनमें से जानकारी "डीएससी के बारे में सूचना" लिंक के तहत वेब साइट https://eprocure.gov.in/eprocure/app पर उपलब्ध है।

3) Tenderer are advised to follow the instructions provided in the 'Instructions to the Tenderer for the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at https://eprocure.gov.in/eprocure/app.
निविदाकर्ता को सलाह दी जाती है कि वे निविदाकार को निर्देश दिए गए हों ताकि ई-प्रोक्योरमेंट के लिए सेंट्रल पब्लिक प्रोकॉर्ममेंट पोर्टल के जरिए https://eprocure.gov.in/eprocure/app पर ऑनलाइन निविदाएं जमा कर सकें।

#### Department of Civil Engineering Indian Institute of Technology Delhi Hauz Khas, New Delhi-110 016

### NOTICE INVITING QUOTATIONS

Dated:18/12/2018

#### Subject: Handheld XRF Analyzer for Geochemical Analysis of Soils

#### **Invitation for Tender Offers**

Indian Institute of Technology Delhi invites online Bids (Technical bid and Commercial bid) from eligible and experienced OEM (Original Equipment Manufacturer) OR OEM Authorized Dealer for **Supply, installation &field demonstration of Handheld XRF Analyzer for Geochemical Analysis of Soils** >with (warranty period as stated at page #1 of this tender) on site comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document, which is available on CPP Portal http://eprocure.gov.in/eprocure/app

#### **TECHNICAL SPECIFICATION:**

Sl.	Technical Specifications
No.	
110.	
1	Equipment capable of identifying metal contaminants in soils, sediments, fluids and filters and
	comprising of the following specifications:
	(a) On-Site screening of heavy metals in soils and sediments
	(b) On-site screening of heavy metals for the determination of total elemental compositions of soil
	samples
	(c) Should be capable of measuring most of the pollutants such as Ag, As, Cd, Cr, Cu, Hg, Ni, Pb, Se, Tl,
	Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.
	(d) Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and
	Rare Earth Elements (REEs)
	(e) X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the
	beam radiation for field use of full 50kV power.
	(f) The excitation source should be 4W Au anode X-ray tube
	(g) Detector shall be large area silicon drift detector.
	(h) Capable of measuring in the temperature range $-10 \deg C$ to $+50 \deg C$ .
	(i) Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick
	targeting and real-time decision-making.
	(j) Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.
	(k) Battery should be rechargeable and hot swappable during battery charging
	(1) Should have USB, Bluetooth for data transfer.
	(m) Should have slot for Micro-SD card of at least 1 GB capacity
	(n) Should have soil stick with adjustable height and push button operation of XRF.
	(o) Should be light weight (less than 2 kg), portable and field-ruggedized
	(p) Should include 2 Li-Ion batteries
	(q) Should have optimized three-beam kV-filter-current settings such that different elements/pollutants
	are detectable with good accuracy. Beam settings should be selectable and programmable.
	<ul> <li>(r) Software package for analysis and reporting.</li> <li>(a) It should have an option of portable workstation with remote instrument control</li> </ul>
	(s) It should have an option of portable workstation with remote instrument control

- (t) Should include soil foot for providing hands-free analysis for long resting times.
- (u) Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap
- (v) Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.
- (w) In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included
- (x) The bid should be lump sum and CIF/CIP to IIT Delhi.
- (y) The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.
- (z) Consumables for one year, if any

Sl. No.	List of optional items
	Optional Extra Items: a. Cost of Comprehensive AMC (including Spares and minimum one visit per annum) after warranty period b. Standard spare parts/accessories (additional)

A complete set of tender documents\* may be Download by prospective bidder free of cost from the website <u>http://eprocure.gov.in/eprocure/app</u>. Bidder has to make payment of requisite fees (i.e. Tender fees (if any) and EMD) online through RTGS/NEFT only.

# **Terms & Conditions Details**

Sl. No.	Specification
1.	<b>Due date</b> : The tender has to be submitted on-line before the due date. The offers received after the
	due date and time will not be considered. No manual bids will be considered.
2.	<b>Preparation of Bids</b> : The offer/bid should be submitted in two bid systems (i.e.) Technical bid and
	financial bid. The technical bid should consist of all technical details along with commercial terms
	and conditions. Financial bid should indicate item wise price for the items mentioned in the
	technical bid. Financial Bids to be submitted in PDF format.
	The Technical bid and the financial bid should be submitted Online.
3.	EMD (if applicable): The tenderer should submit an EMD amount through RTGS/NEFT. The
	Technical Bid without EMD would be considered as UNRESPONSIVE and will not be accepted.
	The EMD will be refunded without any interest to the unsuccessful bidders after the award of
	contract. Refer to Schedule (at page 1 of this document) for its actual place of submission.
4.	<b>Refund of EMD</b> : The EMD will be returned to unsuccessful Tenderer only after the Tenders are
	finalized. In case of successful Tenderer, it will be retained till the successful and complete
	installation of the equipment.
5.	<b>Opening of the tender</b> : The online bid will be opened by a committee duly constituted for this
	purpose. Online bids (complete in all respect) received along with EMD (if any) will be opened as
	mentioned at "Annexure: Schedule" in presence of bidders representative if available. Only one
	representative will be allowed to participate in the tender opening. Bid received without EMD (if
	present) will be rejected straight way. The technical bid will be opened online first and it will be
	examined by a technical committee (as per specification and requirement). The financial offer/bid
	will be opened only for the offer/bid which technically meets all requirements as per the
	specification, and will be opened in the presence of the vendor's representatives subsequently for
	further evaluation. The bidders if interested may participate on the tender opening Date and Time.
	The bidder should produce authorization letter from their company to participate in the tender
	opening.
6.	Acceptance/ Rejection of bids: The Committee reserves the right to reject any or all offers without
	assigning any reason.
7.	Pre-qualification criteria:
	(i) Bidders should be the manufacturer / authorized dealer. Letter of Authorization from original
	equipment manufacturer (OEM) on the same and specific to the tender should be enclosed.
	(ii) An undertaking from the OEM is required stating that they would facilitate the bidder on a
	regular basis with technology/product updates and extend support for the warranty as well. (Ref.
	Annexure-II)
	(iii) OEM should be internationally reputed Branded Company.
	(iv) Non-compliance of tender terms, non-submission of required documents, lack of clarity of the
	specifications, contradiction between bidder specification and supporting documents etc. may lead to
	rejection of the bid.
	(v) In the tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEM itself can
	bid but both cannot bid simultaneously for the same item/product in the same tender.
	(vi) If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on
0	behalf of another Principal/OEM in the same tender for the same item/product.
8.	<b>Performance Security</b> : The supplier shall require to submit the performance security in the form of
	irrevocable bank guarantee issued by any Indian Nationalized Bank for an amount which is stated at
	page #1 of the tender document within 21 days from the date of receipt of the purchase order/LC and
	should be kept valid for a period of 60 days beyond the date of completion of warranty period.
9.	Force Majeure: The Supplier shall not be liable for forfeiture of its performance security, liquidated
	damages or termination for default, if and to the extent that, it's delay in performance or other failure

	to perform its obligations under the Contract is the result of an event of Force Majeure.
	• For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
	• If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
10.	<b>Risk Purchase Clause</b> : In event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from the other source on the total risk of the supplier under risk purchase clause.
11.	<b>Packing Instructions</b> : Each package will be marked on three sides with proper paint/indelible ink, the
	following:
	i.Item Nomenclature
	ii.Order/Contract No.
	iii.Country of Origin of Goods
	iv.Supplier's Name and Address
	v.Consignee details
	vi.Packing list reference number
12.	Delivery and Documents:
	Delivery of the goods should be made within a maximum of 12 weeks from the date of the opening of LC. Within 24 hours of shipment, the supplier shall notify the purchaser and the insurance company by achla(talay/faw/a mail the full datails of the shipment including contrast number, resilvery, received
	cable/telex/fax/e mail the full details of the shipment including contract number, railway receipt number/ AAP etc. and date, description of goods, quantity, name of the consignee, invoice etc. The
	supplier shall mail the following documents to the purchaser with a copy to the insurance company:
	1. 4 Copies of the Supplier invoice showing contract number, goods' description, quantity
	2. unit price, total amount;
	3. Insurance Certificate if applicable;
	<ol> <li>Manufacturer's/Supplier's warranty certificate;</li> <li>Inspection Certificate issued by the nominated inspection agency, if any</li> </ol>
	<ol> <li>6. Supplier's factory inspection report; and</li> </ol>
	<ol> <li>Supplet's factory inspection report, and</li> <li>Certificate of Origin (if possible by the beneficiary);</li> </ol>
	<ol> <li>8. Two copies of the packing list identifying the contents of each package.</li> </ol>
	<ol> <li>9. The above documents should be received by the Purchaser before arrival of the Goods (except)</li> </ol>
	where the Goods have been delivered directly to the Consignee with all documents) and, if not
	received, the Supplier will be responsible for any consequent expenses.
13.	<b>Delayed delivery:</b> If the delivery is not made within the due date for any reason, the Committee will
101	have the right to impose penalty 1% per week and the maximum deduction is 10% of the contract
	value / price.
14.	<b>Prices</b> : The price should be quoted in lumpsum per unit with breakup of all packing and delivery
	charges and optional extra items. The offer/bid should be exclusive of taxes and duties, which will be
	paid by the purchaser as applicable. However the percentage of taxes & duties shall be clearly
	indicated.
	The price should be quoted without custom duty and excise duty, since IIT Delhi is exempted from
	payment of Excise Duty and is eligible for concessional rate of custom duty. Necessary certificate
	will be issued on demand.
	In case of imports, the price should be quoted on CIF upto IIT Delhi. IIT Delhi will make
	necessary arrangements for the clearance of imported goods at the Airport/Seaport. Hence the price

	should not include the above charges. "In case of <b>CIF</b> shipments, kindly provide the shipment information at least 2 days in advance before landing the shipment along with the documents i.e. invoice, packing list, forwarder Name, address, contact No. in India to save penalty/demurrage charges (imposed by Indian Customs). Otherwise these charges will be recovered from the supplier/Indian Agent."
15.	Notices: For the purpose of all notices, the following shall be the address of the Purchaser and Supplier.         Purchaser:       Prof. ManojDatta,         Principal Investigator       Department of Civil Engineering         Indian Institute of Technology       HauzKhas, New Delhi - 110016.
	Supplier:       (To be filled in by the supplier)         (All supplier's should submit its supplies information as per Annexure-II).
16.	<ul> <li>Progress of Supply: Wherever applicable, supplier shall regularly intimate progress of supply, in writing, to the Purchaser as under:</li> <li>1. Quantity offered for inspection and date;</li> <li>2. Quantity accepted/rejected by inspecting agency and date;</li> <li>3. Quantity dispatched/delivered to consignees and date;</li> <li>4. Quantity where incidental services have been satisfactorily completed with date;</li> <li>5. Quantity where rectification/repair/replacement effected/completed on receipt of any communication from consignee/Purchaser with date;</li> <li>6. Date of completion of entire Contract including incidental services, if any; and</li> <li>7. Date of receipt of entire payments under the Contract (In case of stage-wise inspection, details required may also be specified).</li> </ul>
17.	<ul> <li>Inspection and Tests: Inspection and tests prior to shipment of Goods and at final acceptance are as follows:</li> <li>After the goods are manufactured and assembled, inspection and testing of the goods shall be carried out at the supplier's plant by the supplier, prior to shipment to check whether the goods are in conformity with the technical specifications attached to the purchase order. Manufacturer's test certificate with data sheet shall be issued to this effect and submitted along with the delivery documents. The purchaser shall be present at the supplier's premises during such inspection and testing if need is felt. The location where the inspection is required to be conducted should be clearly indicated. The supplier shall inform the purchaser about the site preparation, if any, needed for installation of the goods at the purchaser's site at the time of submission of order acceptance.</li> <li>The acceptance test will be conducted by the Purchaser, their consultant or other such person nominated by the Purchaser at its option after the equipment is installed at purchaser's site in the presence of supplier's representatives. The acceptance will involve trouble free operation and ascertaining conformity with the ordered specifications and quality. There shall not be any additional charges for carrying out acceptance test. No malfunction, partial or complete failure of any part of the equipment is expected to occur. The Supplier shall maintain necessary log in respect of the result of the test to establish to the entire satisfaction of the Purchaser, the successful completion of the ordered item failing to pass the acceptance test, failing which the Purchaser reserve the right to get the equipment replaced by the Supplier at no extra cost to the Purchaser.</li> </ul>

	shall also be the responsibility and at the cost of the Supplier.
18.	<b>Resolution of Disputes</b> : The dispute resolution mechanism to be applied pursuant shall be as follows:
	• In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any
	matter arising out of or connected with this agreement, such disputes or difference shall be settled in
	accordance with the Indian Arbitration& Conciliation Act, 1996, the rules there under and any
	statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The
	dispute shall be referred to the Director, Indian Institute of Technology (IIT) Delhi and if he is unable
	or unwilling to act, to the sole arbitration of some other person appointed by him willing to act as
	such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all
	parties to this order.
	• In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by
	arbitration in accordance with provision of sub-clause (a) above. But if this is not acceptable to the
	supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United
	Nations Commission on International Trade Law) Arbitration Rules.
	• The venue of the arbitration shall be the place from where the order is issued.
19.	Applicable Law: The place of jurisdiction would be New Delhi (Delhi) INDIA.
20.	Right to Use Defective Goods
	If after delivery, acceptance and installation and within the guarantee and warranty period, the
	operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or
	complete replacement is made without interfering with the Purchaser's operation.
21.	Supplier Integrity
21.	The Supplier is responsible for and obliged to conduct all contracted activities in accordance with the
	Contract using state of the art methods and economic principles and exercising all means available to
	achieve the performance specified in the contract.
22.	Training
22.	Training The Supplier is required to provide training to the designated Purchaser's technical and end user
	The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.
22.	The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment. Installation & Demonstration
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li><b>Installation &amp; Demonstration</b></li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance</li> </ul>
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form.</li> </ul>
	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li><b>Installation &amp; Demonstration</b></li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form.</li> <li><b>Insurance:</b> For delivery of goods at the purchaser's premises, the insurance shall be obtained by the</li> </ul>
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration</li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form.</li> </ul>
23.	The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment. <b>Installation &amp; Demonstration</b> The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. <b>Insurance:</b> For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li><b>Installation &amp; Demonstration</b></li> <li>The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.</li> <li>In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form.</li> <li><b>Insurance:</b> For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final</li> </ul>
23.	The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment. <b>Installation &amp; Demonstration</b> The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. <b>Insurance:</b> For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. <b>For orders placed on CIF basis, the insurance should be up</b>
23.	The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment. <b>Installation &amp; Demonstration</b> The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. <b>Insurance:</b> For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. <b>For orders placed on CIF basis, the insurance should be up to IIT Delhi.</b>
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. For orders placed on CIF basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include: <ul> <li>Furnishing of 01 set of detailed operations &amp; maintenance manual.</li> </ul></li></ul>
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. For orders placed on CIF basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include: <ul> <li>Furnishing of 01 set of detailed operations &amp; maintenance manual.</li> <li>Arranging the shifting/moving of the item to their location of final installation within IITD</li> </ul></li></ul>
23. 24. 25.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. For orders placed on CIF basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include: <ul> <li>Furnishing of 01 set of detailed operations &amp; maintenance manual.</li> <li>Arranging the shifting/moving of the item to their location of final installation within IITD premises at the cost of Supplier through their Indian representatives. </li> </ul></li></ul>
23.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. For orders placed on CIF basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include: <ul> <li>Furnishing of 01 set of detailed operations &amp; maintenance manual.</li> <li>Arranging the shifting/moving of the item to their location of final installation within IITD premises at the cost of Supplier through their Indian representatives. </li> </ul></li></ul>
23. 24. 25.	<ul> <li>The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.</li> <li>Installation &amp; Demonstration The supplier is required to carry out the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials. In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance company as per his convenience. IITD will not be liable to any type of losses in any form. Insurance: For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. For orders placed on CIF basis, the insurance should be up to IIT Delhi. Incidental services: The incidental services also include: <ul> <li>Furnishing of 01 set of detailed operations &amp; maintenance manual.</li> <li>Arranging the shifting/moving of the item to their location of final installation within IITD premises at the cost of Supplier through their Indian representatives. </li> </ul></li></ul>

-	or in part, the Supplier shall at its discretion make such
	o the Goods or any part thereof as may be necessary in
e	pecified in the Contract at its own cost and expense and warranty should be comprehensive on site.
	Supplier in writing of any claims arising under this
	Supplier shall immediately within in 02 days arrange
	r parts thereof free of cost at the ultimate destination.
· · · ·	parts/goods at the time of their replacement. No claim
	the replaced parts/goods thereafter. The period for
correction of defects in the warranty period	d is 02 days. If the supplier having been notified fails
	purchaser may proceed to take such remedial action as
	nd expenses and without prejudice to any other rights,
which the purchaser may have against the	
	mentioned. The maintenance charges (AMC) under f the warranty should also be mentioned. The
comprehensive warranty will com	
1 2	pment against the defect of any manufacturing,
workmanship and poor quality of the com	
7. Governing Language	
The contract shall be written in English lang	uage. English language version of the Contract shall
	nd other documents pertaining to the Contract, which
are exchanged by the parties, shall be written in	the same language.
8. Applicable Law The Contrast shall be interpreted in accordance	a with the laws of the Union of India and all disputes
shall be subject to place of jurisdiction.	e with the laws of the Union of India and all disputes
9. Notices	
	her pursuant to this contract/order shall be sent to the
	, FAX or e mail and confirmed in writing to the other
party's address.	
	d or on the notice's effective date, whichever is later.
0. <b>Taxes</b>	
	taxes, duties, license fees, octroi, road permits, etc.,
	to the Purchaser. However, GST etc, in respect of the lier shall be payable extra, if so stipulated in the order.
transaction between the rutenaser and the Supp	nel shan de payadle exita, il so supulated il the order.
For research purpose(s) ONLY, 5% GST will b	e applicable with concessional GST Certificate.
1. Duties	
1 100	uty under notification No.51/96 (partially orfull) and
	e" can be issued after providing following information
• -	e issued to the shipment in the name of the Institute, (no
and research work only.	rocured product should be used for teaching, scientific
a) Shipping details i.e. Master Airway Bill No.	and House Airway No. (if exists)
b) Forwarder details i.e. Name, Contact No., etc.	•
	GST and necessary GST ExemptionCertificate will be
provided for which following information are re	-
<ul><li>b) Quotation with details of Basic Price, Rate, T</li><li>c) Supply Order Copy</li></ul>	as a Amount on which ED is applicable
d) Proforma-Invoice Copy.	
	f any will be paid to the Indian agent in Rupees on

	receipt of the equipment and after satisfactory installation. Agency Commission will not be paid in foreign currency under any circumstances. The details should be explicitly shown in Tender even in case of Nil commission. The tenderer should indicate the percentage of agency commission to be paid to the Indian agent.
33.	Payment:
55.	<ul> <li>(i) For imported items Payment will be made through irrevocable Letter of Credit (LC) Cash Against Documents (CAD)/Against delivery/after satisfactory installation by T.T. Letter of Credit (LC) will be established in favour of foreign Supplier after the submission of performance security. The letter of credit (LC) will be established on the exchange rates as applicable on the date of establishment. For Imports, LC will be opened for 100% FOB/CIF value. 95% of the LC amount shall be released on presentation of complete and clear shipping documents and 5% of the LC amount shall be released after the installation and demonstration of the equipment at the INST site of installation in faultless working condition for period of 60 days from the date of the satisfactory installation and subject to the production of unconditional performance bank guarantee as specified in Clause 8 of tender terms and conditions.</li> <li>(ii) For Indigenous supplies, 100% payment shall be made by the Purchaser against delivery, inspection, successful installation, commissioning and acceptance of the equipment at IITD in good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in Clause 9 of tender terms and conditions.</li> <li>(iii) Indian Agency commission (IAC), if any shall be paid after satisfactory installation &amp; commissioning of the goods at the destination at the exchange rate prevailing on the date of negotiation of LC documents, subject to DGS&amp;D registration for restricted items.</li> <li>(iv) All the bank charges within India will be borne by the Institute and outside India will be borne</li> </ul>
	by the Supplier.
34.	<b>User list:</b> Brochure detailing technical specifications and performance, list of industrial and educational establishments where the items enquired have been supplied must be provided. (Ref.
	Annexure-III)
35.	<ul> <li>Manuals and Drawings <ul> <li>Before the goods and equipment are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.</li> <li>The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract.</li> <li>Unless and otherwise agreed, the goods equipment shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser.</li> </ul> </li> </ul>
36.	<b>Application Specialist</b> : The Tenderer should mention in the Techno-Commercial bid the availability and names of Application Specialist and Service Engineers in the nearest regional office. (Ref. to Annexure-III)
37.	<ul> <li>Site Preparation: The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter.</li> <li>The supplier shall visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre-installation requirements.</li> </ul>
38.	Spare Parts         The Supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:         ii. Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this election shall not relieve the Supplier of any warranty obligations under the Contract; and

43.	<b>Disputes and Jurisdiction</b> : Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within New Delhi.
42.	<b>Training of Personnel:</b> The supplier shall be required to undertake to provide the technical training to the personnel involved in the use of the equipment at the Institute premises, immediately after completing the installation of the equipment for a minimum period of one week at the supplier's cost.
41.	<b>Downtime:</b> During the warranty period not more than 5% downtime will be permissible. For every day exceeding permissible downtime, penalty of 1/365 of the 5% FOB value will be imposed. Downtime will be counted from the date and time of the filing of complaint with in the business hours.
	<ul> <li>In the event the Purchaser terminates the Contract in whole or in part, the Purchaser may procure, upon such terms and in such manner, as it deems appropriate, Goods or Services similar to those undelivered, and the Supplier shall be liable to the Purchaser for any excess costs for such similar Goods or Services. However, the Supplier shall continue the performance of the Contract to the extent not terminated.</li> </ul>
	<ul> <li>For the purpose of this Clause:         <ol> <li>"Corrupt practice" means the offering, giving, receiving or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution.</li> <li>"Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Borrower, and includes collusive practice among Bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the Borrower of the benefits of free and open competition;"</li> </ol> </li> </ul>
40.	<ul> <li>Termination for Default</li> <li>The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default sent to the Supplier, terminate the Contract in whole or part: <ol> <li>If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the order, or within any extension thereof granted by the Purchaser; or</li> <li>If the Supplier fails to perform any other obligation(s) under the Contract.</li> <li>If the Supplier, in the judgment of the Purchaser has engaged in corrupt or fraudulent practices in competing for or in executing the Contract.</li> </ol> </li> </ul>
39.	<ul> <li>iii. In the event of termination of production of the spare parts:</li> <li>iv. Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements; and</li> <li>v. Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if requested.</li> <li>Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the Goods, such as gaskets, plugs, washers, belts etc. Other spare parts and components shall be supplied as promptly as possible but in any case within six months of placement of order.</li> <li>Defective Equipment: If any of the equipment supplied by the Tenderer is found to be substandard, refurbished, un-merchantable or not in accordance with the description/specification or otherwise faulty, the committee will have the right to reject the equipment or its part. The prices of such equipment shall be refunded by the Tenderer with 18% interest if such payments for such equipment have already been made. All damaged or unapproved goods shall be returned at supplier. Defective part in equipment, if found before installation and/or during warranty period, shall be replaced within 45 days on receipt of the intimation from this office at the cost and risk of supplier including all other charges. In case supplier fails to replace above item as per above terms &amp; conditions, IIT Delhi may consider "Banning" the supplier.</li> </ul>

44.	Compliancy certificate:	This	certificate	must be	provided	indicating	conformity to	the tec	hnical
	specifications. (Annexure	-I)							

# **COMPLIANCE SHEET**

### **TECHNICAL SPECIFICATION**

Equipment capable of identifying metal contaminants in soils, sediments, fluids and filters and comprising of the following specifications:         1       On-Site screening of heavy metals in soils and sediments         2       On-site screening of heavy metals for the determination of total elemental compositions of soil samples         3       Should be capable of measuring most of the pollutants such as Ag, As, Cd, Cr, Cu, Hg, Ni, Pb, Se, Tl, Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.         4       Should be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)         5       X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full S0kV power.         6       The excitation source should be 4W Au anode X-ray tube         7       Detector shall be large area silicon drift detector.         8       Capable of measuring in the temperature range -10 deg C to +50 deg C.         9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and raa-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB Bluetooth for data transfer.         13       Should have soil stick with adjustable height and push button operation of XRF.	SI. No.	Technical Specifications	Compliance Y/N
comprising of the following specifications:         On-Site screening of heavy metals in soils and sediments           1         On-Site screening of heavy metals for the determination of total elemental compositions of soil samples           3         Should be capable of measuring most of the pollutants such as Ag, As, Cd, Cr, Cu, Hg, Ni, Pb, Se, TI, Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.           4         Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)           5         X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full S0kV power.           6         The excitation source should be 4W Au anode X-ray tube           7         Detector shall be large area silicon drift detector.           8         Capable of measuring in the temperature range -10 deg C to +50 deg C.           9         Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.           10         Should have USB, Bluetooth for data transfer.           11         Battery should be for Micro-SD card of at least 1 GB capacity           14         Should have Soil stick with adjustable height and push button operation of XRF.           15         Should have onjtim detectable with good accuracy. Beam settings should be selectable and programmable.           17         Should have optimized three-beam kV-filter-current		oment capable of identifying metal contaminants in soils, sediments, fluids and filters and	
1         On-Site screening of heavy metals in soils and sediments           2         On-site screening of heavy metals for the determination of total elemental compositions of soil samples           3         Should be capable of idensity for the determination of total elemental compositions of soil samples           3         Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)           5         X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full 50kV power.           6         The excitation source should be 4W Au anode X-ray tube           7         Detector shall be large area silicon drift detector.           8         Capable of measuring in the temperature range -10 deg C to +50 deg C.           9         Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.           10         Should have USB, Bluetooth for data transfer.           13         Should have USB, Bluetooth for data transfer.           13         Should have onjetift elementa 2 ky, portable and field-ruggedized           16         Should have optimized three-beam kV-filter-current settings should be selectable and programmable.           17         Should have optimized three-beam kV-filter-current settings should be selectable and programmable.           18         Software package for analysis			
2         On-site screening of heavy metals for the determination of total elemental compositions of soil samples           3         Should be capable of measuring most of the pollutants such as Ag, As, Cd, Cr, Cu, Hg, Ni, Pb, Se, Tl, Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.           4         Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)           5         X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full 50kV power.           6         The excitation source should be 4W Au anode X-ray tube           7         Detector shall be large area silicon drift detector.           8         Capable of measuring in the temperature range -10 deg C to +50 deg C.           9         Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.           10         Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.           11         Battery should be rechargeable and hot swappable during battery charging           12         Should have solt of Micro-SD card of at least 1 GB capacity           13         Should have solt for Micro-SD card of at least 1 GB capacity           14         Should have solt sitck with adjustable height and push button operation of XRF.           15         Should have optimized three-beam kV-filter-current settings such that differen			
<ul> <li>Should be capable of measuring most of the pollutants such as Ag, As, Cd, Cr, Cu, Hg, Ni, Pb, Se, Ti, Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.</li> <li>Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)</li> <li>X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full 50kV power.</li> <li>The excitation source should be 4W Au anode X-ray tube</li> <li>Detector shall be large area silicon drift detector.</li> <li>Capable of measuring in the temperature range -10 deg C to +50 deg C.</li> <li>Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.</li> <li>Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.</li> <li>Battery should be rechargeable and hot swappable during battery charging</li> <li>Should have USB, Bluetooth for data transfer.</li> <li>Should have soil stick with adjustable height and push button operation of XRF.</li> <li>Should have soil stick with adjustable height and push button operation of XRF.</li> <li>Should have sol stick with adjustable height and push button operation of XRF.</li> <li>Should have optimized three-beam kV-filter-current settings should be selectable and programmable.</li> <li>Software package for analysis and reporting.</li> <li>It should have an option of portable workstation with remote instrument control</li> <li>Should include soil foot for providing hands-free analysis for long resting strap</li> <li>Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap</li> <li>Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.</li> <li>The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years</li></ul>	2		
Se, TI, Zn, and RCRA Metals Ag, As, Ba, Cd, Cr, Hg, Pb, and Se.           4         Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REBs)           5         X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full S0kV power.           6         The excitation source should be 4W Au anode X-ray tube           7         Detector shall be large area silicon drift detector.           8         Capable of measuring in the temperature range -10 deg C to +50 deg C.           9         Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.           10         Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.           11         Battery should be rechargeable and hot swappable during battery charging           12         Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.           11         Battery should be rechargeable and hot swappable during battery charging           12         Should have solt for Micro-SD card of at least 1 GB capacity           13         Should have solt solt with adjustable height and push button operation of XRF.           15         Should have solt solt with good accuracy. Beam settings should be selectable and programmable.           17         Should have an option of portable workstati		samples	
4       Should also be capable of identifying high-Z and challenging elements, such as Sn, Cr, Sb, Te, and Rare Earth Elements (REEs)         5       X-ray tube/incident beam should have high power: 50 kV. Must include safety shield for shielding the beam radiation for field use of full 50kV power.         6       The excitation source should be 4W Au anode X-ray tube         7       Detector shall be large area silicon drift detector.         8       Capable of measuring in the temperature range -10 deg C to +50 deg C.         9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have SIB, Bluetooth for data transfer.         13       Should have soil stick with adjustable height and push button operation of XRF.         15       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should have an option of portable workstation with remote instrument control         21       Standard accessor	3		
shielding the beam radiation for field use of full 50kV power.           6         The excitation source should be 4W Au anode X-ray tube           7         Detector shall be large area silicon drift detector.           8         Capable of measuring in the temperature range -10 deg C to +50 deg C.           9         Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.           10         Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.           11         Battery should be rechargeable and hot swappable during battery charging           12         Should have solt for Micro-SD card of at least 1 GB capacity           13         Should have soil stick with adjustable height and push button operation of XRF.           15         Should have soil stick with adjustable height and push button operation of XRF.           16         Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.           18         Software package for analysis and reporting.           19         It should have an option of portable workstation with remote instrument control           20         Should include soil foot for providing hands-free analysis for long resting times.           21         Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainles	4		
6       The excitation source should be 4W Au anode X-ray tube         7       Detector shall be large area silicon drift detector.         8       Capable of measuring in the temperature range -10 deg C to +50 deg C.         9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB, Bluetooth for data transfer.         13       Should have soil stick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Stan	5		
7       Detector shall be large area silicon drift detector.         8       Capable of measuring in the temperature range -10 deg C to +50 deg C.         9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have solo for Micro-SD card of at least 1 GB capacity         13       Should have solit sick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, statinless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment to be done at an intermational workshop to 100 participants over	6	*	
8       Capable of measuring in the temperature range -10 deg C to +50 deg C.         9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB, Bluetooth for data transfer.         13       Should have slot for Micro-SD card of at least 1 GB capacity         14       Should be using the explored of at least 1 GB capacity         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should niclude 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being s			
9       Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick targeting and real-time decision-making.         10       Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.         11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB, Bluetooth for data transfer.         13       Should have slot for Micro-SD card of at least 1 GB capacity         14       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period o	8		
11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB, Bluetooth for data transfer.         13       Should have slot for Micro-SD card of at least 1 GB capacity         14       Should have soil stick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.      <	9	Should have XRF-GPS-GIS explorer providing connectivity between XRF and GIS for quick	
11       Battery should be rechargeable and hot swappable during battery charging         12       Should have USB, Bluetooth for data transfer.         13       Should have solt for Micro-SD card of at least 1 GB capacity         14       Should have soil stick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.      <	10	Should have colour, high performance, daylight/backlit touchscreen of 3 to 4 inch size.	
13       Should have slot for Micro-SD card of at least 1 GB capacity         14       Should have soil stick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	11		
14       Should have soil stick with adjustable height and push button operation of XRF.         15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	12	Should have USB, Bluetooth for data transfer.	
15       Should be light weight (less than 2 kg), portable and field-ruggedized         16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	13	Should have slot for Micro-SD card of at least 1 GB capacity	
16       Should include 2 Li-Ion batteries         17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	14	Should have soil stick with adjustable height and push button operation of XRF.	
17       Should have optimized three-beam kV-filter-current settings such that different elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.         18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	15	Should be light weight (less than 2 kg), portable and field-ruggedized	
elements/pollutants are detectable with good accuracy. Beam settings should be selectable and programmable.18Software package for analysis and reporting.19It should have an option of portable workstation with remote instrument control20Should include soil foot for providing hands-free analysis for long resting times.21Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap22Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.23In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included24The bid should be lump sum and CIF/CIP to IIT Delhi.25The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	16	Should include 2 Li-Ion batteries	
18       Software package for analysis and reporting.         19       It should have an option of portable workstation with remote instrument control         20       Should include soil foot for providing hands-free analysis for long resting times.         21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	17	elements/pollutants are detectable with good accuracy. Beam settings should be selectable and	
<ul> <li>20 Should include soil foot for providing hands-free analysis for long resting times.</li> <li>21 Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap</li> <li>22 Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.</li> <li>23 In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included</li> <li>24 The bid should be lump sum and CIF/CIP to IIT Delhi.</li> <li>25 The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.</li> </ul>	18		
21       Standard accessories, such as waterproof carrying case, docking station, mini USB cable, stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	19	It should have an option of portable workstation with remote instrument control	
stainless steel calibration check reference coin, spare windows, integrated writ strap         22       Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.         23       In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included         24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	20	Should include soil foot for providing hands-free analysis for long resting times.	
<ul> <li>In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included</li> <li>The bid should be lump sum and CIF/CIP to IIT Delhi.</li> <li>The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.</li> </ul>	21		
<ul> <li>In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical personnel must also be included</li> <li>The bid should be lump sum and CIF/CIP to IIT Delhi.</li> <li>The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.</li> </ul>	22	Cost of hands on training at IIT Delhi for 5 days on the equipment being sold must be included.	
24       The bid should be lump sum and CIF/CIP to IIT Delhi.         25       The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	23	In addition to installation/training, cost of demonstration of the equipment to be done at an international workshop to 100 participants over the period of 3 days by expert technical	
25 The bidder should have supplied 2 or more similar equipment from the same OEM over the past three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.	24		
three years. Details of the clients are to be attached along with the name, address, email ids and mobile numbers of the contact persons.		1	
	23	three years. Details of the clients are to be attached along with the name, address, email ids and	
26 Consumables for one year if any	26	Consumables for one year, if any	

Optional Items:						
Sl.	Technical Specifications	Compliance				
No.		Ϋ́/Ν				
1	Cost of Comprehensive AMC (including Spares and minimum one visit per annum) after warranty period					
2	Standard spare parts/accessories (additional)					

I have also enclosed all relevant documents in support of my claims, (as above) in the following pages.

### Signature of Bidder

Name: \_\_\_\_\_

Designation:

Organization Name: \_\_\_\_\_

Contact No. : \_\_\_\_\_

#### << Organization Letter Head >> DECLARATION SHEET

We, \_\_\_\_\_\_\_hereby certify that all the information and data furnished by our organization with regard to this tender specification are true and complete to the best of our knowledge. I have gone through the specification, conditions and stipulations in details and agree to comply with the requirements and intent of specification.

This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certified that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty.

The prices quoted in the financial bids are subsidized due to academic discount given to IIT Delhi.

We, further specifically certify that our organization has not been Black Listed/De	NAME & ADDRESS OF THE Vendor/ Manufacturer / Agent
Listed or put to any Holiday by any Institutional Agency/ Govt. Department/ Public Sector	
Undertaking in the last three years.	
1 Phone	
2 Fax	
3 E-mail	
4 Contact Person Name	
5 Mobile Number	
6 GST Number	
7 PAN Number	
(In case of on-line payment of Tender Fees)	
8 UTR No. (For Tender Fee)	
(In case of on-line payment of EMD)	
9 UTR No. (For EMD)	
10 Kindly provide bank details of the bidder	
in the following format:	
a) Name of the Bank	
,	
b) Account Number	
c) Kindly attach scanned copy of one	
Chequebook page to enable us to return the	
EMD to unsuccessful bidder	

(Signature of the Tenderer)

Name:

Seal of the Company

List of Government Organizations for whom the Bidder has undertaken such work during last three years (must be supported with work orders)						
Name of the organization	Name of Contact Person	Contact No.				

Name of application specialist / Service Engineer who have the technical competency to handle and support the quoted product during the warranty period.							
Name of the organization	Name of Contact Person	Contact No.					

Signature of Bidder

Name: \_\_\_\_\_ Designation: \_\_\_\_\_

Organization Name: \_\_\_\_\_

Contact No. : \_\_\_\_\_

# **Bid Submission**

## **Online Bid Submission:**

The Online bids (complete in all respect) must be uploaded online intwo Envelops as explained below:-

Envelope – 1 (Following documents to be provided as single PDF file)							
Sl. No.	Document	Content	File Types				
1.	Technical	Compliance Sheet as per Annexure - I	.PDF				
2.	Bid	Organization Declaration Sheet as per Annexure - II	.PDF				
3.		List of organizations/ clients where the same products have been supplied (in last two years) along with their contact number(s). (Annexure-III)	.PDF				
4.		Technical supporting documents in support of all claims made at Annexure-I (Annexure-IV)	.PDF				
Envelope – 2							
Sl. No.	Document	Content					
1.	Financial Bid	Price bid should be submitted in PDF format.	.PDF				

## <Department/Centre Name> Indian Institute of Technology Delhi HauzKhas, New Delhi-110016

## Date: XX/XX/XXXX

### **Subject:Purchase of <Item> (Following format is used for imported items)**

S.	Currency	Description and Specification of	Qty.	Unit	Agency	Discount	Ex-works	Packing +	FOB/FCA	Insurance	CIF Price
No.		the Item	in	Price	Commission		price	Handling	AirportPric	+ Freight	(f+g)
			Units		(If		(d=a+b-c)	+ DOC +	e	(g)	
				(a)	applicable)	(c)		Inland	(f=d+e)		
					(b)			Freight +			
								FCA			
								Charges			
								(e)			
1											

Note: At any circumstances, it is the responsibility of the foreign supplier to hand over the material to our forwarder at the origin airport after completing all the inland clearing. No Ex-works consignment will be entertained.

#### For indigenous items please quote as per following format.

	Tor margenous items please quote as per ronowing rorman									
S.	Description and Specification of the	Qty. in Units	Unit Price in	GST%	Total Price in Rs.					
No.	Item		Rs.							
1.										
2.										

Note: The above financial template should be strictly followed. Any deviation from the above template (in terms of description and specification of the item) may lead to cancellation of the tender.