Notice Inviting Quotation (E-Procurement mode)

INDIAN INSTITUTE OF TECHNOLOGY DELHI HAUZ KHAS, NEW DELHI-110016

Dated: 06/12/2016

Open Tender Notice No.IITD/CIVIL(SP-725)/2016

Indian Institute of Technology Delhi is in the process of purchasing following item(s) as per details as given as under.

Details of the item	Geotechnical Dynamic Model testing Facility
Earnest Money Deposit to be submitted	Rs. NIL
Warranty	3 Years
Performance security	5% of FOB value (to be decided by the faculty concerned)

Tender Documents may be downloaded from Central Public Procurement Portal http://eprocure.gov.in/eprocure/app. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website http://eprocure.gov.in/eprocure/app. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at 'Instructions for online Bid Submission '.

Tenderers can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, Click on "GO" button to view all IIT Delhi tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website http://eprocure.gov.in/eprocure/app as per the schedule given in the next page.

No manual bids will be accepted. All quotation (both Technical and Financial should be submitted in the E-procurement portal).

Schedule

<u>Schedule</u>			
Name of Organization	Indian Institute of Technology Delhi		
Tender Type (Open/Limited/EOI/Auction/Single)	Open		
Tender Category (Services/Goods/works)	Goods		
Type/Form of Contract (Work/Supply/ Auction/Service/Buy/Empanelment/ Sell)	Form of Contract (Work/Supply/Service)		
Product Category (Civil Works/Electrical Works/Fleet Management/ Computer Systems)	Civil Works (Hydraulic System)		
Source of Fund (Institute/Project)	Budget Code RP03208		
Is Multi Currency Allowed	Yes		
Date of Issue/Publishing	06/12/2016 At (15:00 Hrs)		
Document Download/Sale Start Date	06/12/2016 At (15:00 Hrs)		
Document Download/Sale End Date	27/12/2016 At (15:00 Hrs)		
Date for Pre-Bid Conference			
Venue of Pre-Bid Conference			
Last Date and Time for Uploading of Bids	27/12/2016 At (15:00 Hrs)		
Date and Time of Opening of Technical Bids	29/12/2016 At (15:00 Hrs)		
Tender Fee EMD	Rs.NIL/- (For Tender Fee) Rs.NIL/-(For EMD) (To be paid through RTGS/NEFT. IIT Delhi Bank details are as under: Name of the Bank A/C : IITD Revenue Account SBI A/C No. : 10773572622 Name of the Bank : State Bank of India, IIT Delhi, Hauz Khas, New Delhi- 110016 IFSC Code : SBIN0001077 MICR Code : 110002156 Swift No. : SBININBB547 (This is mandatory that UTR Number is provided in the online quotation/bid. (Kindly refer to the UTR Column of the Declaration Sheet at Annexure-II)		
No. of Covers (1/2/3/4)	02		
Bid Validity days (180/120/90/60/30)	90days (From last date of opening of tender)		
Address for Communication	Dr. Bappaditya Manna, Civil Engg. Dept. IIT Delhi, Hauz Khas, New Delhi -110016		
Contact No.	011- 26591211		
Fax No.	011- 26581117		
Email Address	bmanna@civil.iitd.ac.in		

Chairman Purchase Committee (Buyer Member)

Instructions for Online Bid Submission:

As per the directives of Department of Expenditure, this tender document has been published on the Central Public Procurement Portal (<u>URL:http://eprocure.gov.in/eprocure/app)</u>. The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

More information useful for submitting online bids on the CPP Portal may be obtained at:

http://eprocure.gov.in/eprocure/app

REGISTRATION

- 1) Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal (URL: http://eprocure.gov.in/eprocure/app) by clicking on the link "Click here to Enroll". Enrolment on the CPP Portal is free of charge.
- 2) As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
- 3) Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- 4) Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.), with their profile.
- 5) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.
- 6) Bidder then logs in to the site through the secured log-in by entering their userID / password and the password of the DSC / eToken.

SEARCHING FOR TENDER DOCUMENTS

- 1) There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.
- 2) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.
- 3) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

PREPARATION OF BIDS

- 1) Bidder should take into account any corrigendum published on the tender document before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the bid documents have to be submitted, the number of documents including the names and content of

each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.

- 3) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF formats. Bid documents may be scanned with 100 dpi with black and white option.
- 4) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Space" area available to them to upload such documents. These documents may be directly submitted from the "My Space" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

SUBMISSION OF BIDS

- 1) Bidder should log into the site well in advance for bid submission so that he/she upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- 2) The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.
- 3) Bidder has to select the payment option as "on-line" to pay the tender fee / EMD as applicable and enter details of the instrument. Whenever, EMD / Tender fees is sought, bidders need to pay the tender fee and EMD separately on-line through RTGS (Refer to Schedule, Page No.2).
- 4) A standard BoQ format has been provided with the tender document to be filled by all the bidders. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. Bidders are required to download the BoQ file, open it and complete the white colored (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.

OR

In some cases Financial Bids can be submitted in PDF format as well (in lieu of BOQ).

- 5) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- 6) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 7) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 8) Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- 9) Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.

ASSISTANCE TO BIDDERS

- 1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- 2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is 1800 233 7315.

General Instructions to the Bidders

- 1) The tenders will be received online through portal http://eprocure.gov.in/eprocure/app . In the Technical Bids, the bidders are required to upload all the documents in .pdf format.
- 2) Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/e-token in the company's name is a prerequisite for registration and participating in the bid submission activities through https://eprocure.gov.in/eprocure/app. Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site https://eprocure.gov.in/eprocure/app under the link "Information about DSC".
- 3) Tenderer are advised to follow the instructions provided in the 'Instructions to the Tenderer for the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at https://eprocure.gov.in/eprocure/app.

Civil Engineering Department Indian Institute of Technology HauzKhas, New Delhi-110 016

NOTICE INVITING QUOTATIONS

Dated: 06/12/2016

Subject: Geotechnical Dynamic Model Testing Facility

Invitation for Tender Offers

Indian Institute of Technology Delhi invites online Bids (Technical bid and Commercial bid)from eligible and experienced OEM (Original Equipment Manufacturer) OR OEM Authorized Dealer for **Geotechnical Dynamic Model Testing Facility**with three years on site comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document, which is available on CPP Portal http://eprocure.gov.in/eprocure/app

TECHNICAL SPECIFICATION:

Sl. No.	Technical Specifications
01	(A) MANDATORY ITEMS:
	1) Servo-controlled Hydraulic Actuator 100 kN (01 No)
	 ✓ Double-ended, double-acting, fatigue-rated for combined balanced dynamic performance ✓ Rated load capacity: 100kN; Rated stroke: 150 mm(+/- 75 mm double amplitude) ✓ Full-stroke, coaxially-mounted displacement transducer 150 mm (+/- 75 mm double amplitude). ✓ High-capacity, non-metallic Polymerbearings bonded directly to the end caps ✓ Piston rods machined from a single piece of heat-treated alloy steel and surface coated ✓ Cushions to protect the actuator from the effects of high-speed and high-mass forces ✓ Manifold should have provision to accommodate dual Servo-valves with /without port shut-offs ✓ One servo-valve of 63 lpm (minimum) capacity(or to meet the below performance criteria) ✓ Fatigue rated Axial Load Cell (for each actuator capacity) with an overload capacity of 150% of the rated load; Non-linearity: 0.08% of full scale or better; Hysteresis: 0.05% of full scale or better; Repeatability: 0.03% of full scale or better

✓ Performance Criteria (100 kN actuator):

Force	Double Amplitude	Frequency
+/- 90 KN	1.5 mm ±0.25 mm	20 Hz
(90% capacity)	80 mm ±10 mm	1 Hz
+/- 75 KN	1.5 mm ± 0.25 mm	20 Hz
(75% capacity)	90 mm ±10 mm	1 Hz
+/- 50 KN	0.35 mm ±0.05 mm	30 Hz
(50% capacity)	100 mm ±10 mm	1 Hz
+/- 25 KN	0.3 mm ±0.05 mm	40 Hz
(25% capacity)	100 mm ±10 mm	1 Hz

- ✓ Performance curve of the complete system should be provided.
- ✓ One Swivel Head and One Swivel Base: Adjustable bearing for clearance to eliminate backlash; Dynamic force rating: ±160 kN or more(for 100 kN actuator); Tilt angle: ±17° or more (for head and base); Swivel angle: +90°/-75° or more; For use in cyclic, reversing load application(for head and base).
- ✓ Accuracy: $\pm 1\%$ above the threshold value of 10% of full-scale of sensors
- ✓ Both displacement transducer and Load Cell should be factory calibrated to full range

2) Servo-controlled Hydraulic Actuator 50 kN (01 No)

- ✓ Double-ended, double-acting, fatigue-rated for combined balanced dynamic performance
- ✓ Rated load capacity:50kN; Rated stroke: 150 mm(+/- 75 mm double amplitude)
- ✓ Full-stroke, coaxially-mounted displacement transducer 150 mm (+/- 75 mm double amplitude).
- ✓ High-capacity, non-metallic polymer bearings bonded directly to the end caps
- ✓ Piston rods machined from a single piece of heat-treated alloy steel and surface coated
- ✓ Cushions to protect the actuator from the effects of high-speed and high-mass forces
- ✓ Manifold should have provision to accommodate dual Servo-valves with /without port shutoffs
- ✓ One servo-valve of 63 lpm (minimum) capacity(or to meet the below performance criteria)
- ✓ Fatigue rated Axial Load Cell (for each actuator capacity) with an overload capacity of 150% of the rated load; Non-linearity : 0.08% of full scale or better; Hysteresis : 0.05% of full scale or better; Repeatability : 0.03% of full scale or better

✓ Performance Criteria (50 kN actuator):

Force	Double Amplitude	Frequency
+/- 45 KN	4.0 mm ± 0.5 mm	20 Hz
(90% capacity)	150 mm ± 15mm	1 Hz
+/- 37.5 KN	4 mm ± 0.5 mm	20 Hz
(75% capacity)	150 mm ± 15 mm	1 Hz
+/- 25 KN	1.5 mm ± 0.25 mm	30 Hz
(50% capacity)	150 mm ± 15 mm	1 Hz
+/- 12.5 KN	1.0 mm ± 0.20 mm	40 Hz
(25% capacity)	150 mm ± 15 mm	1 Hz

- ✓ Performance curve of the complete system should be provided.
- ✓ One Swivel Head and One Swivel Base: Adjustable bearing for clearance to eliminate backlash; Dynamic force rating: ±70 kN or more(for 50 kN actuator); Tilt angle: ±17° or more (for head and base); Swivel angle: +90°/-75° or more; For use in cyclic, reversing load application(for head and base).
- ✓ Accuracy: $\pm 1\%$ above the threshold value of 10% of full-scale of sensors
- ✓ Both displacement transducer and Load Cell should be factory calibrated to full range

3) Hydraulic Power Unit (01 No)

- ✓ Variable-volume pump, Rated flow capacity: 100 lpm or higher
- ✓ Foot print must be less than 2 m x 1m.
- ✓ 210 bar pressure, and 415 V AC, 50 Hz, 3-phase power
- ✓ Direct-coupled 60 HP submerged motor providing minimum 91% operating efficiency.
- ✓ Nominal noise rating of 63dB(A) or less measured at 3 feet with the pump operating at high pressure and dumping the full flow over the relief valve.
- ✓ Corrosion-resistant stainless steel oil-to-water heat exchanger to dissipate generated heat
- ✓ Submersed pump/motor design to eliminate the release of ambient heat.
- ✓ Temperature controlled water-saver and water shut-off valves to minimize water consumption.
- ✓ Cabinet with lockable cover to reduce noise and to prevent tampering of controls.
- ✓ Full flow 3-micron absolute filter in the return line to provide excellent oil cleaning.
- ✓ Integrated reservoir, pump/motor, heat exchanger
- ✓ Approved electrical disconnect interrupts power to the unit when the door is opened.
- ✓ Interlock circuitry for over-temperature and low-fluid level protection.
- ✓ Switches for start, low/high pressure, and stop on the front panel of the enclosure.
- ✓ High quality face seal fittings minimize the potential for leakage outside of the reservoir.
- ✓ Indicators for power on, low oil level, over temperature, and dirty filter warnings
- ✓ Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each

3. 1) Hydraulic Service Manifold (two Stations):

- ✓ Maximum operating pressure: 21 MPa (3000 PSI)
- ✓ Shall each allow the actuators to be independently controlled.
- ✓ Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel.
- ✓ Fast emergency unload for system depressurization.
- ✓ Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop.
- ✓ Flow rating: 114 LPM (30 GPM) or more.
- ✓ 3.8 litre (1 Gallon) pressure and 1 litre (0.26 Gallon) return line accumulator or more, Filter: 10 micron

4) Digital Controller (01 No)

- ✓ Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability.
- ✓ The controller shall be expandable maximum up to 4 channels
- ✓ Should have two stations to control two actuators
- ✓ Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster
- ✓ Signal conditioner data sample rate: 100 kHz or faster.
- ✓ Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes.
- ✓ Computer controlled transducer limits.
- ✓ Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading.
- ✓ Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters.
- ✓ Ability to save and restore PID tuning settings
- ✓ Adaptive controls compensation: Peak-Valley and Null Pacing
- ✓ Connected with high-speed serial interface with personal computer, No parallel interface is acceptable.
- ✓ Programmable libraries for real time control of system
- ✓ Should be available for use with C++, Visual basic, and Lab View.
- ✓ Online UPS should be included and should communicate with the controller.
- ✓ Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long
- ✓ One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control

4.1) Digital Universal Conditioner/Valve Driver Card

- ✓ The conditioner shall support transducers.
- ✓ Portability of transducer calibration data between conditioners

- ✓ Conditioners interface to DC or AC transducers and eliminate the need for separate conditioner types.
- ✓ Low-noise, low-drift, high- accuracy signal conditioning, and shunt calibration
- ✓ On-board strain gauge bridge completion.
- ✓ Excitation sensing and loss detection
- ✓ Processor-controlled limit detection
- ✓ Variable excitation frequencies for a wide variety of transducer types.
- ✓ Cards configurable to drive a two-stage valve and condition a transducer, or to drive a three-stage valve.

4.2) System Operating Software

- ✓ A software interface for configuring the controller and the user interface to the test station.
- ✓ A two channel on line data display (X vs. Y or Time vs. YY), digital displays, and a system exerciser for setting up tuning parameters and warming up the system prior to testing.
- ✓ Easy program for simple monotonic and cycle test execution including data acquisition.
- ✓ Test data may be stored in choice of ASCII, Lotus, or Excel formats for analysis with your favorite tools.
- ✓ Null pacing adaptive control algorithm for use with ramp command signals
- ✓ Adaptive phase and amplitude algorithm for use with (constant/block) sine wave command signals
- ✓ Perform amplitude control to ensure desired amplitudes are achieved.
- ✓ Data acquisition (timed, P/V, Level Crossing, Cyclic/Logarithmic)
- ✓ Function generation up to 600 Hz on all channels; Sine, square, triangle, ramp, hold, processes, and ability to play digitized profiles
- ✓ Soft start/stop available on all channels
- ✓ Amplitude Phase Control S/W to adjust the amplitude and phase of the command to realize the desired feedback signal.
- ✓ Access to test status information from any web-enabled, internet-connected device
- ✓ System Views Provide information about a single test system or multiple test system.
- ✓ Lab Views Provide information about all the test status of your lab from anywhere at any time
- ✓ Security Data transfers use SSL/TLS 256-bit encryption
- ✓ User Access Control Provides ability to create individual passwords for a page, or limit access to users
- ✓ Alerts Allow for email, text message or Tweet notification to interested parties when the state of a test changes

4.3) General Application Software

- ✓ Shall include the ability to control or to capture data from any installed AC or DC transducer, or externally conditioned transducer.
- ✓ Shall allow detection of digital inputs to the test system controller and allow digital

- output signals to be used as output from one of the control system digital output channels.
- ✓ The application software shall allow the following kinds of tests to be defined and executed.
 - ➤ Block loading fatigue
 - ➤ Constant amplitude fatigue
 - ➤ Random fatigue using an input files to define end levels and rates any of the above in combination.
 - Graphical drag and drop test layout design
 - > Test template creation
 - > System software application + custom waveform activities
 - ➤ Parallel branches for test execution and logical operators (if/then, while)
 - Limit sensing, sequencing triggers, and interface to digital I/O
 - ➤ Report generation while the test is running
- ✓ Shall have a detector watching for changes in peak / valley readings. The process can be used to trigger data collection or other definable test system processes.
- ✓ Shall allow the operator to selectively acquire data periodically as defined by logarithms, linear intervals or user selected intervals. This data shall be collected in time, level crossing or peak and valley samples.
- ✓ Shall allow a virtually unlimited number of test procedures to be set up and stored.
- ✓ Must be suitable for high rate testing.

B) REMARKS

- ✓ Supplier's engineer shall provide installation support and check-out service.
- ✓ On-line system manuals on the CD.
- ✓ The bidder must note that requirements in the above specifications are intended to be descriptive only and not restrictive. The bids offered under other specifications will be considered, provided that the bid clearly states the offered commodities are substantially superior to those requirements in the above specification. (By descriptive literature, illustrations, etc.)
- ✓ Demonstration of same or similar system in Civil/Geotechnical laboratory (preferably in IIT's) during the course of technical evaluation.
- ✓ User List preferably in civil/structural laboratory with contact person name, e-mail ids and direct contact number, user certificate for satisfactory installation and workmanship along with signature and stamp to be provided as testimonials, the institute reserves the right to call each individual person for reference.
- ✓ PO's of same/similar items should be provided along-with the technical bids
- ✓ Bidder should have a sound bank balance, annual report for the last five years to be provided.
- ✓ The supplier must be ISO 9000 series certificated.
- ✓ Service Support: The Vendor shall define the service support plan in India and shall utilize either factory direct service or through factory trained service personnel located in India with a minimum of five service engineers on staff for servo-hydraulic

application. A commitment to the local community shall be evidenced by the existence of this service support for at least 10 years.

- ✓ One year system warranty after acceptance
- ✓ Spares and consumables to be quoted
- ✓ PO's of same/similar items should be provided along-with the technical bid

A complete set of tender documents* may be Download by prospective bidder free of cost from the website http://eprocure.gov.in/eprocure/app. Bidder has to make payment of requisite fees (i.e. Tender fees (if any) and EMD) online through RTGS/NEFT only.

Terms & Conditions Details

Sl.No.	Specification
1.	Due date : The tender has to be submitted on-line before the due date. The offers received after the
	due date and time will not be considered. No manual bids will be considered.
2.	Preparation of Bids: The offer/bid should be submitted in two bid systems (i.e.) Technical bid and financial bid. The technical bid should consist of all technical details along with commercial terms and conditions. Financial Bids to be submitted in PDF format. The Technical bid and the financial bid should be submitted Online.
3.	EMD (if applicable): The tenderer should submit an EMD amount through RTGS/NEFT. The
3.	Technical Bid without EMD would be considered as UNRESPONSIVE and will not be accepted. The EMD will be refunded without any interest to the unsuccessful bidders after the award of contract. Refer to Schedule (at page 1 of this document) for its actual place of submission.
4.	Refund of EMD : The EMD will be returned to unsuccessful Tenderer only after the Tenders are finalized. In case of successful Tenderer, it will be retained till the successful and complete installation of the equipment.
5.	Opening of the tender: The online bid will be opened by a committee duly constituted for this purpose. Online bids (complete in all respect) received along with EMD (if any) will be opened as mentioned at "Annexure: Schedule" in presence of bidders representative if available. Only one representative will be allowed to participate in the tender opening. Bid received without EMD (if present) will be rejected straight way. The technical bid will be opened online first and it will be examined by a technical committee (as per specification and requirement). The financial offer/bid will be opened only for the offer/bid which technically meets all requirements as per the specification, and will be opened in the presence of the vendor's representatives subsequently for further evaluation. The bidders if interested may participate on the tender opening Date and Time. The bidder should produce authorization letter from their company to participate in the tender opening.
6.	Acceptance/ Rejection of bids: The Committee reserves the right to reject any or all offers without assigning any reason.
7.	Pre-qualification criteria: (i) Bidders should be the manufacturer / authorized dealer. Letter of Authorization from original equipment manufacturer (OEM) on the same and specific to the tender should be enclosed. (ii) An undertaking from the OEM is required stating that they would facilitate the bidder on a regular basis with technology/product updates and extend support for the warranty as well. (Ref. Annexure-II) (iii) OEM should be internationally reputed Branded Company. (iv) Non-compliance of tender terms, non-submission of required documents, lack of clarity of the specifications, contradiction between bidder specification and supporting documents etc. may lead to rejection of the bid. (v) In the tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEMitself can bid but both cannot bid simultaneously for the same item/product in the same tender. (vi) If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product. (vii) Vendor should have supplied at least 3 nos. of similar equipment within India (preferably in IIT's) and abroad. A detailed reference of supply of equipment must be provided along with the bids. Any negative comments from any one referred would disqualify the bid. IIT Delhi reserves the right to interact/ visit with the referred customer as per its convenience.
	(viii) Installation, Commissioning and Terms to IIT Delhi: The cost should include fabrication, delivery, unloading at IIT Delhi site, installation, deputation of competent engineers for installation and systems required for smooth running of the equipment. Interested bidders are

- welcome to see the space availability at the Foundation Engineering Laboratory, IIT Delhi by taking prior appointment.
- (ix) Vendor is required to supply, fabricate, install and ensure proper commissioning of the equipment within 24 weeks of the Supply order.
- (x) The supplier should demonstrate the performance of the equipment of the specifications by conducting trial tests at the Foundation Engineering Laboratory, IIT Delhi. Complete set of Manuals for operation, maintenance and safety should be provided.
- 8. **Performance Security**: The supplier shall require to submit the performance security in the form of irrevocable bank guarantee issued by any Indian Nationalized Bank for an amount which is stated at page #1 of the tender document within 21 days from the date of receipt of the purchase order/LC and should be kept valid for a period of 60 days beyond the date of completion of warranty period.
- 9. **Force Majeure:** The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
 - For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
 - If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 10. **Risk Purchase Clause**: In event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from the other source on the total risk of the supplier under risk purchase clause.
- 11. **Packing Instructions**: Each package will be marked on three sides with proper paint/indelible ink, the following:
 - i. Item Nomenclature
 - ii. Order/Contract No.
 - iii. Country of Origin of Goods
 - iv. Supplier's Name and Address
 - v. Consignee details
 - vi. Packing list reference number

12. **Delivery and Documents:**

Delivery of the goods should be made within a maximum of 08 to 16 weeks from the date of placement of purchase order and the opening of LC. Within 24 hours of shipment, the supplier shall notify the purchaser and the insurance company by cable/telex/fax/e mail the full details of the shipment including contract number, railway receipt number/ AAP etc. and date, description of goods, quantity, name of the consignee, invoice etc. The supplier shall mail the following documents to the purchaser with a copy to the insurance company:

- 1. 4 Copies of the Supplier invoice showing contract number, goods' description, quantity
- 2. unit price, total amount;
- 3. Insurance Certificate if applicable;
- 4. Manufacturer's/Supplier's warranty certificate;
- 5. Inspection Certificate issued by the nominated inspection agency, if any
- 6. Supplier's factory inspection report; and
- 7. Certificate of Origin (if possible by the beneficiary);

8. Two copies of the packing list identifying the contents of each package. The above documents should be received by the Purchaser before arrival of the Goods (except where the Goods have been delivered directly to the Consignee with all documents) and, if not received, the Supplier will be responsible for any consequent expenses. 13. **Delayed delivery:** If the delivery is not made within the due date for any reason, the Committee will have the right to impose penalty 1% per week and the maximum deduction is 10% of the contract value / price. Prices: The price should be quoted in net per unit (after breakup) and must include all packing 14. and delivery charges. The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. However the percentage of taxes & duties shall be clearly indicated. The price should be quoted without custom duty and excise duty, since IIT Delhi is exempted from payment of Excise Duty and is eligible for concessional rate of custom duty. Necessary certificate will be issued on demand. In case of imports, the price should be quoted on FOB Basis only. Under special circumstances (e.g. perishable chemicals), when the item is imported on CIF/CIP, please indicate CIF/CIP charges separately up to IIT Delhi indicating the mode of shipment. IIT Delhi will make necessary arrangements for the clearance of imported goods at the Airport/Seaport. Hence the price should not include the above charges. Notices: For the purpose of all notices, the following shall be the address of the Purchaser and 15. Supplier. Purchaser:Dr. Bappaditya Manna Civil Engineering Department Indian Institute of Technology Hauz Khas, New Delhi - 110016. **Supplier:** (To be filled in by the supplier) (All suppliers should submit its supplies information as per Annexure-II). **Progress of Supply**: Wherever applicable, supplier shall regularly intimate progress of supply, in 16. writing, to the Purchaser as under: 1. Quantity offered for inspection and date; 2. Quantity accepted/rejected by inspecting agency and date; 3. Quantity dispatched/delivered to consignees and date; 4. Quantity where incidental services have been satisfactorily completed with date; 5. Quantity where rectification/repair/replacement effected/completed on receipt of any communication from consignee/Purchaser with date; 6. Date of completion of entire Contract including incidental services, if any; and 7. Date of receipt of entire payments under the Contract (In case of stage-wise inspection, details required may also be specified). 17. **Inspection and Tests:** Inspection and tests prior to shipment of Goods and at final acceptance are as After the goods are manufactured and assembled, inspection and testing of the goods shall be carried out at the supplier's plant by the supplier, prior to shipment to check whether the goods are in conformity with the technical specifications attached to the purchase order. Manufacturer's test certificate with data sheet shall be issued to this effect and submitted along with the delivery documents. The purchaser shall be present at the supplier's premises during such inspection and testing if need is felt. The location where the inspection is required to be conducted should be clearly indicated. The supplier shall inform the purchaser about the site preparation, if any, needed for installation of the goods at the purchaser's site at the time of

- submission of order acceptance.
- The acceptance test will be conducted by the Purchaser, their consultant or other such person nominated by the Purchaser at its option after the equipment is installed at purchaser's site in the presence of supplier's representatives. The acceptance will involve trouble free operation and ascertaining conformity with the ordered specifications and quality. There shall not be any additional charges for carrying out acceptance test. No malfunction, partial or complete failure of any part of the equipment is expected to occur. The Supplier shall maintain necessary log in respect of the result of the test to establish to the entire satisfaction of the Purchaser, the successful completion of the test specified.
- In the event of the ordered item failing to pass the acceptance test, a period not exceeding one weeks will be given to rectify the defects and clear the acceptance test, failing which the Purchaser reserve the right to get the equipment replaced by the Supplier at no extra cost to the Purchaser.
- Successful conduct and conclusion of the acceptance test for the installed goods and equipment shall also be the responsibility and at the cost of the Supplier.
- **Resolution of Disputes**: The dispute resolution mechanism to be applied pursuant shall be as follows:
 - In case of Dispute or difference arising between the Purchaser and a domestic supplier relating to any matter arising out of or connected with this agreement, such disputes or difference shall be settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules there under and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Director, Indian Institute of Technology (IIT) Delhi and if he is unable or unwilling to act, to the sole arbitration of some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order.
 - In the case of a dispute between the purchaser and a Foreign Supplier, the dispute shall be settled by arbitration in accordance with provision of sub-clause (a) above. But if this is not acceptable to the supplier then the dispute shall be settled in accordance with provisions of UNCITRAL (United Nations Commission on International Trade Law) Arbitration Rules.
 - The venue of the arbitration shall be the place from where the order is issued.
- 19. **Applicable Law:** The place of jurisdiction would be New Delhi (Delhi) INDIA.

20. **Right to Use Defective Goods**

If after delivery, acceptance and installation and within the guarantee and warranty period, the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.

21. **Supplier Integrity**

The Supplier is responsible for and obliged to conduct all contracted activities in accordance with the Contract using state of the art methods and economic principles and exercising all means available to achieve the performance specified in the contract.

22. **Training**

The Supplier is required to provide training to the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.

23. **Installation & Demonstration**

The supplier is required to done the installation and demonstration of the equipment within one month of the arrival of materials at the IITD site of installation, otherwise the penalty clause will be the same as per the supply of materials.

In case of any mishappening/damage to equipment and supplies during the carriage of supplies from the origin of equipment to the installation site, the supplier has to replace it with new equipment/supplies immediately at his own risk. Supplier will settle his claim with the insurance

company as per his convenience. IITD will not be liable to any type of losses in any form. 24. **Insurance:** For delivery of goods at the purchaser's premises, the insurance shall be obtained by the supplier in an amount equal to 110% of the value of the goods from "warehouse to warehouse" (final destinations) on "All Risks" basis including War Risks and Strikes. The insurance shall be valid for a period of not less than 3 months after installation and commissioning. In case of orders placed on FOB/FCA basis, the purchaser shall arrange Insurance. If orders placed on CIF/CIP basis, the insurance should be up to IIT Delhi. 25. **Incidental services:** The incidental services also include: • Furnishing of 01 set of detailed operations & maintenance manual. Arranging the shifting/moving of the item to their location of final installation within IITD premises at the cost of Supplier through their Indian representatives. 26. Warranty: Warranty period shall be (as stated at page #2 of this tender) from date of installation of Goods at the IITD site of installation. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the contract. If for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall at its discretion make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests. The warranty should be comprehensive on site. (ii) The Purchaser shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall immediately within in 02 days arrange to repair or replace the defective goods or parts thereof free of cost at the ultimate The Supplier shall take over the replaced parts/goods at the time of their replacement. No claim whatsoever shall lie on the Purchaser for the replaced parts/goods thereafter. The period for correction of defects in the warranty period is 02 days. If the supplier having been notified fails to remedy the defects within 02 days, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expenses and without prejudice to any other rights, which the purchaser may have against the supplier under the contract. (iii) The warranty period should be clearly mentioned. The maintenance charges (AMC) under different schemes after the expiry of the warranty should also be mentioned. The warranty will commence from the date of the satisfactory installation/commissioning of the equipment against the defect of any manufacturing, workmanship and poor quality of the components. (iv) After the warranty period is over, Annual Maintenance Contract (AMC)/Comprehensive Maintenance Contract (CMC) up to next two years should be started. The AMC/CMC charges will not be included in computing the total cost of the equipment. 27. **Governing Language** The contract shall be written in English language. English language version of the Contract shall govern its interpretation. All correspondence and other documents pertaining to the Contract, which are exchanged by the parties, shall be written in the same language. 28. Applicable Law The Contract shall be interpreted in accordance with the laws of the Union of India and all disputes shall be subject to place of jurisdiction. 29. **Notices** • Any notice given by one party to the other pursuant to this contract/order shall be sent to the other party in writing or by cable, telex, FAX or e mail and confirmed in writing to the other party's address. A notice shall be effective when delivered or on the notice's effective date, whichever is

20	T
30.	Taxes Suppliers shall be entirely responsible for all taxes, duties, license fees, octroi, road permits, etc., incurred until delivery of the contracted Goods to the Purchaser. However, VAT in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order.
31.	Duties IIT Delhi is exempted from paying custom duty under notification No.51/96 (partially orfull) and necessary "Custom Duty Exemption Certificate" can be issued after providing following information and Custom Duty Exemption Certificate will be issued to the shipment in the name of the Institute, no certificate will be issued to third party: a) Shipping details i.e. Master Airway Bill No. and House Airway No. (if exists) b) Forwarder details i.e. Name, Contact No., etc.
	IIT Delhi is exempted from paying Excise Duty and necessary Excise Duty ExemptionCertificate will be provided for which following information are required. b) Quotation with details of Basic Price, Rate, Tax & Amount on which ED is applicable c) Supply Order Copy d) Performa-Invoice Copy.
32.	Agency Commission: Agency commission if any will be paid to the Indian agent in Rupees on receipt of the equipment and after satisfactory installation. Agency Commission will not be paid in foreign currency under any circumstances. The details should be explicitly shown in Tender even in case of Nil commission. The tenderer should indicate the percentage of agency commission to be paid to the Indian agent.
33.	 Payment: For imported items Payment will be made through irrevocable Letter of Credit (LC). Letter of Credit (LC) will be established in favor of foreign Supplier after the submission of performance security. The letter of credit (LC) will be established on the exchange rates as applicable on the date of establishment. For Imports, LC will be opened for 100% FOB/CIF value. 80% of the LC amount shall be released on presentation of complete and clear shipping documents and 20% of the LC amount shall be released after the installation and demonstration of the equipment at the INST site of installation in faultless working condition for period of 60 days from the date of the satisfactory installation and subject to the production of unconditional performance bank guarantee as specified in Clause 8 of tender terms and conditions. For Indigenous supplies, 100% payment shall be made by the Purchaser against delivery, inspection, successful installation, commissioning and acceptance of the equipment at IITD in good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in Clause 9 of tender terms and conditions. Indian Agency commission (IAC), if any shall be paid after satisfactory installation & commissioning of the goods at the destination at the exchange rate prevailing on the date of negotiation of LC documents, subject to DGS&D registration for restricted items. All the bank charges within India will be borne by the Institute and outside India will be borne by the Supplier.
34.	User list: Brochure detailing technical specifications and performance, list of industrial and educational establishments where the items enquired have been supplied must be provided. (Ref. Annexure-III)
35.	Manuals and Drawings
33.	(i) Before the goods and equipment are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.

- The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract. (iii) Unless and otherwise agreed, the goods equipment shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser. Application Specialist: The Tenderer should mention in the Techno-Commercial bid the 36. availability and names of Application Specialist and Service Engineers in the nearest regional office. (Ref. to Annexure-III) 37. **Site Preparation**: The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter. The supplier shall visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre-installation requirements. 38. **Spare Parts** The Supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier: ii. Such spare parts as the Purchaser may elect to purchase from the Supplier, providing that this election shall not relieve the Supplier of any warranty obligations under the Contract; and iii. In the event of termination of production of the spare parts: iv. Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements; and v. Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if requested.

Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the Goods, such as gaskets, plugs, washers, belts etc. Other spare parts and components shall be supplied as promptly as possible but in any case within six months of placement of order.

Defective Equipment: If any of the equipment supplied by the Tenderer is found to be 39. substandard, refurbished, un-merchantable or not in accordance with the description/specification or otherwise faulty, the committee will have the right to reject the equipment or its part. The prices of such equipment shall be refunded by the Tenderer with 18% interest if such payments for such equipment have already been made. All damaged or unapproved goods shall be returned at suppliers cost and risk and the incidental expenses incurred thereon shall be recovered from the supplier. Defective part in equipment, if found before installation and/or during warranty period, shall be replaced within 45 days on receipt of the intimation from this office at the cost and risk of supplier including all other charges. In case supplier fails to replace above item as per above terms & conditions, IIT Delhi may consider "Banning" the supplier.

Termination for Default 40.

The Purchaser may, without prejudice to any other remedy for breach of contract, by written notice of default sent to the Supplier, terminate the Contract in whole or part:

- i. If the Supplier fails to deliver any or all of the Goods within the period(s) specified in the order, or within any extension thereof granted by the Purchaser; or
- ii If the Supplier fails to perform any other obligation(s) under the Contract.
- iii If the Supplier, in the judgment of the Purchaser has engaged in corrupt or fraudulent practices in competing for or in executing the Contract.
- For the purpose of this Clause:
 - "Corrupt practice" means the offering, giving, receiving or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution.

	ii. "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Borrower, and includes collusive practice among Bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the Borrower of the benefits of free and open competition;"
	• In the event the Purchaser terminates the Contract in whole or in part, the Purchaser may procure, upon such terms and in such manner, as it deems appropriate, Goods or Services similar to those undelivered, and the Supplier shall be liable to the Purchaser for any excess costs for such similar Goods or Services. However, the Supplier shall continue the performance of the Contract to the extent not terminated.
41.	Shifting : After 1-2 years once our new Academic Block will be ready, the supplier has to shift and reinstall the instrument free of cost (if required).
42.	Downtime: During the warranty period not more than 5% downtime will be permissible. For every day exceeding permissible downtime, penalty of 1/365 of the 5% FOB value will be imposed. Downtime will be counted from the date and time of the filing of complaint with in the business hours.
43.	Training of Personnel: The supplier shall be required to undertake to provide the technical training to the personnel involved in the use of the equipment at the Institute premises, immediately after completing the installation of the equipment for a minimum period of one week at the supplier's cost.
44.	Disputes and Jurisdiction : Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within New Delhi.
45.	Compliancy certificate : This certificate must be provided indicating conformity to the technical specifications. (Annexure-I)
46.	"In case of CIF/CIP shipments, kindly provide the shipment information at least 2 days in advance before landing the shipment along with the documents i.e. invoice, packing list, forwarder Name, address, contact No. in India to save demurrage charges (imposed by Indian Customs). Otherwise these charges will be recovered from the supplier/Indian Agent."

COMPLIANCE SHEET

TECHNICAL SPECIFICATION

I have also enclosed all relevant documents in support of my claims, (as above) in the following pages.

S. No	Technical Specifica	tions		COMPLIANCE Y/N
	DATORY ITEMS:			1/11
		aulic Actuator 100kN(01 No)	
1	<u>. </u>	`	combined balanced dynamic performance	
2			50 mm(+/- 75 mm double amplitude)	
3	Full-stroke, coaxially amplitude)	y-mounted displacement	t transducer 150 mm (+/- 75 mm double	
4	<u> </u>	etallic Polymerbearings h	bonded directly to the end caps	
5	<u> </u>	<u> </u>	of heat-treated alloy steel and surface	
6	Cushions to protect t	he actuator from the effe	ects of high-speed and high-mass forces	
7			modate dual Servo-valves with /without	
8	One servo-valve of 6 criteria)	3 lpm (minimum) capac	city(or to meet the below performance	
9	of 150% of the rated	l load; Non-linearity: 0	nator capacity) with an overload capacity 0.08% of full scale or better; Hysteresis: 0.03% of full scale or better	
10	Performance Criteria	(100 kN actuator):		
	Force	Double Amplitude	Frequency	
	+/- 90 KN	1.5 mm ±0.25 mm	20 Hz	
	(90% capacity)	80 mm ±10 mm	1 Hz	
	+/- 75 KN	1.5 mm ± 0.25 mm	20 Hz	
	(75% capacity)	90 mm ±10 mm	1 Hz	
	+/- 50 KN	0.35 mm ±0.05 mm	30 Hz	
	(50% capacity)	100 mm ±10 mm	1 Hz	
	+/- 25 KN	0.3 mm ±0.05 mm	40 Hz	
	(25% capacity)	100 mm ±10 mm	1 Hz	
11	Performance curve o	f the complete system sh	nould be provided	
12	One Swivel Head and One Swivel Base: Adjustable bearing for clearance to eliminate backlash; Dynamic force rating: ±160 kN or more(for 100 kN actuator); Tilt angle: ±17° or more (for head and base); Swivel angle: +90°/-75° or more; For use in cyclic, reversing load application(for head and base)			
13	Accuracy: ±1% above the threshold value of 10% of full-scale of sensors			
14	•		should be factory calibrated to full range	

(2) S	ervo-controlled	Hydraulic Actuator 5	0kN (01 No)			
15	Double-ended,	· ·	gue-rated for combine	d balanced dy	namic	
	performance					
16	Rated load cap	acity:50kN; Rated strol	ke: 150 mm(+/- 75 mm de	ouble amplitude)		
17	Full-stroke, co	axially-mounted displa	cement transducer 150 n	nm (+/- 75 mm d	louble	
	amplitude).					
18			bearings bonded directly			
19		achined from a single	e piece of heat-treated a	ılloy steel and sı	urface	
	coated					
20			the effects of high-speed			
21		ld have provision to	accommodate dual Serve	o-valves with /w	ithout	
22	port shut-offs	vo of 62 lpm (minimu	m) capacity(or to meet t	ha halavy manfann	manaa	
22	criteria)	ve of 63 ipin (minimu	iii) capacity(or to meet t	ne below perion	nance	
23		Axial Load Cell (for ea	ch actuator capacity) wit	h an overload car	nacity	
23		`	rity: 0.08% of full scale	-		
			oility: 0.03% of full scale			
22		riteria (50kN actuator)				
		Force	Double Amplitude	Frequency		
		+/- 45 KN	$4.0 \text{ mm} \pm 0.5 \text{ mm}$	20 Hz		
		(90% capacity)	150 mm ± 15mm	1 Hz		
		+/- 37.5 KN	4.0 mm ± 0.5 mm	20 Hz		
		(75% capacity)	150 mm ± 15 mm	1 Hz		
		+/- 25 KN	1.5 mm ± 0.25 mm	30 Hz		
		(50% capacity)	150 mm ± 15 mm	1 Hz		
		+/- 12.5 KN	1.0 mm ± 0.20 mm	40 Hz		
		(25% capacity)	150 mm ± 15 mm	1 Hz		
23	Performance cu	arve of the complete sy	stem should be provided.			
24	One Swivel He	ead and One Swivel Bas	se: Adjustable bearing for	r clearance to elin	ninate	
			kN or more(for 50 kN act		±17°	
	`	, ,	ngle: $+90^{\circ}/-75^{\circ}$ or more;	For use in cyclic,		
25		application(for head an		C		
25	Accuracy: ±1% above the threshold value of 10% of full-scale of sensors				wan 32	
26 (3) E	Both displacement transducer and Load Cell should be factory calibrated to full range Hydraulic Power Unit (01 No)					
(3) E 27	T .		anacity: 100 lpm or highe	r		
28	Variable-volume pump, Rated flow capacity: 100 lpm or higher Foot print must be less than 2 m x 1m					
29	210 bar pressure, and 415 V AC, 50 Hz, 3-phase power					
30	Direct-coupled 60 HP submerged motor providing minimum 91% operating					
21	efficiency		1 2 2	tat at		
31			less measured at 3 feet w		rating	
32	at high pressure and dumping the full flow over the relief valve Corrosion-resistant stainless steel oil-to-water heat exchanger to dissipate generated				erated	
J <u>~</u>	heat					
33	Submersed pump/motor design to eliminate the release of ambient heat					
	= ==================================		are release of affile			

consumption 50 Cabinet with lockable cover to reduce noise and to prevent tampering of controls 51 Full flow 3-micron absolute filter in the return line to provide excellent oil cleaning 52 Integrated reservoir, pump/motor, heat exchanger 53 Approved electrical disconnect interrupts power to the unit when the door is opened 54 Interlock circuitry for over-temperature and low-fluid level protection 55 Witches for start, low/high pressure, and stop on the front panel of the enclosure 56 High quality face seal fittings minimize the potential for leakage outside of the reservoir 57 Indicators for power on, low oil level, over temperature, and dirty filter warnings 58 Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each 58 J. Hydratid (two Stations): 58 J. Hydratid (two Stations): 59 All each allow the actuators to be independently controlled 50 Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel 50 Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel 50 Fessure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop 50 J. Bitiers (I Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 51 Digital Controller (01 No) 52 Digital Controller (01 No) 53 Douglat serve control, function generation, data acquisition, hydraulic control, and digital I/O capability 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Punction generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (Dypically load, strain or displacement) including load limited displacement during sp	34	Temperature controlled water-saver and water shut-off valves to minimize water	
Cabinet with lockable cover to reduce noise and to prevent tampering of controls		<u> </u>	
Full flow 3-micron absolute filter in the return line to provide excellent oil cleaning The theretad reservoir, pump/motor, heat exchanger The theretad disconnect interrupts power to the unit when the door is opened Interlock circuitry for over-temperature and low-fluid level protection Switches for start, low-high pressure, and stop on the front panel of the enclosure High quality face seal fittings minimize the potential for leakage outside of the reservoir Indicators for power on, low oil level, over temperature, and dirty filter warnings Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each 3. D Hydraulid (two Stations): All Maximum operating pressure: 21 MPa (3000 PSI) Shall each allow the actuators to be independently controlled Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel Fast emergency unload for system depressurization Pessure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop Flow rating: 114 LPM (30 GPM) or more 3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 4D Digital Controller (01 No) Digital Servo control, function generation, data acquisition, hydraulic control, and digital MO capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators Should have two stations to control two actuators Should have two stations to control two actuators Digital Control (DPC) bandwidth update rate: 6 kHz or faster Digital Control (DPC) bandwidth update rate: 6 kHz or faster Punction generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Direct Digital Control (DPC) bandwidth update rate: 6 kHz or faster Dinator of displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog	35	1	
Integrated reservoir, pump/motor, heat exchanger Approved electrical disconnect interrupts power to the unit when the door is opened Interlock circuitry for over-temperature and low-fluid level protection Switches for start, low/high pressure, and stop on the front panel of the enclosure High quality face seal fittings minimize the potential for leakage outside of the reservoir Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power on, low oil level, over temperature, and dirty filter warnings Indicators for power oil level, over temperature, and dirty filter warnings Indicators for power oil level, over temperature, and dirty filter varnings Indicators for power oil level, over temperature, and dirty indicators Indicators for power filters for p	36		
Approved electrical disconnect interrupts power to the unit when the door is opened linterlock circuitry for over-temperature and low-fluid level protection linterlock circuitry for over-temperature and low-fluid level protection linterlock circuitry for over-temperature, and stop on the front panel of the enclosure High quality face seal fittings minimize the potential for leakage outside of the reservoir lindicators for power on, low oil level, over temperature, and dirty filter warnings lindicators for power on, low oil level, over temperature, and dirty filter warnings lindicators for power on, low oil level, over temperature, and dirty filter warnings lindicators for power on, low oil level, over temperature, and dirty filter warnings lindicators for power on, low oil level, over temperature, and dirty filter warnings lindicators for power of the minimum lindicators on/off control of each channel on/off control of each channel on/off control of each channel provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop provided to prevent specimen damage caused by abnormal hydraulic control, and digital Poo apability of the power start of the filter for lindicator of specimen damage in minimum lindicator of specimen start of the filter for data acquisition, hydraulic control, specimen start of the filter for d	37	· ·	
Interlock circuitry for over-temperature and low-fluid level protection	38		
High quality face seal fittings minimize the potential for leakage outside of the reservoir Indicators for power on, low oil level, over temperature, and dirty filter warnings Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each 3. I) Hydraulic Service Manifold (two Stations): 4 Maximum operating pressure: 21 MPa (3000 PSI) Shall each allow the actuators to be independently controlled Solatonio valves should be included on each actuator channel to allow independent on/off control of each channel Fast emergency unload for system depressurization Past emergency unload for system depressurization Fast emergency unload for system depressurization Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop Flow rating: 114 LPM (30 GPM) or more 3. Siters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 4) Digital Controller (01 No) Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Direct Digital conditioner data sample rate: 100 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controled transducer imits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load imitted displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Compected with high-speed serial interface with personal computer,	39	••	
High quality face seal fittings minimize the potential for leakage outside of the reservoir Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each Hose pipes for actuators the pressure of 15 m long each Hose pipes for actuators the pressure of 15 m long each Hose pipes for actuators to be independently pentrolled Maximum operating pressure: 21 MPa (3000 PSI) Shall each allow the actuators to be independently controlled Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel Fast emergency unload for system depressurization Fast emergency unload for system depressurization Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop Flow rating: 114 LPM (30 GPM) or more Bigital Controller (01 No) Digital Controller (01 No) Digital Servo control, function generation, data acquisition, hydraulic control, and digital I/O capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system C	40	Switches for start, low/high pressure, and stop on the front panel of the enclosure	
Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each	41		
Hose pipes for actuator: Pressure, Return and Drain pipes of 15 m long each	42	Indicators for power on, low oil level, over temperature, and dirty filter warnings	
3.1 Hydraulic Service Manifold (two Stations): 44 Maximum operating pressure: 21 MPa (3000 PSI) 45 Shall each allow the actuators to be independently controlled 46 Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel 47 Fast emergency unload for system depressurization 48 Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop 49 Flow rating: 114 LPM (30 GPM) or more 50 3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 40 Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability 52 The controller shall be expandable maximum up to 4 channels 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraires for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the			
45 Shall each allow the actuators to be independently controlled 46 Solenoid valves should be included on each actuator channel to allow independent 47 Fast emergency unload for system depressurization 48 Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic 49 on/off and inspection start/stop 49 Flow rating: 114 LPM (30 GPM) or more 50 3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or 50 more, Filter: 10 micron 4) Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and 52 digital I/O capability 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp 57 waveforms and downloaded wave shapes 58 Automated dynamic control mode switching between any connected transducer. Any 58 connected transducer or calculation can be selected for control (typically load, strain 59 or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access 50 to transducer signals and other critical parameters 50 Ability to save and restore PID tuning settings 51 Adoptive controls compensation: Peak-Valley and Null Pacing 52 Connected with high-speed serial interface with personal computer, No parallel 53 interface is acceptable 54 Should be available for use with C++, Visual basic, and Lab View 55 Online UPS should be included and should communicate with the controller 56 Cables for hydraulic control, servovalves, conditioners, and communications to 57 personal computer should be 15 m long 58 One A/D CARD: 8 channel analogue input per controller for data acquisition & 58 feedback control	3. 1) H		
Solenoid valves should be included on each actuator channel to allow independent on/off control of each channel		Maximum operating pressure: 21 MPa (3000 PSI)	
on/off control of each channel 47 Fast emergency unload for system depressurization 48 Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop 49 Flow rating: 114 LPM (30 GPM) or more 50 3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 4) Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability 52 The controller shall be expandable maximum up to 4 channels 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control	45	Shall each allow the actuators to be independently controlled	
Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop	46	_	
Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic on/off and inspection start/stop	47	Fast emergency unload for system depressurization	
Flow rating: 114 LPM (30 GPM) or more 3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 4) Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability 52 The controller shall be expandable maximum up to 4 channels 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	48	Pressure shall be provided to prevent specimen damage caused by abnormal hydraulic	
3.8 liters (1 Gallon) pressure and 1 liter (0.26 Gallon) return line accumulator or more, Filter: 10 micron 4) Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability 52 The controller shall be expandable maximum up to 4 channels 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	40	1	
more, Filter: 10 micron 4) Digital Controller (01 No) 51 Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability 52 The controller shall be expandable maximum up to 4 channels 53 Should have two stations to control two actuators 54 Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster 55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card			
Digital servo control, function generation, data acquisition, hydraulic control, and digital I/O capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Signal conditioner data sample rate: 100 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system About the variable of the savailable for use with C++, Visual basic, and Lab View Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long Cables for hydraulic control, servovalves, conditioners, and communication & feedback control Ali Digital Universal Conditioner/Valve Driver Card	50		
digital I/O capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators birect Digital Control (DDC) bandwidth update rate: 6 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Abould be available for use with C++, Visual basic, and Lab View Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control Jugital Universal Conditioner/Valve Driver Card	4) Dig	gital Controller (01 No)	
digital I/O capability The controller shall be expandable maximum up to 4 channels Should have two stations to control two actuators birect Digital Control (DDC) bandwidth update rate: 6 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Abould be available for use with C++, Visual basic, and Lab View Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control Jugital Universal Conditioner/Valve Driver Card	51	Digital servo control, function generation, data acquisition, hydraulic control, and	
Should have two stations to control two actuators Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Signal conditioner data sample rate: 100 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system About the available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control Jujetal Universal Conditioner/Valve Driver Card			
Direct Digital Control (DDC) bandwidth update rate: 6 kHz or faster Signal conditioner data sample rate: 100 kHz or faster Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Abould be available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control	52	The controller shall be expandable maximum up to 4 channels	
55 Signal conditioner data sample rate: 100 kHz or faster 56 Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		Should have two stations to control two actuators	
Function generation by 32 bit processor, standard haversine, square, and ramp waveforms and downloaded wave shapes Computer controlled transducer limits Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Should be available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	54		
waveforms and downloaded wave shapes 57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	55	Signal conditioner data sample rate: 100 kHz or faster	
57 Computer controlled transducer limits 58 Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading 59 Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	56		
Automated dynamic control mode switching between any connected transducer. Any connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Should be available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control Jipital Universal Conditioner/Valve Driver Card	57	_	
connected transducer or calculation can be selected for control (typically load, strain or displacement) including load limited displacement during specimen loading Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters Ability to save and restore PID tuning settings Adaptive controls compensation: Peak-Valley and Null Pacing Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Should be available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 1.1) Digital Universal Conditioner/Valve Driver Card		1	
Three computer-selectable channels of 16 bit resolution analog output for easy access to transducer signals and other critical parameters 60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		connected transducer or calculation can be selected for control (typically load, strain	
60 Ability to save and restore PID tuning settings 61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	59	Three computer-selectable channels of 16 bit resolution analog output for easy access	
61 Adaptive controls compensation: Peak-Valley and Null Pacing 62 Connected with high-speed serial interface with personal computer, No parallel interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		ŭ i	
Connected with high-speed serial interface with personal computer, No parallel interface is acceptable Programmable libraries for real time control of system Should be available for use with C++, Visual basic, and Lab View Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 1.1) Digital Universal Conditioner/Valve Driver Card		·	
interface is acceptable 63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		· · · · · · · · · · · · · · · · · · ·	
63 Programmable libraries for real time control of system 64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	62		
64 Should be available for use with C++, Visual basic, and Lab View 65 Online UPS should be included and should communicate with the controller 66 Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		*	
Online UPS should be included and should communicate with the controller Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card		· ·	
Cables for hydraulic control, servovalves, conditioners, and communications to personal computer should be 15 m long One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card			
personal computer should be 15 m long 67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card			
67 One A/D CARD: 8 channel analogue input per controller for data acquisition & feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	66	<u> </u>	
feedback control 4.1) Digital Universal Conditioner/Valve Driver Card	67		
4.1) Digital Universal Conditioner/Valve Driver Card			
The conditioner shall support transducers	4.1) Di	gital Universal Conditioner/Valve Driver Card	
	68	The conditioner shall support transducers	

69	Portability of transducer calibration data between conditioners	
70	Conditioners interface to DC or AC transducers and eliminate the need for separate	
	conditioner types	
71	Low-noise, low-drift, high- accuracy signal conditioning, and shunt calibration	
72	On-board strain gauge bridge completion	
73	Excitation sensing and loss detection	
74	Processor-controlled limit detection	
75	Variable excitation frequencies for a wide variety of transducer types	
76	Cards configurable to drive a two-stage valve and condition a transducer, or to drive a three-stage valve	
4 2) Si	ystem Operating Software	
77	A software interface for configuring the controller and the user interface to the test	
	station	
78	A two channel on line data display (X vs. Y or Time vs. YY), digital displays, and a	
	system exerciser for setting up tuning parameters and warming up the system prior to	
	testing	
79	Easy program for simple monotonic and cycle test execution including data acquisition.	
80	Test data may be stored in choice of ASCII, Lotus, or Excel formats for analysis with	
	your favourite tools	
81	Null pacing adaptive control algorithm for use with ramp command signals	
82	Adaptive phase and amplitude algorithm for use with (constant/block) sine wave	
	command signals	
83	Perform amplitude control to ensure desired amplitudes are achieved	
84	Data acquisition (timed, P/V, Level Crossing, Cyclic/Logarithmic)	
85	Function generation up to 600 Hz on all channels; Sine, square, triangle, ramp, hold,	
	processes, and ability to play digitized profiles	
86	Soft start/stop available on all channels	
87	Amplitude Phase Control S/W to adjust the amplitude and phase of the command to	
	realize the desired feedback signal	
88	Access to test status information from any web-enabled, internet-connected device	
89	System Views - Provide information about a single test system or multiple test system.	
90	Lab Views - Provide information about all the test status of your lab from anywhere	
	at any time	
91	Security - Data transfers use SSL/TLS 256-bit encryption	
92	User Access Control - Provides ability to create individual passwords for a page, or	
	limit access to users	
93	Alerts - Allow for email, text message or Tweet notification to interested parties when	
	the state of a test changes	
	eneral Application Software	
94	Shall include the ability to control or to capture data from any installed AC or DC	
	transducer, or externally conditioned transducer	
95	Shall allow detection of digital inputs to the test system controller and allow digital	
	output signals to be used as output from one of the control system digital output channels	
96	The application software shall allow the following kinds of tests to be defined	
	and executed	
97	(1) Block loading fatigue	
98	(2) Constant amplitude fatigue	
99	(3) Random fatigue using an input files to define end levels and rates any of the	

	above in combination	
100	(4) Graphical drag - and - drop test layout design	
101	(5) Test template creation	
102	(6) System software application + custom waveform activities	
103	(7) Parallel branches for test execution and logical operators (if/then, while)	
104	(8) Limit sensing, sequencing triggers, and interface to digital I/O	
105	(9) Report generation while the test is running	
106	Shall allow the operator to selectively acquire data periodically as defined by	
100	logarithms, linear intervals or user selected intervals. This data shall be collected in	
	time, level crossing or peak and valley samples	
107	Shall allow a virtually unlimited number of test procedures to be set up and stored	
108	Must be suitable for high rate testing	
	EMARKS	
109	Supplier's engineer shall provide installation support and check-out service.	
110	On-line system manuals on the CD	
111	The bidder must note that requirements in the above specifications are intended to be	
	descriptive only and not restrictive. The bids offered under other specifications will	
	be considered, provided that the bid clearly states the offered commodities are	
	substantially superior to those requirements in the above specification. (by descriptive	
	literature, illustrations, etc.)	
112	Demonstration of same or similar system in Civil/Geotechnical laboratory (preferably	
	in IIT's) during the course of technical evaluation	
113	User List preferably in civil/structural laboratory with contact person name, e-mail ids	
	and direct contact number, user certificate for satisfactory installation and	
	workmanship along with signature and stamp to be provided as testimonials, the	
	institute reserves the right to call each individual person for reference	
114	PO's of same/similar items should be provided along-with the technical bids	
115	Bidder should have a sound bank balance, annual report for the last five years to be	
	provided	
116	The supplier must be ISO 9000 series certificated	
117	Service Support: The Vendor shall define the service support plan in India and shall	
	utilize either factory direct service or through factory trained service personnel	
	located in India with a minimum of five service engineers on staff for servo-hydraulic	
	application. A commitment to the local community shall be evidenced by the	
	existence of this service support for at least 10 years	
118	One year system warranty after acceptance	
119	Spares and consumables to be quoted	
120	PO's of same/similar items should be provided along-with the technical bid	

Signature of Bidder

Name:	
Designation: _	
Organization Name:	
Contact No. :	

<< Organization Letter Head >> DECLARATION SHEET

gone through the specification, conditions and sintent of specification. This is certified that our organization has been further certified that our organization meets document. Moreover, OEM has agreed to support for the warranty. The prices quoted in the financial bids are subsident to the support for the warranty.	hereby certify that all the information and data furnished by diffication are true and complete to the best of our knowledge. I have stipulations in details and agree to comply with the requirements and authorized (Copy attached) by the OEM to participate in Tender. We all the conditions of eligibility criteria laid down in this tender port on regular basis with technology / product updates and extend dized due to academic discount given to IIT Delhi.
We, further specifically certify that our organization has not been Black Listed/De Listed or put to any Holiday by any Institutional Agency/ Govt. Department/ Public Sector Undertaking in the last three years.	NAME & ADDRESS OF THE Vendor/ Manufacturer / Agent
1 Phone	
2 Fax	
3 E-mail	
4 Contact Person Name	
5 Mobile Number	
6 TIN Number	
7 PAN Number	
(In case of on-line payment of Tender Fees)	
8 UTR No. (For Tender Fee)	
(In case of on-line payment of EMD)	
9 UTR No. (For EMD)	
10 Kindly provide bank details of the bidder in the following format: a) Name of the Bank	
b) Account Number	
c) Kindly attach scanned copy of one Cheque book page to enable us to return the EMD to unsuccessful bidder	

(Signature of the Tenderer)

Name:

Seal of the Company

List of Govt. Organization/Dept.

years (must be supported with work ord		
Name of the organization	Name of Contact Person	Contact No.
_		
Name of application specialist / Service	e Engineer who have the technical compet	ency to handle and
support the quoted product during the		·
Name of the organization	Name of Contact Person	Contact No.
		Signature of Bidde
	Name:	G
		Signature of Bidde

Bid Submission

Online Bid Submission:

The Online bids (complete in all respect) must be uploaded onlineintwo Envelops as explained below:-

Envelope – 1 (Following documents to be provided as single PDF file)							
Sl. No.	Documents	Content	File Types				
1.	Technical Bid	Compliance Sheet as per Annexure - I	.PDF				
2.		Organization Declaration Sheet as per Annexure - II	.PDF				
3.		List of organizations/ clients where the same products have been supplied (in last two years) along with their contact number(s). (Annexure-III)	.PDF				
4.		Technical supporting documents in support of all claims made at Annexure-I (Annexure-IV)	.PDF				
Sl. No.	Sl. No. TYPES Content						
1.	Financial Bid	Price bid should be submitted in PDF format.	.PDF				

Civil Engineering Department Indian Institute of Technology Delhi Hauz Khas, New Delhi-110016

Date:

Subject:Purchase of Loading Frame with Lifting Arrangement

	J	0			0 0						
S.	Currency	Description and Specification of	Qty.	Unit	Agency	Discount	Ex-works	Packing +	FOB/FCA	Insurance	CIF Price
No.		the Item	in	Price	Commission		price	Handling	Price	+ Freight	(f+g)
			Units		(If		(d=a+b-c)	+ DOC +	(f=d+e)	(g)	
				(a)	applicable)	(c)		Inland			
					(b)			Freight +			
								FCA			
								Charges			
								(e)			
1											

For indigenous items please quote as per following format.

S.	Description and Specification of the	Qty. in Units	Unit Price in	Excise Duty %	CST/VAT%	Octroi%	Total Price in
No.	Item		Rs.				Rs.
1.							
2.							

Note: The above financial template should be strictly followed. Any deviation from the above template (in terms of description and specification of the item) may lead to cancellation of the tender.