

2.	Column switching valve (As per specification given in the tender document)	01								
3.	Pressure Sensor: After system pump, pre-column & post column. Range: 0-3.5 MPa and above with accuracy ± 2 Psi. (As per specification given in the tender document)	01								

For indigenous items please quote as per following format.

S. No.	Description and Specification of the Item	Qty. in Units	Unit Price in Rs.	GST%	Total Price in Rs.
1.	Fast Protein Liquid Chromatography (FPLC) (As per specification given in the tender document)	01			

List of optional items

S. No.	Description and Specification of the Item	Qty. in Units	Unit Price in Rs.	GST%	Total Price in Rs.
1.	Analytical flow cell for 5 mm path length (As per specification given in the tender document)	01			

2.	Column switching valve (As per specification given in the tender document)	01			
3.	Pressure Sensor: After system pump, pre-column & post column. Range: 0-3.5 MPa and above with accuracy ± 2 Psi. (As per specification given in the tender document)	01			

Note: The above financial template should be strictly followed. Any deviation from the above template (in terms of description and specification of the item) may lead to cancellation of the tender.